

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061215\
 Data File : BM001660.D
 Acq On : 12 Jun 2015 03:38
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 12 16:48:28 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 16:27:37 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	89	0.00
2	1,4-Dioxane	1.000	1.021	-2.1	90	0.00
3	n-Nitrosodimethylamine	1.000	1.013	-1.3	92	0.00
4 S	2-Fluorophenol	1.000	0.943	5.7	88	0.00
5 S	Phenol-d6	1.000	0.946	5.4	89	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	90	0.00
7 S	Nitrobenzene-d5	1.000	0.928	7.2	88	0.00
8	Nitrobenzene	1.000	0.925	7.5	88	0.00
9	Naphthalene	1.000	0.942	5.8	90	0.00
10	Hexachlorobutadiene	1.000	0.954	4.6	93	0.00
11	2-Methylnaphthalene	1.000	0.963	3.7	92	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	92	0.00
13 S	2,4,6-Tribromophenol	1.000	0.994	0.6	90	0.00
14 S	2-Fluorobiphenyl	1.000	0.955	4.5	93	0.00
15	Acenaphthylene	1.000	0.957	4.3	92	-0.01
16	Acenaphthene	1.000	0.939	6.1	92	0.00
17	Fluorene	1.000	1.012	-1.2	99	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	92	0.00
19	4-Bromophenyl-phenylether	1.000	0.967	3.3	87	0.00
20	Hexachlorobenzene	1.000	1.095	-9.5	101	0.00
21	Pentachlorophenol	1.000	0.994	0.6	108	0.00
22	Phenanthrene	1.000	0.877	12.3	86	0.00
23	Anthracene	1.000	1.061	-6.1	98	0.00
24	Fluoranthene	1.000	0.972	2.8	92	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	90	0.00
26	Pyrene	1.000	0.974	2.6	91	0.00
27 S	Terphenyl-d14	1.000	0.967	3.3	89	-0.01
28	Benzo(a)anthracene	1.000	0.878	12.2	82	0.00
29	Chrysene	1.000	0.984	1.6	92	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.992	0.8	95	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.820	18.0	81	0.00
32 I	Perylene-d12	5.000	5.000	0.0	83	0.00
33	Benzo(b)fluoranthene	1.000	0.969	3.1	86	0.00
34	Benzo(k)fluoranthene	1.000	0.940	6.0	82	0.00
35 C	Benzo(a)pyrene	1.000	0.928	7.2	82	0.00
36	Dibenzo(a,h)anthracene	1.000	0.880	12.0	80	0.00
37	Benzo(g,h,i)perylene	1.000	0.992	0.8	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0