

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061215\
 Data File : BM001660.D
 Acq On : 12 Jun 2015 03:38
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 12 16:48:28 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 16:27:37 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.624	0.557	10.7	90	0.00
3	n-Nitrosodimethylamine	0.872	0.884	-1.4	92	0.00
4 S	2-Fluorophenol	1.184	1.116	5.7	88	0.00
5 S	Phenol-d6	1.488	1.408	5.4	89	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
7 S	Nitrobenzene-d5	0.383	0.355	7.3	88	0.00
8	Nitrobenzene	0.414	0.383	7.5	88	0.00
9	Naphthalene	1.212	1.142	5.8	90	0.00
10	Hexachlorobutadiene	0.201	0.192	4.5	93	0.00
11	2-Methylnaphthalene	0.782	0.753	3.7	92	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
13 S	2,4,6-Tribromophenol	0.182	0.181	0.5	90	0.00
14 S	2-Fluorobiphenyl	1.746	1.667	4.5	93	0.00
15	Acenaphthylene	2.337	2.236	4.3	92	-0.01
16	Acenaphthene	1.487	1.396	6.1	92	0.00
17	Fluorene	1.519	1.538	-1.3	99	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
19	4-Bromophenyl-phenylether	0.502	0.485	3.4	87	0.00
20	Hexachlorobenzene	0.455	0.498	-9.5	101	0.00
21	Pentachlorophenol	0.124	0.131	-5.6	108	0.00
22	Phenanthrene	1.640	1.438	12.3	86	0.00
23	Anthracene	1.938	2.057	-6.1	98	0.00
24	Fluoranthene	2.429	2.361	2.8	92	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
26	Pyrene	1.716	1.671	2.6	91	0.00
27 S	Terphenyl-d14	0.754	0.729	3.3	89	-0.01
28	Benzo(a)anthracene	1.571	1.379	12.2	82	0.00
29	Chrysene	1.503	1.479	1.6	92	0.00
30	Bis(2-ethylhexyl)phthalate	0.979	0.902	7.9	95	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	0.969	18.0	81	0.00
32 I	Perylene-d12	1.000	1.000	0.0	83	0.00
33	Benzo(b)fluoranthene	1.631	1.581	3.1	86	0.00
34	Benzo(k)fluoranthene	1.584	1.488	6.1	82	0.00
35 C	Benzo(a)pyrene	1.359	1.261	7.2	82	0.00
36	Dibenzo(a,h)anthracene	1.142	1.006	11.9	80	0.00
37	Benzo(g,h,i)perylene	1.234	1.130	8.4	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0