

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061215\
 Data File : BM001675.D
 Acq On : 12 Jun 2015 13:17
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 12 18:09:23 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 16:27:37 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	0.00
2	1,4-Dioxane	1.000	0.975	2.5	102	0.00
3	n-Nitrosodimethylamine	1.000	0.992	0.8	106	0.00
4 S	2-Fluorophenol	1.000	1.026	-2.6	113	0.00
5 S	Phenol-d6	1.000	1.045	-4.5	116	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	107	0.00
7 S	Nitrobenzene-d5	1.000	1.059	-5.9	120	0.00
8	Nitrobenzene	1.000	1.043	-4.3	118	0.00
9	Naphthalene	1.000	0.950	5.0	108	0.00
10	Hexachlorobutadiene	1.000	0.943	5.7	109	0.00
11	2-Methylnaphthalene	1.000	0.975	2.5	111	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	110	0.00
13 S	2,4,6-Tribromophenol	1.000	1.243	-24.3	135	0.00
14 S	2-Fluorobiphenyl	1.000	0.950	5.0	111	0.00
15	Acenaphthylene	1.000	1.018	-1.8	117	-0.01
16	Acenaphthene	1.000	0.959	4.1	113	0.00
17	Fluorene	1.000	1.042	-4.2	122	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	109	0.00
19	4-Bromophenyl-phenylether	1.000	0.946	5.4	102	0.00
20	Hexachlorobenzene	1.000	1.071	-7.1	118	0.00
21	Pentachlorophenol	1.000	1.097	-9.7	148	0.00
22	Phenanthrene	1.000	0.859	14.1	101	0.00
23	Anthracene	1.000	1.135	-13.5	125	0.00
24	Fluoranthene	1.000	0.949	5.1	108	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	93	0.00
26	Pyrene	1.000	1.074	-7.4	104	0.00
27 S	Terphenyl-d14	1.000	1.073	-7.3	103	0.00
28	Benzo(a)anthracene	1.000	0.982	1.8	95	0.00
29	Chrysene	1.000	0.964	3.6	94	0.02
30	Bis(2-ethylhexyl)phthalate	1.000	1.360	-36.0#	145	0.02
31	Indeno(1,2,3-cd)pyrene	1.000	1.001	-0.1	102	0.01
32 I	Perylene-d12	5.000	5.000	0.0	91	0.01
33	Benzo(b)fluoranthene	1.000	0.979	2.1	95	0.01
34	Benzo(k)fluoranthene	1.000	0.937	6.3	89	0.01
35 C	Benzo(a)pyrene	1.000	1.014	-1.4	98	0.01
36	Dibenzo(a,h)anthracene	1.000	1.050	-5.0	104	0.00
37	Benzo(g,h,i)perylene	1.000	1.089	-8.9	101	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0