

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061315\
 Data File : BM001685.D
 Acq On : 13 Jun 2015 00:34
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM061315

Quant Time: Jun 13 02:15:23 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 13 02:11:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.609	0.551	9.5	97	0.00
3	n-Nitrosodimethylamine	0.904	0.860	4.9	100	0.00
4 S	2-Fluorophenol	1.206	1.122	7.0	97	0.00
5 S	Phenol-d6	1.520	1.419	6.6	97	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
7 S	Nitrobenzene-d5	0.394	0.364	7.6	97	0.00
8	Nitrobenzene	0.426	0.388	8.9	97	0.00
9	Naphthalene	1.206	1.128	6.5	98	0.00
10	Hexachlorobutadiene	0.202	0.192	5.0	98	0.00
11	2-Methylnaphthalene	0.773	0.723	6.5	98	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
13 S	2,4,6-Tribromophenol	0.117	0.109	6.8	102	0.00
14 S	2-Fluorobiphenyl	1.774	1.675	5.6	99	0.00
15	Acenaphthylene	2.389	2.216	7.2	98	0.00
16	Acenaphthene	1.469	1.355	7.8	97	0.00
17	Fluorene	2.171	1.835	15.5	93	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
19	4-Bromophenyl-phenylether	0.095	0.093	2.1	99	0.00
20	Hexachlorobenzene	0.180	0.162	10.0	103	0.00
21	Pentachlorophenol	0.084	0.070	16.7	93	0.00
22	Phenanthrene	1.220	1.167	4.3	100	0.00
23	Anthracene	0.892	0.807	9.5	100	0.00
24	Fluoranthene	0.911	0.847	7.0	100	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
26	Pyrene	1.689	1.578	6.6	101	0.00
27 S	Terphenyl-d14	0.747	0.697	6.7	100	0.00
28	Benzo(a)anthracene	1.578	1.474	6.6	102	0.00
29	Chrysene	1.467	1.384	5.7	100	0.00
30	Bis(2-ethylhexyl)phthalate	0.920	0.738	19.8	98	0.00
31	Indeno(1,2,3-cd)pyrene	1.545	1.454	5.9	101	0.00
32 I	Perylene-d12	1.000	1.000	0.0	106	0.00
33	Benzo(b)fluoranthene	1.568	1.458	7.0	103	0.00
34	Benzo(k)fluoranthene	1.478	1.406	4.9	98	-0.01
35 C	Benzo(a)pyrene	1.367	1.277	6.6	100	0.00
36	Dibenzo(a,h)anthracene	1.356	1.282	5.5	100	0.00
37	Benzo(g,h,i)perylene	1.423	1.343	5.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0