

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061315\
 Data File : BM001693.D
 Acq On : 13 Jun 2015 05:24
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 13 06:49:56 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 13 02:11:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.609	0.591	3.0	96	0.00
3	n-Nitrosodimethylamine	0.904	0.898	0.7	96	0.00
4 S	2-Fluorophenol	1.206	1.212	-0.5	97	0.00
5 S	Phenol-d6	1.520	1.529	-0.6	96	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
7 S	Nitrobenzene-d5	0.394	0.389	1.3	95	0.00
8	Nitrobenzene	0.426	0.418	1.9	96	0.00
9	Naphthalene	1.206	1.196	0.8	95	0.00
10	Hexachlorobutadiene	0.202	0.199	1.5	94	0.00
11	2-Methylnaphthalene	0.773	0.766	0.9	95	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.117	0.125	-6.8	106	0.00
14 S	2-Fluorobiphenyl	1.774	1.789	-0.8	96	0.00
15	Acenaphthylene	2.389	2.376	0.5	95	0.00
16	Acenaphthene	1.469	1.450	1.3	94	0.00
17	Fluorene	2.171	2.015	7.2	93	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
19	4-Bromophenyl-phenylether	0.095	0.115	-21.1	111	0.00
20	Hexachlorobenzene	0.180	0.162	10.0	94	0.00
21	Pentachlorophenol	0.084	0.075	10.7	91	0.00
22	Phenanthrene	1.220	1.314	-7.7	102	0.00
23	Anthracene	0.892	0.805	9.8	91	0.00
24	Fluoranthene	0.911	0.921	-1.1	99	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
26	Pyrene	1.689	1.690	-0.1	101	0.00
27 S	Terphenyl-d14	0.747	0.730	2.3	97	0.00
28	Benzo(a)anthracene	1.578	1.577	0.1	102	0.00
29	Chrysene	1.467	1.492	-1.7	100	0.00
30	Bis(2-ethylhexyl)phthalate	0.920	0.866	5.9	107	0.00
31	Indeno(1,2,3-cd)pyrene	1.545	1.550	-0.3	100	0.00
32 I	Perylene-d12	1.000	1.000	0.0	100	0.00
33	Benzo(b)fluoranthene	1.568	1.566	0.1	104	0.00
34	Benzo(k)fluoranthene	1.478	1.515	-2.5	99	-0.01
35 C	Benzo(a)pyrene	1.367	1.378	-0.8	101	0.00
36	Dibenzo(a,h)anthracene	1.356	1.368	-0.9	100	0.00
37	Benzo(g,h,i)perylene	1.423	1.418	0.4	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0