


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061316\
 Data File : BM005770.D
 Acq On : 13 Jun 2016 15:25
 Operator : UM/SJ
 Sample : DFTPP55
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
 Title : SVOA CALIBRATION
 Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2546, 2547, 2548; Background Corrected with Scan 2540

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	28984	PASS
68	69	0.00	2	1.7	502	PASS
69	198	0.00	100	34.4	30221	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.4	40757	PASS
197	198	0.00	2	0.3	274	PASS
198	198	100	100	100.0	87757	PASS
199	198	5	9	7.0	6116	PASS
275	198	10	60	26.4	23181	PASS
365	198	1	100	3.4	2983	PASS
441	443	0.01	100	75.7	11216	PASS
442	198	50	100	85.6	75117	PASS
443	442	15	24	19.7	14809	PASS

m/z	Abundance
38.00	322.0
39.00	2472.0
49.00	303.0
50.00	8080.0
51.00	33896.0
52.00	1592.0
55.00	311.0
56.00	882.0
57.00	2173.0
61.00	401.0
61.90	466.0
63.00	1415.0
65.00	679.0
68.00	598.0
69.00	33456.0
74.00	3493.0
75.00	5612.0
76.10	1842.0
77.00	41200.0
78.00	2823.0
79.00	2604.0
79.90	1805.0
81.00	2842.0
82.00	844.0
82.90	775.0
85.00	513.0
85.90	851.0
86.90	341.0
90.90	619.0
92.00	729.0
93.00	4316.0
98.00	3216.0
99.00	2933.0
101.00	1668.0
103.00	429.0
104.00	932.0
105.00	945.0
107.00	12577.0
108.00	2077.0
109.00	489.0
110.00	22648.0
111.00	3393.0
112.00	407.0
116.00	612.0
117.00	10393.0
118.10	717.0
122.00	727.0
123.00	1218.0
124.00	573.0
125.00	457.0
127.00	42808.0
128.00	3270.0
129.00	15833.0
130.00	1426.0
133.90	448.0
135.00	1446.0
136.00	459.0
137.10	759.0
141.00	2174.0
141.90	632.0
143.00	419.0
146.00	392.0
147.00	1078.0
148.00	2509.0
149.00	591.0
151.00	376.0
153.00	674.0
153.90	521.0
155.00	1227.0
156.00	1742.0
157.00	491.0
158.00	479.0
160.00	638.0
160.90	1037.0
161.90	303.0
165.00	753.0
165.90	582.0
167.00	4331.0
168.00	2502.0
168.90	342.0
172.00	333.0
172.90	536.0
174.00	760.0
175.00	1820.0
176.10	516.0
177.00	681.0
178.00	303.0
178.90	3066.0
180.00	2354.0
180.90	1125.0
183.90	329.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

185.00	1513.0
186.00	12797.0
187.00	3575.0
188.10	352.0
188.90	832.0
190.90	426.0
192.00	1125.0
193.00	1155.0
196.00	2520.0
196.80	480.0
198.00	87512.0
199.00	6414.0
200.00	421.0
201.50	367.0
202.90	572.0
204.00	3457.0
205.00	5377.0
206.00	23816.0
207.00	3666.0
208.00	842.0
211.00	840.0
215.00	324.0
216.00	452.0
217.00	6084.0
218.00	753.0
221.00	3438.0
221.90	800.0
223.00	1338.0
224.00	12845.0
225.00	3494.0
226.00	345.0
227.00	5287.0
228.00	709.0
229.00	1002.0
231.00	446.0
233.90	328.0
234.90	328.0
237.00	462.0
242.00	713.0
243.00	740.0
244.00	9884.0
245.00	1382.0
246.00	1983.0
247.00	464.0
248.90	341.0
254.10	341.0
255.00	50264.0
256.00	7323.0
257.00	482.0
258.00	2823.0
258.90	416.0
265.00	1295.0
273.00	1649.0
274.00	3953.0
275.00	21856.0
276.00	3280.0
277.00	1853.0
278.00	357.0
285.00	402.0
292.90	503.0
296.00	6544.0
296.90	783.0
303.00	812.0
315.00	754.0
316.00	338.0
323.00	2030.0
327.00	318.0
334.00	1106.0
345.90	443.0
352.00	561.0
353.00	431.0
354.10	605.0
365.00	2564.0
365.90	374.0
372.00	987.0
402.00	415.0
403.00	524.0
421.00	555.0
422.00	429.0
423.00	3012.0
424.00	784.0
441.00	9030.0
442.00	60144.0
443.00	11953.0
444.00	1203.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
38.00	446.0
39.00	2485.0
49.00	407.0
50.00	8472.0
51.00	32824.0
52.00	1531.0
56.00	1065.0
57.00	2311.0
61.00	393.0
62.00	483.0
63.00	1323.0
65.00	610.0
68.00	562.0
69.00	35536.0
73.00	359.0
74.00	3580.0
75.00	5784.0
76.00	2139.0
77.00	43552.0
78.00	2988.0
79.00	2670.0
80.00	2076.0
81.00	3062.0
82.00	749.0
83.00	795.0
84.90	766.0
86.00	913.0
87.00	418.0
91.00	662.0
92.00	698.0
93.00	4809.0
94.00	358.0
98.00	3575.0
99.00	3126.0
101.00	1824.0
103.00	528.0
104.00	902.0
105.00	1032.0
106.00	382.0
107.00	14082.0
108.00	2077.0
109.00	370.0
110.00	24664.0
111.00	3700.0
111.90	322.0
115.90	721.0
117.00	11005.0
118.00	910.0
122.00	832.0
122.90	1221.0
124.00	582.0
125.00	531.0
127.00	48560.0
128.00	3653.0
129.00	17752.0
130.00	1466.0
134.00	462.0
134.90	1442.0
136.00	646.0
137.00	889.0
141.00	2414.0
142.00	740.0
142.90	472.0
145.90	486.0
147.00	1305.0
148.00	2775.0
149.00	564.0
151.00	334.0
153.00	804.0
154.00	497.0
155.00	1344.0
156.00	2256.0
157.10	369.0
157.90	380.0
159.00	386.0
160.00	685.0
161.00	1065.0
161.90	423.0
165.00	1019.0
166.00	815.0
167.00	5177.0
168.00	2864.0
169.00	518.0
171.90	434.0
173.00	604.0
174.00	1156.0
175.00	2086.0
176.00	635.0
177.00	1023.0
178.00	386.0
179.00	3855.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

180.00	2713.0
181.00	1275.0
185.00	1869.0
186.00	14566.0
187.00	4488.0
188.00	530.0
189.00	849.0
190.90	480.0
192.00	1320.0
193.00	1510.0
194.00	334.0
196.00	2644.0
197.10	488.0
198.00	104368.0
198.90	7256.0
199.90	545.0
201.60	356.0
202.90	708.0
204.00	4139.0
205.00	6912.0
206.00	27984.0
207.00	4196.0
208.00	1188.0
208.90	393.0
210.00	445.0
211.00	1100.0
215.00	304.0
216.00	619.0
217.00	7096.0
217.90	802.0
221.00	3898.0
221.80	1053.0
223.00	1521.0
224.00	15218.0
225.00	4207.0
225.90	463.0
227.00	6839.0
228.00	910.0
229.00	1403.0
231.00	637.0
234.00	401.0
234.90	497.0
236.00	359.0
237.00	583.0
241.00	336.0
242.00	804.0
243.10	847.0
244.00	12206.0
245.00	1637.0
246.00	2462.0
247.00	437.0
248.90	476.0
254.10	480.0
255.00	63152.0
256.00	9384.0
257.00	635.0
258.00	3678.0
258.90	558.0
265.00	1429.0
273.00	1728.0
274.00	5124.0
275.00	27224.0
276.00	3548.0
277.00	2284.0
278.00	340.0
283.00	333.0
285.00	418.0
292.90	542.0
296.00	8242.0
297.00	1039.0
303.00	979.0
314.00	376.0
315.00	1080.0
316.00	586.0
323.00	2254.0
324.00	558.0
327.00	495.0
328.00	314.0
334.00	1706.0
335.00	416.0
341.00	350.0
346.00	676.0
352.10	665.0
353.00	517.0
354.00	584.0
365.00	3500.0
365.90	498.0
372.00	1409.0
373.00	336.0
383.00	309.0
401.90	520.0
403.00	807.0
421.00	621.0
421.90	654.0
423.00	4445.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

424.00	995.0
441.00	12798.0
442.10	87808.0
443.10	17576.0
444.00	1464.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

m/z	Abundance
39.00	1371.0
49.00	301.0
50.00	5035.0
51.00	20232.0
52.00	989.0
56.00	607.0
57.00	1415.0
63.00	791.0
65.00	493.0
68.00	347.0
69.00	21672.0
74.00	2340.0
75.00	3526.0
76.10	1211.0
77.00	26512.0
78.00	1841.0
79.00	1629.0
79.90	1264.0
81.00	1916.0
81.90	523.0
83.00	516.0
85.00	436.0
86.00	496.0
91.00	409.0
91.90	448.0
93.00	3157.0
98.00	2382.0
99.00	1840.0
100.90	875.0
103.00	391.0
104.00	583.0
104.90	698.0
107.00	8586.0
108.00	1269.0
110.00	15492.0
111.00	2289.0
115.90	514.0
117.00	6957.0
117.90	506.0
121.90	617.0
123.00	865.0
123.90	341.0
125.00	433.0
127.00	30904.0
128.00	2256.0
129.00	11679.0
130.00	914.0
133.90	349.0
135.00	938.0
136.00	401.0
136.90	535.0
141.00	1450.0
141.90	523.0
142.90	335.0
146.90	1020.0
148.00	1862.0
149.00	381.0
152.90	446.0
154.00	350.0
155.00	937.0
156.00	1489.0
158.90	332.0
160.00	597.0
161.00	712.0
164.90	741.0
165.90	548.0
167.00	3459.0
168.00	1979.0
171.90	382.0
172.90	369.0
173.90	714.0
175.00	1349.0
176.90	622.0
179.00	2443.0
180.00	1742.0
181.00	809.0
185.00	1329.0
186.00	9998.0
187.00	2854.0
188.00	376.0
188.90	569.0
191.00	302.0
192.00	767.0
193.00	1050.0
196.00	2158.0
196.70	342.0
197.00	323.0
197.90	71392.0
199.00	4678.0
199.90	398.0
201.40	357.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55

203.00	500.0
204.00	2852.0
205.00	4451.0
206.00	19816.0
207.00	2576.0
208.00	727.0
210.10	346.0
211.00	790.0
215.90	440.0
217.00	5049.0
218.00	705.0
221.00	3377.0
221.80	649.0
223.00	1077.0
224.00	10765.0
225.00	2879.0
227.00	4757.0
228.00	605.0
229.00	914.0
230.90	424.0
235.00	362.0
241.00	317.0
242.00	613.0
243.00	576.0
244.00	9109.0
245.00	1203.0
246.00	1785.0
246.90	326.0
249.00	306.0
254.00	363.0
255.00	45800.0
256.00	6484.0
257.00	473.0
258.00	2764.0
259.00	377.0
264.90	1149.0
273.00	1430.0
274.00	4180.0
275.00	20464.0
276.00	2632.0
277.00	1860.0
278.00	308.0
285.10	326.0
292.90	469.0
296.00	5871.0
297.00	779.0
303.00	714.0
315.00	590.0
316.00	392.0
323.00	1989.0
324.00	312.0
326.90	457.0
334.00	1181.0
335.00	366.0
345.90	484.0
352.00	551.0
353.00	345.0
354.00	575.0
365.00	2887.0
365.90	396.0
372.00	1229.0
382.90	391.0
401.90	421.0
403.00	628.0
421.00	458.0
422.00	561.0
423.00	3756.0
424.00	712.0
441.00	11822.0
442.00	77400.0
443.00	14900.0
444.00	1061.0

Instrument :

BNA_M

ClientSampleId :

DFTPP55

m/z	Abundance
206.90	330.0

Instrument :
BNA_M
ClientSampleId :
DFTPP55