

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061421\
 Data File : BM030445.D
 Acq On : 14 Jun 2021 20:14
 Operator : CG/JU
 Sample : M2651-08
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DEP-CRANE-PAD-EXCAVATION

Quant Time: Jun 15 03:51:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 14:10:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	175468	20.000	ng	0.00
21) Naphthalene-d8	10.622	136	632185	20.000	ng	0.00
39) Acenaphthene-d10	14.457	164	369081	20.000	ng	0.00
64) Phenanthrene-d10	17.192	188	737811	20.000	ng	0.00
76) Chrysene-d12	21.351	240	838059	20.000	ng	0.00
86) Perylene-d12	23.627	264	942282	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	1482612	137.360	ng	0.01
7) Phenol-d6	7.010	99	1639298	104.913	ng	0.01
23) Nitrobenzene-d5	8.987	82	1190983	91.780	ng	0.00
42) 2,4,6-Tribromophenol	15.945	330	693208	149.386	ng	0.00
45) 2-Fluorobiphenyl	13.080	172	2548722	90.813	ng	0.00
79) Terphenyl-d14	19.804	244	4162795	87.246	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.268	149	182244	8.470	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061421\
Data File : BM030445.D
Acq On : 14 Jun 2021 20:14
Operator : CG/JU
Sample : M2651-08
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DEP-CRANE-PAD-EXCAVATION

Quant Time: Jun 15 03:51:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 11 14:10:50 2021
Response via : Initial Calibration

