

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061422\
 Data File : BM035469.D
 Acq On : 14 Jun 2022 17:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 03:07:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM061022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 10 16:05:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	39.036	2.4	99	0.00
3	Pyridine	40.000	41.085	-2.7	101	0.00
4	n-Nitrosodimethylamine	40.000	40.935	-2.3	101	0.00
5 S	2-Fluorophenol	80.000	80.749	-0.9	100	0.00
6	Aniline	40.000	42.161	-5.4	105	0.00
7 S	Phenol-d6	80.000	82.023	-2.5	103	0.00
8	2-Chlorophenol	40.000	41.414	-3.5	104	0.00
9	Benzaldehyde	40.000	37.854	5.4	103	0.00
10 C	Phenol	40.000	41.292	-3.2	103	0.00
11	bis(2-Chloroethyl)ether	40.000	40.936	-2.3	103	0.00
12	1,3-Dichlorobenzene	40.000	40.137	-0.3	101	0.00
13 C	1,4-Dichlorobenzene	40.000	40.304	-0.8	102	0.00
14	1,2-Dichlorobenzene	40.000	40.627	-1.6	102	0.00
15	Benzyl Alcohol	40.000	42.280	-5.7	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.356	6.6	97	0.00
17	2-Methylphenol	40.000	42.521	-6.3	105	0.00
18	Hexachloroethane	40.000	40.512	-1.3	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.430	-3.6	103	0.00
20	3+4-Methylphenols	40.000	42.331	-5.8	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	40.805	-2.0	105	0.00
23 S	Nitrobenzene-d5	80.000	80.290	-0.4	104	0.00
24	Nitrobenzene	40.000	40.504	-1.3	106	0.00
25	Isophorone	40.000	41.316	-3.3	106	0.00
26 C	2-Nitrophenol	40.000	43.650	-9.1	109	0.00
27	2,4-Dimethylphenol	40.000	41.675	-4.2	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.864	-2.2	106	0.00
29 C	2,4-Dichlorophenol	40.000	42.152	-5.4	107	0.00
30	1,2,4-Trichlorobenzene	40.000	40.572	-1.4	105	0.00
31	Naphthalene	40.000	40.405	-1.0	105	0.00
32	Benzoic acid	40.000	42.266	-5.7	114	0.01
33	4-Chloroaniline	40.000	43.018	-7.5	110	0.00
34 C	Hexachlorobutadiene	40.000	40.164	-0.4	104	0.00
35	Caprolactam	40.000	46.653	-16.6	122	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.033	-5.1	109	0.00
37	2-Methylnaphthalene	40.000	40.945	-2.4	108	0.00
38	1-Methylnaphthalene	40.000	40.885	-2.2	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.914	2.7	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.405	4.0	111	0.00
42 S	2,4,6-Tribromophenol	80.000	90.882	-13.6	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.686	-1.7	111	0.00
44	2,4,5-Trichlorophenol	40.000	40.596	-1.5	111	0.00
45 S	2-Fluorobiphenyl	80.000	77.515	3.1	109	0.00
46	1,1'-Biphenyl	40.000	38.721	3.2	110	0.00
47	2-Chloronaphthalene	40.000	39.254	1.9	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061422\
 Data File : BM035469.D
 Acq On : 14 Jun 2022 17:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 03:07:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM061022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 10 16:05:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.126	-2.8	115	0.00
49	Acenaphthylene	40.000	39.759	0.6	112	0.00
50	Dimethylphthalate	40.000	40.273	-0.7	116	0.00
51	2,6-Dinitrotoluene	40.000	41.307	-3.3	116	0.00
52 C	Acenaphthene	40.000	40.570	-1.4	112	0.00
53	3-Nitroaniline	40.000	43.303	-8.3	120	0.00
54 P	2,4-Dinitrophenol	40.000	43.125	-7.8	131	0.00
55	Dibenzofuran	40.000	40.819	-2.0	112	0.00
56 P	4-Nitrophenol	40.000	43.281	-8.2	129	0.00
57	2,4-Dinitrotoluene	40.000	45.098	-12.7	121	0.00
58	Fluorene	40.000	41.706	-4.3	118	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.987	-10.0	118	0.00
60	Diethylphthalate	40.000	42.213	-5.5	119	0.00
61	4-Chlorophenyl-phenylether	40.000	41.221	-3.1	116	0.00
62	4-Nitroaniline	40.000	47.514	-18.8	131	0.00
63	Azobenzene	40.000	40.588	-1.5	118	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	126	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.542	-11.4	135	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.366	-3.4	121	0.00
67	4-Bromophenyl-phenylether	40.000	39.845	0.4	120	0.00
68	Hexachlorobenzene	40.000	40.044	-0.1	122	0.00
69	Atrazine	40.000	43.332	-8.3	131	0.00
70 C	Pentachlorophenol	40.000	45.450	-13.6	139	0.00
71	Phenanthrene	40.000	40.741	-1.9	125	0.00
72	Anthracene	40.000	41.358	-3.4	127	0.00
73	Carbazole	40.000	43.204	-8.0	132	0.00
74	Di-n-butylphthalate	40.000	44.497	-11.2	136	0.00
75 C	Fluoranthene	40.000	43.952	-9.9	136	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	142	0.00
77	Benzydine	40.000	41.369	-3.4	129	0.00
78	Pyrene	40.000	38.151	4.6	136	0.00
79 S	Terphenyl-d14	80.000	75.732	5.3	134	0.00
80	Butylbenzylphthalate	40.000	42.613	-6.5	152	0.00
81	Benzo(a)anthracene	40.000	40.272	-0.7	141	0.00
82	3,3'-Dichlorobenzidine	40.000	42.510	-6.3	141	0.00
83	Chrysene	40.000	40.302	-0.8	142	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.031	-7.6	151	0.00
85 c	Di-n-octyl phthalate	40.000	45.429	-13.6	156	0.00
86 I	Perylene-d12	20.000	20.000	0.0	135	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.684	10.8	116	0.00
88	Benzo(b)fluoranthene	40.000	41.063	-2.7	141	0.00
89	Benzo(k)fluoranthene	40.000	41.323	-3.3	136	0.00
90 C	Benzo(a)pyrene	40.000	40.707	-1.8	135	0.00
91	Dibenzo(a,h)anthracene	40.000	36.059	9.9	117	0.00
92	Benzo(g,h,i)perylene	40.000	34.951	12.6	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061422\
Data File : BM035469.D
Acq On : 14 Jun 2022 17:01
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jun 15 03:07:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM061022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 10 16:05:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0