

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM061520\
 Data File : BM026397.D
 Acq On : 15 Jun 2020 18:32
 Operator : JU/CG
 Sample : L3008-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MH-725-D-WATERING-BAG

Quant Time: Jun 16 02:21:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 12:22:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	99206	20.00	ng	0.00
21) Naphthalene-d8	10.16	136	410331	20.00	ng	0.00
39) Acenaphthene-d10	14.06	164	250515	20.00	ng	0.00
64) Phenanthrene-d10	16.81	188	531862	20.00	ng	0.00
76) Chrysene-d12	21.02	240	552617	20.00	ng	0.00
86) Perylene-d12	23.11	264	586965	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.04	112	274738	48.96	ng	0.00
7) Phenol-d6	6.61	99	231202	28.46	ng	0.00
23) Nitrobenzene-d5	8.55	82	726237	86.90	ng	0.00
42) 2,4,6-Tribromophenol	15.56	330	423339	139.50	ng	0.00
45) 2-Fluorobiphenyl	12.67	172	1520263	92.12	ng	0.00
79) Terphenyl-d14	19.47	244	2543484	89.34	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.53	163	279765	15.268	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM061520\
Data File : BM026397.D
Acq On : 15 Jun 2020 18:32
Operator : JU/CG
Sample : L3008-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MH-725-D-WATERING-BAG

Quant Time: Jun 16 02:21:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 12:22:53 2020
Response via : Initial Calibration

