

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061522\
 Data File : BM035472.D
 Acq On : 15 Jun 2022 11:14
 Operator : CG/JU
 Sample : PB145531BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB145531BL

Quant Time: Jun 15 23:47:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM061022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 10 16:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	303754	20.000	ng	0.00
21) Naphthalene-d8	10.616	136	1330346	20.000	ng	0.00
39) Acenaphthene-d10	14.457	164	809355	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	1624285	20.000	ng	# 0.00
76) Chrysene-d12	21.392	240	1321971	20.000	ng	0.00
86) Perylene-d12	23.739	264	1211172	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.405	112	2716862	151.676	ng	0.00
7) Phenol-d6	7.005	99	3533084	142.335	ng	0.00
23) Nitrobenzene-d5	8.981	82	2124928	86.801	ng	0.00
42) 2,4,6-Tribromophenol	15.951	330	1143904	132.751	ng	0.00
45) 2-Fluorobiphenyl	13.081	172	4891102	86.334	ng	0.00
79) Terphenyl-d14	19.833	244	6322256	91.314	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061522\
Data File : BM035472.D
Acq On : 15 Jun 2022 11:14
Operator : CG/JU
Sample : PB145531BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB145531BL

Quant Time: Jun 15 23:47:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM061022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 10 16:05:33 2022
Response via : Initial Calibration

