

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061716\
 Data File : BM005839.D
 Acq On : 16 Jun 2016 22:02
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02051

Quant Time: Jun 17 01:48:51 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:45:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 S	1,4-Dioxane-d8	0.154	0.150	2.6	91	0.00
3 S	Phenol-d5	1.674	1.565	6.5	99	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.868	0.842	3.0	99	0.00
5 S	2-Chlorophenol-d4	1.396	1.373	1.6	99	0.00
6 S	4-Methylphenol-d8	1.497	1.463	2.3	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.141	0.143	-1.4	99	0.00
9 S	2-Nitrophenol-d4	0.172	0.181	-5.2	104	0.00
10 S	2,4-Dichlorophenol-d3	0.341	0.343	-0.6	99	0.00
11	Naphthalene	1.072	1.044	2.6	101	0.00
12 S	4-Chloroaniline-d4	0.231	0.236	-2.2	104	0.00
13	2-Methylnaphthalene	0.817	0.818	-0.1	102	0.00
14	1-Methylnaphthalene	0.851	0.840	1.3	102	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
16 S	Dimethylphthalate-d6	1.892	1.868	1.3	103	0.00
17 S	Acenaphthylene-d8	2.180	2.139	1.9	102	0.00
18	Acenaphthylene	2.207	2.164	1.9	103	0.00
19 C	Acenaphthene	1.473	1.432	2.8	102	0.00
20 S	4-Nitrophenol-d4	0.228	0.205	10.1	103	0.00
21 S	Fluorene-d10	1.612	1.563	3.0	103	0.00
22	Fluorene	1.786	1.739	2.6	102	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.135	0.124	8.1	106	0.00
25	Phenanthrene	1.191	1.168	1.9	103	0.00
26 S	Anthracene-d10	1.044	1.018	2.5	103	0.00
27	Anthracene	1.217	1.207	0.8	104	0.00
28 C	Fluoranthene	1.496	1.461	2.3	103	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
30 S	Pyrene-d10	0.948	0.933	1.6	101	0.00
31	Pyrene	1.182	1.181	0.1	103	0.00
32	Benzo(a)anthracene	1.249	1.238	0.9	104	0.00
33	Chrysene	1.203	1.195	0.7	102	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.245	1.200	3.6	99	0.00
36	Benzo(k)fluoranthene	1.202	1.217	-1.2	107	0.00
37 S	Benzo(a)pyrene-d12	1.000	0.990	1.0	104	0.00
38 C	Benzo(a)pyrene	1.222	1.212	0.8	102	0.00
39	Indeno(1,2,3-cd)pyrene	1.494	1.465	1.9	105	0.00
40	Dibenzo(a,h)anthracene	1.229	1.211	1.5	103	0.00
41	Benzo(g,h,i)perylene	1.272	1.245	2.1	103	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061716\
Data File : BM005839.D
Acq On : 16 Jun 2016 22:02
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02051

Quant Time: Jun 17 01:48:51 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061716.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 01:45:30 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0