

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061716\
 Data File : BM005854.D
 Acq On : 17 Jun 2016 07:02
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02052

Quant Time: Jun 17 07:34:49 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:56:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	160#	0.00
2 S	1,4-Dioxane-d8	0.154	0.139	9.7	135	0.00
3 S	Phenol-d5	1.674	1.543	7.8	158#	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.868	0.799	7.9	151#	0.00
5 S	2-Chlorophenol-d4	1.396	1.402	-0.4	163#	0.00
6 S	4-Methylphenol-d8	1.497	1.397	6.7	154#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	152#	0.00
8 S	Nitrobenzene-d5	0.141	0.153	-8.5	159#	0.00
9 S	2-Nitrophenol-d4	0.172	0.186	-8.1	160#	0.00
10 S	2,4-Dichlorophenol-d3	0.341	0.355	-4.1	154#	0.00
11	Naphthalene	1.072	1.048	2.2	151#	0.00
12 S	4-Chloroaniline-d4	0.231	0.307	-32.9#	202#	0.00
13	2-Methylnaphthalene	0.817	0.792	3.1	148	0.00
14	1-Methylnaphthalene	0.851	0.790	7.2	143	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	136	0.00
16 S	Dimethylphthalate-d6	1.892	1.747	7.7	127	0.00
17 S	Acenaphthylene-d8	2.180	2.158	1.0	135	0.00
18	Acenaphthylene	2.207	2.152	2.5	135	0.00
19 C	Acenaphthene	1.473	1.431	2.9	135	0.00
20 S	4-Nitrophenol-d4	0.228	0.199	12.7	132	0.00
21 S	Fluorene-d10	1.612	1.520	5.7	132	0.00
22	Fluorene	1.786	1.668	6.6	129	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.135	0.104	23.0#	104	0.00
25	Phenanthrene	1.191	1.163	2.4	122	0.00
26 S	Anthracene-d10	1.044	1.024	1.9	122	0.00
27	Anthracene	1.217	1.201	1.3	122	0.00
28 C	Fluoranthene	1.496	1.380	7.8	115	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
30 S	Pyrene-d10	0.948	1.071	-13.0	111	0.00
31	Pyrene	1.182	1.363	-15.3	113	0.00
32	Benzo(a)anthracene	1.249	1.243	0.5	99	0.00
33	Chrysene	1.203	1.183	1.7	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	75	0.00
35	Benzo(b)fluoranthene	1.245	1.251	-0.5	75	0.00
36	Benzo(k)fluoranthene	1.202	1.257	-4.6	80	0.00
37 S	Benzo(a)pyrene-d12	1.000	0.978	2.2	75	0.00
38 C	Benzo(a)pyrene	1.222	1.190	2.6	73	0.01
39	Indeno(1,2,3-cd)pyrene	1.494	1.336	10.6	69	0.00
40	Dibenzo(a,h)anthracene	1.229	1.098	10.7	68	0.00
41	Benzo(g,h,i)perylene	1.272	1.123	11.7	67	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061716\
Data File : BM005854.D
Acq On : 17 Jun 2016 07:02
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02052

Quant Time: Jun 17 07:34:49 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061716.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 01:56:56 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0