

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061721\
 Data File : BN014939.D
 Acq On : 17 Jun 2021 12:57
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV007

Quant Time: Jun 17 13:31:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN061721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 13:23:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	6385	0.40	ng/ul	-0.01
4) Naphthalene-d8	10.51	136	23833	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.37	164	13558	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.11	188	28725	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	35868	0.40	ng/ul	0.00
23) Perylene-d12	23.58	264	27056	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.30	96	3679	0.41	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	16166	0.40	ng/ul	0.00
18) Fluoranthene-d10	19.15	212	33921	0.30	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.33	88	3460	0.411	ng/ul	97
5) Naphthalene	10.56	128	31900	0.422	ng/ul	100
7) 2-Methylnaphthalene	12.18	142	20800	0.419	ng/ul	100
8) 1-Methylnaphthalene	12.40	142	19428	0.405	ng/ul	100
10) Acenaphthylene	14.08	152	29364	0.411	ng/ul	100
11) Acenaphthene	14.43	153	21155	0.418	ng/ul	99
12) Fluorene	15.42	166	24604	0.424	ng/ul	99
14) Pentachlorophenol	16.76	266	3358	0.467	ng/ul	99
15) Phenanthrene	17.16	178	40902	0.427	ng/ul	99
16) Anthracene	17.25	178	41397	0.444	ng/ul	100
19) Fluoranthene	19.18	202	50528	0.352	ng/ul	99
20) Pyrene	19.54	202	52281	0.355	ng/ul	99
21) Benzo(a)anthracene	21.31	228	47844	0.368	ng/ul	99
22) Chrysene	21.36	228	56775	0.402	ng/ul	99
24) Benzo(b)fluoranthene	22.90	252	43152	0.407	ng/ul	100
25) Benzo(k)fluoranthene	22.95	252	51791	0.421	ng/ul	99
26) Benzo(a)pyrene	23.48	252	38766	0.399	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	25.86	276	45829	0.409	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.88	278	36904	0.414	ng/ul	99
29) Benzo(g,h,i)perylene	26.55	276	42194	0.411	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061721\
Data File : BN014939.D
Acq On : 17 Jun 2021 12:57
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SICV007

Quant Time: Jun 17 13:31:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN061721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 13:23:24 2021
Response via : Initial Calibration

