



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061721\
 Data File : BM030486.D
 Acq On : 17 Jun 2021 09:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP052

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM060621.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jun 17 13:10:35 2021



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.7	146673	PASS
68	69	0.00	2	1.6	2386	PASS
69	198	0.00	100	35.2	148964	PASS
70	69	0.00	2	0.5	752	PASS
127	198	10	80	45.1	190757	PASS
197	198	0.00	2	0.5	2043	PASS
198	198	100	100	100.0	422720	PASS
199	198	5	9	6.9	29115	PASS
275	198	10	60	30.1	127160	PASS
365	198	1	100	4.3	18023	PASS
441	198	0.01	100	11.9	50331	PASS
442	198	50	100	79.2	334997	PASS
443	442	15	24	19.1	64096	PASS

m/z	Abundance
36.00	1141.0
37.00	1251.0
37.90	412.0
39.00	379.0
40.00	1052.0
41.10	2034.0
42.00	934.0
43.10	1277.0
44.00	2210.0
45.00	508.0
47.00	6989.0
48.00	3318.0
49.00	37744.0
50.00	1614.0
51.00	11687.0
52.10	188.0
53.10	157.0
55.00	824.0
56.10	304.0
57.00	535.0
58.00	193.0
60.00	343.0
67.10	259.0
67.90	159.0
69.00	340.0
70.00	332.0
70.90	281.0
71.90	219.0
73.00	960.0
73.80	213.0
75.00	258.0
76.90	408.0
79.00	183.0
81.00	249.0
82.00	401.0
82.90	1001.0
84.00	24368.0
85.00	1095.0
86.00	15988.0
86.90	282.0
88.00	2552.0
90.90	302.0
93.00	170.0
95.00	299.0
96.00	248.0
97.00	216.0
103.00	155.0
105.00	354.0
109.10	243.0
117.00	217.0
119.00	153.0
121.10	162.0
129.10	244.0
132.90	355.0
134.90	229.0
144.90	169.0
146.90	355.0
148.90	202.0
165.10	211.0
167.00	175.0
180.00	153.0
190.90	363.0
193.00	409.0
207.00	2426.0
208.00	812.0
209.10	401.0
221.10	315.0
281.00	783.0
281.90	322.0
282.90	279.0
341.00	182.0
355.00	260.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

m/z	Abundance
36.00	1422.0
37.00	2101.0
38.00	2855.0
39.10	14527.0
40.00	1404.0
41.10	1950.0
41.90	1051.0
43.00	1245.0
44.00	2307.0
45.00	887.0
47.00	6991.0
48.00	3795.0
49.00	39040.0
50.00	50120.0
51.00	196160.0
52.00	8968.0
53.00	544.0
54.00	205.0
55.00	1519.0
56.00	5780.0
57.00	12905.0
58.00	823.0
59.00	417.0
60.00	370.0
61.00	2002.0
62.00	2865.0
63.00	7254.0
64.10	863.0
65.00	3449.0
66.00	452.0
67.00	433.0
68.00	3202.0
69.00	186752.0
70.00	1244.0
71.00	287.0
71.90	297.0
73.00	2872.0
74.00	19056.0
75.00	29600.0
76.10	10491.0
77.10	212864.0
78.10	15200.0
79.00	13525.0
80.00	11142.0
81.00	15047.0
82.00	4751.0
83.00	4154.0
84.00	26680.0
85.00	3294.0
86.00	21992.0
87.00	2379.0
88.00	3350.0
88.90	362.0
91.00	3299.0
92.00	3535.0
93.00	25192.0
94.00	1733.0
95.10	595.0
96.00	1213.0
96.90	669.0
98.00	19336.0
99.00	14472.0
100.00	1223.0
101.00	8574.0
101.90	391.0
103.00	2613.0
104.00	5475.0
105.00	5518.0
106.00	1753.0
107.00	63776.0
108.00	10822.0
109.10	1789.0
110.00	114504.0
111.00	17864.0
112.00	2220.0
113.00	764.0
113.90	184.0
114.90	379.0
116.00	3253.0
117.00	59536.0
118.00	4636.0
118.90	551.0
120.00	826.0
121.10	471.0
122.00	3925.0
123.00	6063.0
124.00	2806.0
125.00	2637.0
126.10	932.0
127.00	231936.0
128.00	17168.0
129.00	88128.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

130.00	7102.0
131.00	1490.0
131.90	843.0
132.90	731.0
134.00	2360.0
135.00	6909.0
136.00	2862.0
137.00	3934.0
138.00	1040.0
138.90	431.0
139.90	1116.0
141.00	11614.0
142.00	3691.0
142.90	2456.0
143.90	733.0
145.00	651.0
146.00	1889.0
147.00	6516.0
148.00	14689.0
149.00	3495.0
150.00	786.0
151.00	1604.0
152.00	925.0
153.00	3774.0
154.00	2998.0
155.00	6441.0
156.00	9681.0
157.10	2460.0
158.00	2085.0
159.00	1627.0
160.00	3485.0
161.00	5669.0
162.00	1760.0
163.00	421.0
164.10	713.0
165.00	4724.0
166.00	3347.0
167.00	25584.0
168.00	13510.0
169.00	2294.0
170.00	1170.0
170.90	1020.0
172.00	2309.0
173.00	3185.0
174.00	5400.0
175.00	10655.0
176.10	2786.0
177.00	4303.0
178.00	1566.0
179.00	19104.0
180.00	13228.0
181.00	6121.0
182.00	1000.0
183.00	469.0
184.00	1355.0
185.00	8396.0
186.00	70240.0
187.00	19872.0
188.00	2043.0
189.00	4520.0
189.90	648.0
191.00	2298.0
192.00	5914.0
193.00	7559.0
194.00	1300.0
195.00	688.0
196.00	13444.0
197.00	2275.0
198.00	483392.0
199.00	33352.0
200.00	2654.0
201.40	2517.0
202.00	724.0
203.00	3584.0
204.00	19352.0
205.00	31528.0
206.00	130240.0
207.00	20608.0
208.00	5259.0
209.00	2054.0
210.00	1750.0
211.00	5364.0
211.80	1092.0
213.10	636.0
215.00	1611.0
216.00	2862.0
217.00	36616.0
218.00	4609.0
218.90	406.0
221.00	20568.0
221.90	4436.0
223.00	7795.0
224.00	74560.0
225.10	19392.0
226.00	2311.0
227.00	32576.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

228.00	4819.0
229.00	6650.0
230.00	1210.0
231.00	2684.0
232.00	456.0
233.00	655.0
234.00	1911.0
235.00	2499.0
236.00	1547.0
237.00	3249.0
237.90	463.0
239.00	1310.0
240.00	983.0
241.00	1691.0
242.00	4146.0
243.00	4232.0
244.00	57072.0
245.00	7647.0
246.00	11829.0
246.90	2209.0
248.00	657.0
249.00	1931.0
250.00	587.0
250.90	488.0
252.00	546.0
253.00	1348.0
254.00	1952.0
255.00	302976.0
256.00	44072.0
257.00	3614.0
258.00	18072.0
259.00	3174.0
260.00	532.0
260.80	467.0
261.90	151.0
263.10	172.0
264.00	490.0
265.00	7830.0
266.00	1285.0
266.80	226.0
270.00	252.0
270.90	615.0
272.10	656.0
273.00	9565.0
274.00	22280.0
275.00	135872.0
276.00	18688.0
277.00	11134.0
278.00	2163.0
278.90	429.0
281.00	881.0
281.90	448.0
283.00	1344.0
284.00	842.0
285.10	2051.0
286.00	355.0
288.00	191.0
289.00	516.0
290.10	436.0
290.90	471.0
292.00	580.0
293.00	2580.0
294.00	817.0
294.90	698.0
296.00	38016.0
297.00	6413.0
297.90	232.0
301.10	690.0
301.90	785.0
303.00	4285.0
304.00	1338.0
307.90	516.0
309.00	414.0
309.90	571.0
311.00	184.0
311.90	165.0
312.90	406.0
314.00	1509.0
315.00	4250.0
316.00	2269.0
317.00	442.0
321.00	954.0
322.00	643.0
323.00	11916.0
324.00	2536.0
325.00	218.0
325.90	276.0
327.00	2591.0
328.00	1183.0
329.10	248.0
332.00	1020.0
333.00	1172.0
334.00	7853.0
335.00	1919.0
336.00	404.0
339.90	150.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

341.00	1654.0
342.00	497.0
346.00	3014.0
346.90	444.0
350.90	258.0
352.00	3701.0
353.00	2400.0
354.00	3634.0
355.00	975.0
356.00	212.0
359.00	374.0
364.00	253.0
365.00	18144.0
366.00	2452.0
366.90	219.0
369.90	262.0
371.00	868.0
372.10	5646.0
373.10	1437.0
374.00	176.0
382.90	1592.0
383.90	354.0
384.90	259.0
390.10	930.0
391.00	468.0
392.00	316.0
401.00	235.0
402.00	2334.0
403.00	3582.0
404.00	1163.0
405.00	318.0
415.10	229.0
421.00	3002.0
422.00	2507.0
423.00	17824.0
424.00	5078.0
425.10	529.0
440.10	224.0
441.10	46608.0
442.10	305152.0
443.10	59880.0
444.10	5554.0
445.10	315.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

m/z	Abundance
35.90	1088.0
37.00	2293.0
38.00	2552.0
39.10	12846.0
40.10	1254.0
41.00	2358.0
42.00	1186.0
43.00	1380.0
44.00	2280.0
45.00	787.0
47.00	7383.0
48.00	4001.0
49.00	39864.0
50.00	45024.0
51.00	180224.0
52.00	8763.0
52.90	579.0
54.00	155.0
55.00	1495.0
56.00	5432.0
57.00	12202.0
58.00	783.0
58.90	317.0
60.00	454.0
61.00	2266.0
62.00	2253.0
63.00	6951.0
64.00	912.0
65.10	3099.0
65.90	361.0
67.10	395.0
68.00	2992.0
69.00	172736.0
70.00	1231.0
71.00	236.0
72.00	227.0
73.00	2450.0
74.00	17624.0
75.00	27576.0
76.10	9821.0
77.00	200064.0
78.00	13387.0
79.00	12559.0
80.00	10124.0
81.00	14010.0
82.00	3685.0
83.00	4491.0
84.00	26048.0
85.00	3225.0
86.00	20880.0
87.00	2256.0
88.00	3270.0
89.00	612.0
91.00	3164.0
92.00	3893.0
93.00	22168.0
94.00	1589.0
95.00	431.0
96.00	1565.0
97.10	626.0
98.00	18224.0
99.00	14628.0
100.00	1247.0
101.00	7886.0
102.00	524.0
103.00	2678.0
104.00	4613.0
105.00	4986.0
106.00	1461.0
107.00	62824.0
108.00	8797.0
109.10	1808.0
110.00	107328.0
111.00	16488.0
112.00	2162.0
113.00	652.0
114.10	232.0
114.90	451.0
116.00	3503.0
117.00	55920.0
118.00	3896.0
119.00	734.0
120.00	725.0
121.10	317.0
122.00	3956.0
123.00	6825.0
124.00	2925.0
125.00	3018.0
126.10	1161.0
127.00	220224.0
128.00	16440.0
129.00	83104.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

130.00	6983.0
131.00	1450.0
131.90	692.0
133.00	692.0
134.00	2456.0
135.00	7078.0
136.00	2555.0
137.00	3958.0
137.90	733.0
138.90	549.0
140.00	964.0
141.00	11610.0
142.00	3585.0
143.00	2424.0
144.10	686.0
144.90	619.0
146.00	2410.0
147.00	6723.0
148.00	14290.0
149.00	3151.0
149.90	730.0
151.00	1397.0
151.80	910.0
153.00	3697.0
154.00	2930.0
155.00	7660.0
156.00	10013.0
157.00	1563.0
158.00	2492.0
159.10	1541.0
160.00	3820.0
161.00	5851.0
162.00	1903.0
162.90	620.0
163.90	604.0
165.00	4761.0
166.00	3073.0
167.00	24592.0
168.00	13676.0
169.00	2085.0
170.00	861.0
170.90	1077.0
171.90	2196.0
173.00	3126.0
174.00	5173.0
175.00	10424.0
176.00	2884.0
177.00	4278.0
177.90	1566.0
179.00	19296.0
180.00	13167.0
181.00	5467.0
182.10	1042.0
183.00	482.0
184.00	1468.0
185.00	8392.0
186.00	69536.0
187.00	18872.0
188.00	2242.0
189.00	4571.0
190.00	733.0
191.00	2623.0
192.00	6137.0
193.00	7193.0
194.00	1542.0
194.90	707.0
196.00	13902.0
197.00	2246.0
198.00	498112.0
199.00	34520.0
200.00	2605.0
201.40	2051.0
203.00	3791.0
204.00	20320.0
205.00	34064.0
206.00	137600.0
207.10	20952.0
208.00	5033.0
209.00	2221.0
210.00	1906.0
211.00	5499.0
211.80	945.0
212.90	513.0
215.00	1515.0
216.00	3310.0
217.00	39512.0
218.00	4974.0
219.00	390.0
221.00	21408.0
222.00	4526.0
223.00	8777.0
224.00	80120.0
225.00	19856.0
226.00	2191.0
227.00	34456.0
228.00	5268.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

229.00	7077.0
230.00	780.0
231.00	2863.0
232.00	525.0
233.00	547.0
234.00	2276.0
235.00	2369.0
235.90	1758.0
237.00	3094.0
238.00	580.0
239.00	1347.0
240.00	1286.0
241.00	1918.0
242.00	4393.0
243.00	4267.0
244.00	63816.0
245.00	7685.0
246.00	13099.0
247.00	2846.0
247.90	543.0
249.00	2192.0
250.00	373.0
250.90	601.0
251.90	566.0
253.00	1673.0
254.00	2061.0
255.00	331968.0
256.00	46544.0
257.00	3400.0
258.00	19400.0
259.00	2775.0
260.00	453.0
261.00	517.0
262.90	248.0
264.00	490.0
265.00	7672.0
266.00	1033.0
266.90	367.0
269.90	396.0
271.00	640.0
272.00	830.0
273.00	10279.0
274.00	26464.0
275.00	151104.0
276.00	20032.0
277.00	12261.0
278.00	2292.0
278.90	440.0
281.00	1192.0
282.00	656.0
283.00	1698.0
284.00	928.0
285.00	2774.0
286.00	336.0
288.90	609.0
290.00	500.0
291.00	333.0
291.90	468.0
293.00	2742.0
294.00	787.0
295.00	989.0
296.00	46048.0
297.00	6418.0
297.90	331.0
301.00	448.0
301.90	666.0
303.00	4881.0
304.00	1297.0
305.00	193.0
307.90	634.0
309.10	400.0
309.90	461.0
311.00	228.0
313.00	221.0
314.00	1975.0
315.00	4208.0
316.00	2244.0
316.90	434.0
320.10	160.0
321.00	1510.0
322.00	533.0
323.00	12701.0
324.00	2203.0
324.90	167.0
325.90	334.0
327.00	2927.0
328.00	1304.0
329.00	239.0
332.00	1017.0
333.00	1276.0
334.00	9080.0
335.10	2238.0
336.10	283.0
339.00	313.0
340.00	198.0
341.00	1807.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

342.00	338.0
343.10	158.0
346.00	3575.0
347.00	699.0
351.00	242.0
352.00	4303.0
353.00	3085.0
354.10	4079.0
355.00	963.0
359.00	325.0
364.10	205.0
365.00	21600.0
366.00	2645.0
367.00	222.0
370.00	488.0
371.10	1166.0
372.10	7270.0
373.10	1634.0
377.00	231.0
383.00	2170.0
384.00	655.0
385.00	251.0
390.10	992.0
390.90	558.0
391.90	445.0
400.90	481.0
402.00	2836.0
403.00	4162.0
404.00	1430.0
405.10	292.0
421.00	3446.0
422.00	3315.0
423.00	23208.0
424.00	6038.0
425.00	687.0
441.10	58640.0
442.10	395200.0
443.10	76768.0
444.10	6697.0
445.10	360.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

m/z	Abundance
36.00	1197.0
37.00	1817.0
38.00	1737.0
39.00	6793.0
40.00	1152.0
41.00	2255.0
42.00	970.0
43.00	1095.0
44.00	2484.0
45.00	486.0
47.00	7655.0
48.00	4215.0
49.00	40104.0
50.00	24880.0
51.00	98696.0
52.00	4609.0
52.90	383.0
54.00	172.0
55.00	984.0
56.00	2457.0
57.00	6569.0
57.90	335.0
59.00	333.0
59.90	314.0
61.00	1121.0
62.00	1268.0
63.00	3512.0
64.00	408.0
65.00	1839.0
66.10	189.0
67.00	299.0
68.00	1440.0
69.00	88424.0
70.00	777.0
71.00	289.0
72.00	222.0
73.00	1984.0
74.00	9776.0
75.00	15015.0
76.10	4650.0
77.10	102696.0
78.00	7333.0
79.00	6840.0
80.00	5218.0
81.00	7317.0
82.00	2610.0
83.00	2850.0
84.00	25720.0
85.00	2201.0
85.90	18440.0
87.00	1308.0
87.90	3196.0
89.00	189.0
91.00	1635.0
92.00	2214.0
93.00	12221.0
93.90	962.0
95.00	394.0
96.00	754.0
97.00	544.0
98.00	9812.0
99.00	8021.0
100.00	676.0
101.00	4310.0
102.00	243.0
103.00	1725.0
104.00	2721.0
105.00	2629.0
106.00	1249.0
107.00	32584.0
108.00	4902.0
109.00	915.0
110.00	56600.0
111.00	8656.0
112.10	1010.0
113.00	572.0
114.90	256.0
116.10	1943.0
117.00	30752.0
118.00	2249.0
118.90	334.0
119.90	368.0
121.00	301.0
122.00	2383.0
123.00	4059.0
124.00	1670.0
124.90	1530.0
126.00	583.0
127.00	120112.0
128.00	9068.0
129.00	44656.0
130.00	3887.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

131.00	859.0
132.00	477.0
132.90	494.0
134.00	1387.0
135.00	4476.0
135.90	1360.0
137.00	2274.0
137.70	281.0
138.90	356.0
140.00	459.0
141.00	6542.0
142.00	1808.0
142.90	1313.0
143.90	305.0
145.00	491.0
146.00	1179.0
147.00	3745.0
148.00	8620.0
149.00	1676.0
149.90	350.0
151.00	739.0
152.00	381.0
153.00	2136.0
154.00	1521.0
155.00	3589.0
156.00	5430.0
157.00	1242.0
158.00	1192.0
159.00	1039.0
160.00	2105.0
161.00	2956.0
161.90	856.0
163.00	471.0
164.00	420.0
165.00	2749.0
166.00	2035.0
167.00	14810.0
168.00	7762.0
169.00	1239.0
169.90	586.0
170.90	541.0
172.00	1440.0
173.00	1580.0
174.00	2673.0
175.00	5519.0
176.00	1590.0
177.00	2509.0
178.00	1096.0
179.00	10979.0
180.00	7632.0
181.00	3561.0
182.10	463.0
183.00	324.0
184.00	977.0
185.00	5029.0
186.00	38496.0
187.00	11449.0
188.00	938.0
189.00	2782.0
190.00	395.0
191.00	1579.0
192.00	3604.0
193.00	4861.0
194.00	878.0
195.00	476.0
196.00	8206.0
196.90	1608.0
198.00	286656.0
199.00	19472.0
199.90	1420.0
201.50	1354.0
203.00	2140.0
204.10	11697.0
205.00	18520.0
206.00	81632.0
207.10	12238.0
208.00	3116.0
209.00	1193.0
210.10	1205.0
211.00	3327.0
212.00	684.0
213.00	265.0
215.00	824.0
216.00	1704.0
217.00	22952.0
218.00	2785.0
219.00	302.0
219.90	360.0
221.00	13085.0
222.00	2786.0
223.00	5226.0
224.00	46256.0
225.00	12467.0
226.00	1316.0
227.00	20320.0
228.00	2755.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

229.00	4195.0
230.00	481.0
231.00	1537.0
232.00	319.0
232.90	383.0
234.00	1370.0
235.00	1643.0
236.00	1059.0
237.00	1801.0
237.90	339.0
238.90	955.0
240.00	563.0
241.00	1363.0
242.00	2721.0
243.00	2975.0
244.10	37328.0
245.00	5416.0
246.00	7752.0
247.00	1345.0
248.00	380.0
248.90	1334.0
249.90	250.0
251.00	400.0
251.90	321.0
252.90	857.0
254.00	1730.0
255.00	204736.0
256.00	29064.0
257.00	2374.0
258.00	11465.0
259.00	1683.0
260.00	350.0
260.90	306.0
264.00	335.0
265.00	4759.0
266.00	802.0
267.00	273.0
268.00	185.0
269.90	245.0
271.00	450.0
271.90	534.0
273.00	6017.0
274.00	15537.0
275.00	94504.0
276.00	12073.0
277.00	7701.0
278.00	1491.0
279.00	382.0
281.00	927.0
282.00	601.0
283.00	1116.0
284.00	686.0
285.00	1543.0
286.10	311.0
288.90	358.0
289.90	317.0
291.00	218.0
292.00	544.0
293.00	1800.0
293.90	297.0
295.10	621.0
296.00	27448.0
297.00	4352.0
297.90	199.0
301.00	376.0
302.10	400.0
303.00	3196.0
304.00	768.0
307.90	332.0
309.00	344.0
309.80	257.0
312.80	250.0
314.00	1408.0
315.00	3681.0
316.00	1855.0
317.00	269.0
320.10	156.0
320.90	917.0
321.90	334.0
323.00	8560.0
324.00	1842.0
325.10	205.0
325.90	179.0
327.00	1951.0
328.00	953.0
328.90	212.0
332.00	918.0
333.00	887.0
334.00	6543.0
335.00	1691.0
336.10	161.0
338.80	210.0
339.90	165.0
340.90	1083.0
342.00	352.0
346.00	2281.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052

347.00	391.0
350.90	215.0
352.00	3024.0
353.00	1985.0
354.00	2685.0
355.00	809.0
356.10	153.0
358.90	191.0
364.10	165.0
365.00	14325.0
366.00	2293.0
367.00	191.0
369.90	322.0
371.00	763.0
372.00	5043.0
372.90	1237.0
383.00	1502.0
384.00	340.0
389.90	650.0
391.00	558.0
392.10	286.0
401.00	417.0
402.00	2272.0
403.00	2995.0
404.00	971.0
405.00	197.0
421.00	2374.0
422.00	2397.0
423.10	16576.0
424.00	3836.0
425.00	359.0
441.10	45744.0
442.10	304640.0
443.10	55640.0
444.10	4503.0
444.90	314.0

Instrument :
BNA_M
ClientSampleId :
DFTPP052