

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061815\
 Data File : BM001730.D
 Acq On : 17 Jun 2015 22:27
 Operator : TP/IZ
 Sample : SSTD3.209
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.209

Quant Time: Jun 18 04:30:46 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM061815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 04:21:50 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	1082	0.40	ng/ul	0.00
2) Naphthalene-d8	10.46	136	3964	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.32	164	1919	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.06	188	3849	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	3584	0.40	ng/ul	0.00
20) Perylene-d12	23.47	264	2944	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.70	266	3034	3.61	ng/ul	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061815\
Data File : BM001730.D
Acq On : 17 Jun 2015 22:27
Operator : TP/IZ
Sample : SSTD3.209
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.209

Quant Time: Jun 18 04:30:46 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM061815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 18 04:21:50 2015
Response via : Initial Calibration

