

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061815\
 Data File : BM001733.D
 Acq On : 18 Jun 2015 10:30
 Operator : TP/IZ
 Sample : PB84013BL
 Misc : SBLK70
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK70

Quant Time: Jun 18 17:14:01 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM061815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 04:33:02 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	1002	0.40	ng/ul	0.00
2) Naphthalene-d8	10.46	136	3531	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.32	164	1645	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.06	188	3561	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	2731	0.40	ng/ul	0.00
20) Perylene-d12	23.47	264	2410	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.06	152	1395	0.32	ng/ul	0.01
14) Fluoranthene-d10	19.09	212	2507	0.33	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061815\
Data File : BM001733.D
Acq On : 18 Jun 2015 10:30
Operator : TP/IZ
Sample : PB84013BL
Misc : SBLK70
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK70

Quant Time: Jun 18 17:14:01 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM061815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 18 04:33:02 2015
Response via : Initial Calibration

