

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061815\
 Data File : BM001737.D
 Acq On : 18 Jun 2015 13:52
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.411

Quant Time: Jun 18 17:31:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM061815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 04:33:02 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
3	Naphthalene	1.058	1.068	-0.9	112	0.00
4 SURR	2-Methylnaphthalene-d10	0.491	0.494	-0.6	112	0.00
5	2-Methylnaphthalene	0.672	0.676	-0.6	112	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
7	Acenaphthylene	1.929	1.897	1.7	113	0.00
8 C	Acenaphthene	1.468	1.432	2.5	111	0.00
9	Fluorene	1.646	1.608	2.3	112	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
11	Pentachlorophenol	0.086	0.081	5.8	113	-0.02
12	Phenanthrene	1.203	1.224	-1.7	116	-0.02
13	Anthracene	1.095	1.119	-2.2	116	0.00
14 SURR	Fluoranthene-d10	0.848	0.866	-2.1	115	0.00
15 C	Fluoranthene	1.279	1.317	-3.0	117	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
17	Pyrene	1.598	1.575	1.4	116	0.00
18	Benzo(a)anthracene	1.301	1.328	-2.1	121	0.00
19	Chrysene	1.389	1.404	-1.1	119	0.00
20 I	Perylene-d12	1.000	1.000	0.0	122	0.00
21	Benzo(b)fluoranthene	1.548	1.685	-8.9	137	0.00
22	Benzo(k)fluoranthene	1.481	1.327	10.4	107	0.00
23 C	Benzo(a)pyrene	1.388	1.399	-0.8	124	0.00
24	Indeno(1,2,3-cd)pyrene	1.603	1.571	2.0	121	0.00
25	Dibenzo(a,h)anthracene	1.302	1.262	3.1	120	0.00
26	Benzo(g,h,i)perylene	1.408	1.395	0.9	121	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0