

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061816\
 Data File : BM005856.D
 Acq On : 17 Jun 2016 09:47
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Jun 17 11:13:24 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 10:48:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	155#	0.00
2 S	1,4-Dioxane-d8	0.154	0.148	3.9	140	0.00
3 S	Phenol-d5	1.674	1.533	8.4	152#	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.868	0.786	9.4	144	0.00
5 S	2-Chlorophenol-d4	1.396	1.374	1.6	154#	0.00
6 S	4-Methylphenol-d8	1.497	1.388	7.3	149	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	148	0.00
8 S	Nitrobenzene-d5	0.141	0.152	-7.8	154#	0.00
9 S	2-Nitrophenol-d4	0.172	0.187	-8.7	157#	0.00
10 S	2,4-Dichlorophenol-d3	0.341	0.357	-4.7	151#	0.00
11	Naphthalene	1.072	1.055	1.6	148	0.00
12 S	4-Chloroaniline-d4	0.231	0.285	-23.4#	183#	0.00
13	2-Methylnaphthalene	0.817	0.799	2.2	146	0.00
14	1-Methylnaphthalene	0.851	0.810	4.8	143	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	139	0.00
16 S	Dimethylphthalate-d6	1.892	1.745	7.8	130	0.00
17 S	Acenaphthylene-d8	2.180	2.152	1.3	138	0.00
18	Acenaphthylene	2.207	2.171	1.6	140	0.00
19 C	Acenaphthene	1.473	1.438	2.4	139	0.00
20 S	4-Nitrophenol-d4	0.228	0.210	7.9	143	0.00
21 S	Fluorene-d10	1.612	1.523	5.5	136	0.00
22	Fluorene	1.786	1.717	3.9	137	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	132	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.135	0.112	17.0	121	0.00
25	Phenanthrene	1.191	1.169	1.8	132	0.00
26 S	Anthracene-d10	1.044	1.016	2.7	131	0.00
27	Anthracene	1.217	1.198	1.6	132	0.00
28 C	Fluoranthene	1.496	1.414	5.5	128	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
30 S	Pyrene-d10	0.948	1.072	-13.1	124	0.00
31	Pyrene	1.182	1.341	-13.5	125	0.00
32	Benzo(a)anthracene	1.249	1.239	0.8	111	0.00
33	Chrysene	1.203	1.174	2.4	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.245	1.246	-0.1	85	0.00
36	Benzo(k)fluoranthene	1.202	1.250	-4.0	91	0.00
37 S	Benzo(a)pyrene-d12	1.000	0.972	2.8	84	0.00
38 C	Benzo(a)pyrene	1.222	1.184	3.1	82	0.00
39	Indeno(1,2,3-cd)pyrene	1.494	1.334	10.7	79	0.00
40	Dibenzo(a,h)anthracene	1.229	1.131	8.0	79	0.00
41	Benzo(g,h,i)perylene	1.272	1.126	11.5	77	0.01

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061816\
Data File : BM005856.D
Acq On : 17 Jun 2016 09:47
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02053

Quant Time: Jun 17 11:13:24 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061716.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 10:48:37 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0