

Data Path : Z:\HPCHEM1\BNA M\DATA\BM061816\
 Data File : BM005872.D
 Acq On : 17 Jun 2016 21:47
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02058

Quant Time: Jun 18 23:57:55 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM061716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 10:48:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	155#	0.00
2 S	1,4-Dioxane-d8	0.154	0.143	7.1	136	0.00
3 S	Phenol-d5	1.674	1.545	7.7	153#	0.02
4 S	Bis-(2-Chloroethyl)ether-d8	0.868	0.771	11.2	141	0.00
5 S	2-Chlorophenol-d4	1.396	1.396	0.0	157#	0.01
6 S	4-Methylphenol-d8	1.497	1.394	6.9	150	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	146	0.01
8 S	Nitrobenzene-d5	0.141	0.155	-9.9	155#	0.00
9 S	2-Nitrophenol-d4	0.172	0.184	-7.0	152#	0.01
10 S	2,4-Dichlorophenol-d3	0.341	0.362	-6.2	151#	0.01
11	Naphthalene	1.072	1.059	1.2	147	0.01
12 S	4-Chloroaniline-d4	0.231	0.346	-49.8#	219#	0.00
13	2-Methylnaphthalene	0.817	0.801	2.0	144	0.01
14	1-Methylnaphthalene	0.851	0.793	6.8	138	0.01
15 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.01
16 S	Dimethylphthalate-d6	1.892	1.770	6.4	124	0.00
17 S	Acenaphthylene-d8	2.180	2.187	-0.3	131	0.01
18	Acenaphthylene	2.207	2.193	0.6	133	0.01
19 C	Acenaphthene	1.473	1.464	0.6	133	0.01
20 S	4-Nitrophenol-d4	0.228	0.188	17.5	120	0.03
21 S	Fluorene-d10	1.612	1.537	4.7	128	0.01
22	Fluorene	1.786	1.731	3.1	129	0.02
23 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.01
24 S	4,6-Dinitro-2-methylphenol-	0.135	0.069	48.9#	68	0.02
25	Phenanthrene	1.191	1.168	1.9	121	0.01
26 S	Anthracene-d10	1.044	1.018	2.5	120	0.01
27	Anthracene	1.217	1.198	1.6	121	0.01
28 C	Fluoranthene	1.496	1.292	13.6	107	0.01
29 I	Chrysene-d12	1.000	1.000	0.0	85	0.01
30 S	Pyrene-d10	0.948	1.132	-19.4	102	0.01
31	Pyrene	1.182	1.412	-19.5	102	0.01
32	Benzo(a)anthracene	1.249	1.253	-0.3	87	0.02
33	Chrysene	1.203	1.175	2.3	83	0.02
34 I	Perylene-d12	1.000	1.000	0.0	69	0.02
35	Benzo(b)fluoranthene	1.245	1.319	-5.9	73	0.02
36	Benzo(k)fluoranthene	1.202	1.198	0.3	71	0.02
37 S	Benzo(a)pyrene-d12	1.000	1.004	-0.4	71	0.02
38 C	Benzo(a)pyrene	1.222	1.201	1.7	68	0.02
39	Indeno(1,2,3-cd)pyrene	1.494	1.451	2.9	69	0.04
40	Dibenzo(a,h)anthracene	1.229	1.203	2.1	69	0.03
41	Benzo(g,h,i)perylene	1.272	1.185	6.8	66	0.05

Data Path : Z:\HPCHEM1\BNA M\DATA\BM061816\
Data File : BM005872.D
Acq On : 17 Jun 2016 21:47
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02058

Quant Time: Jun 18 23:57:55 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM061716.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 10:48:37 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0