

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061915\
 Data File : BM001741.D
 Acq On : 18 Jun 2015 17:48
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

apatel
 6/22/2015 2:45:04 PM

Quant Time: Jun 19 04:25:41 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 04:18:06 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	15446	5.00	ng	0.00
6) Naphthalene-d8	10.46	136	57144	5.00	ng	0.00
12) Acenaphthene-d10	14.32	164	28439	5.00	ng	0.00
18) Phenanthrene-d10	17.06	188	72806	5.00	ng	0.00
25) Chrysene-d12	21.25	240	49751	5.00	ng	0.00
32) Perylene-d12	23.47	264	44912	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.25	112	17329	4.72	ng	0.00
5) Phenol-d6	6.83	99	22819	4.94	ng	0.00
7) Nitrobenzene-d5	8.83	82	22066	5.01	ng	0.00
13) 2,4,6-Tribromophenol	15.80	330	4993	4.95	ng	0.00
14) 2-Fluorobiphenyl	12.94	172	46019	4.59	ng	0.00
27) Terphenyl-d14	19.70	244	34664	4.64	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.18	88	7944	4.08	ng	95
3) n-Nitrosodimethylamine	3.48	42	13231	4.89	ng	97
8) Nitrobenzene	8.87	77	24578	5.16	ng	99
9) Naphthalene	10.49	128	63633	4.56	ng	99
10) Hexachlorobutadiene	10.78	225	10396	4.48	ng	99
11) 2-Methylnaphthalene	12.12	142	41696	4.64	ng	99
15) Acenaphthylene	14.03	152	68537	5.13	ng	100
16) Acenaphthene	14.37	154	39659	4.66	ng	98
17) Fluorene	15.36	166	36590	4.24	ng	99
19) 4-Bromophenyl-phenylether	16.24	248	14040	2.15	ng	94
20) Hexachlorobenzene	16.38	284	9616	1.59	ng	# 100
21) Pentachlorophenol	16.72	266	5472	3.20	ng	96
22) Phenanthrene	17.09	178	110123	4.56	ng	98
23) Anthracene	17.19	178	46122m	1.77	ng	
24) Fluoranthene	19.13	202	76952	2.39	ng	99
26) Pyrene	19.49	202	82545	4.84	ng	99
28) Benzo(a)anthracene	21.24	228	72646	4.66	ng	99
29) Chrysene	21.28	228	71584m	4.74	ng	
30) Bis(2-ethylhexyl)phthalate	21.16	149	46962	4.86	ng	100
31) Indeno(1,2,3-cd)pyrene	25.71	276	73764	6.00	ng	100
33) Benzo(b)fluoranthene	22.80	252	69068	4.68	ng	98
34) Benzo(k)fluoranthene	22.84	252	64722	4.58	ng	97
35) Benzo(a)pyrene	23.38	252	60939	4.96	ng	97
36) Dibenzo(a,h)anthracene	25.72	278	59109	5.57	ng	97
37) Benzo(g,h,i)perylene	26.39	276	60323	5.28	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM061915\
Data File : BM001741.D
Acq On : 18 Jun 2015 17:48
Operator : TP/IZ
Sample : SSTDIC005
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Manual Integrations
APPROVED

apatel
6/22/2015 2:45:04 PM

Quant Time: Jun 19 04:25:41 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 19 04:18:06 2015
Response via : Initial Calibration

