

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061915\
 Data File : BM001748.D
 Acq On : 18 Jun 2015 21:59
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM061915

Quant Time: Jun 19 05:13:28 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 05:05:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	120	0.00
2	1,4-Dioxane	1.000	0.974	2.6	113	0.00
3	n-Nitrosodimethylamine	1.000	0.981	1.9	117	-0.02
4 S	2-Fluorophenol	1.000	0.943	5.7	113	0.00
5 S	Phenol-d6	1.000	0.929	7.1	114	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	120	0.00
7 S	Nitrobenzene-d5	1.000	0.930	7.0	114	0.00
8	Nitrobenzene	1.000	0.937	6.3	114	0.00
9	Naphthalene	1.000	0.942	5.8	114	0.00
10	Hexachlorobutadiene	1.000	0.949	5.1	117	0.00
11	2-Methylnaphthalene	1.000	0.950	5.0	115	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	122	-0.01
13 S	2,4,6-Tribromophenol	1.000	0.972	2.8	122	0.00
14 S	2-Fluorobiphenyl	1.000	0.942	5.8	115	0.00
15	Acenaphthylene	1.000	0.936	6.4	116	0.00
16	Acenaphthene	1.000	0.939	6.1	116	0.00
17	Fluorene	1.000	0.994	0.6	109	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	114	0.00
19	4-Bromophenyl-phenylether	1.000	1.013	-1.3	128	0.00
20	Hexachlorobenzene	1.000	0.929	7.1	114	0.00
21	Pentachlorophenol	1.000	0.933	6.7	115	0.00
22	Phenanthrene	1.000	0.904	9.6	109	0.00
23	Anthracene	1.000	0.910	9.0	103	0.00
24	Fluoranthene	1.000	0.963	3.7	114	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	117	0.00
26	Pyrene	1.000	0.931	6.9	115	-0.01
27 S	Terphenyl-d14	1.000	0.945	5.5	115	0.00
28	Benzo(a)anthracene	1.000	0.935	6.5	113	0.00
29	Chrysene	1.000	0.964	3.6	115	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.924	7.6	118	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.937	6.3	110	0.00
32 I	Perylene-d12	5.000	5.000	0.0	117	0.00
33	Benzo(b)fluoranthene	1.000	0.974	2.6	115	-0.01
34	Benzo(k)fluoranthene	1.000	0.904	9.6	108	0.00
35 C	Benzo(a)pyrene	1.000	0.941	5.9	112	0.00
36	Dibenzo(a,h)anthracene	1.000	0.940	6.0	110	0.00
37	Benzo(g,h,i)perylene	1.000	0.938	6.2	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0