

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061915\
 Data File : BM001754.D
 Acq On : 19 Jun 2015 10:15
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 19 12:24:30 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 05:05:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.570	0.575	-0.9	107	0.00
3	n-Nitrosodimethylamine	0.880	0.909	-3.3	113	-0.02
4 S	2-Fluorophenol	1.169	1.174	-0.4	110	0.00
5 S	Phenol-d6	1.505	1.495	0.7	111	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	110	-0.02
7 S	Nitrobenzene-d5	0.390	0.386	1.0	111	0.00
8	Nitrobenzene	0.420	0.419	0.2	111	0.00
9	Naphthalene	1.222	1.209	1.1	109	0.00
10	Hexachlorobutadiene	0.206	0.204	1.0	112	0.00
11	2-Methylnaphthalene	0.778	0.769	1.2	110	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	109	-0.01
13 S	2,4,6-Tribromophenol	0.156	0.170	-9.0	122	0.00
14 S	2-Fluorobiphenyl	1.787	1.793	-0.3	109	0.00
15	Acenaphthylene	2.376	2.357	0.8	110	0.00
16	Acenaphthene	1.489	1.480	0.6	110	0.00
17	Fluorene	1.276	1.297	-1.6	99	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
19	4-Bromophenyl-phenylether	0.205	0.237	-15.6	124	0.00
20	Hexachlorobenzene	0.141	0.135	4.3	100	0.00
21	Pentachlorophenol	0.063	0.065	-3.2	109	0.00
22	Phenanthrene	1.713	1.690	1.3	101	0.00
23	Anthracene	0.724	0.742	-2.5	100	0.00
24	Fluoranthene	1.080	1.175	-8.8	110	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
26	Pyrene	1.741	1.680	3.5	111	-0.01
27 S	Terphenyl-d14	0.737	0.729	1.1	112	0.00
28	Benzo(a)anthracene	1.507	1.468	2.6	109	0.00
29	Chrysene	1.584	1.546	2.4	108	0.00
30	Bis(2-ethylhexyl)phthalate	0.869	0.824	5.2	113	0.00
31	Indeno(1,2,3-cd)pyrene	1.523	1.494	1.9	107	0.00
32 I	Perylene-d12	1.000	1.000	0.0	109	0.00
33	Benzo(b)fluoranthene	1.641	1.656	-0.9	111	-0.01
34	Benzo(k)fluoranthene	1.469	1.413	3.8	107	0.00
35 C	Benzo(a)pyrene	1.382	1.363	1.4	109	0.00
36	Dibenzo(a,h)anthracene	1.368	1.337	2.3	106	0.00
37	Benzo(g,h,i)perylene	1.439	1.408	2.2	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0