

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061915\
 Data File : BM001768.D
 Acq On : 19 Jun 2015 18:51
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 20 02:29:13 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM061915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 05:05:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.570	0.568	0.4	119	0.00
3	n-Nitrosodimethylamine	0.880	0.893	-1.5	125	-0.02
4 S	2-Fluorophenol	1.169	1.170	-0.1	124	0.00
5 S	Phenol-d6	1.505	1.487	1.2	124	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
7 S	Nitrobenzene-d5	0.390	0.385	1.3	126	0.00
8	Nitrobenzene	0.420	0.417	0.7	126	0.00
9	Naphthalene	1.222	1.212	0.8	125	0.00
10	Hexachlorobutadiene	0.206	0.202	1.9	126	0.00
11	2-Methylnaphthalene	0.778	0.783	-0.6	127	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
13 S	2,4,6-Tribromophenol	0.156	0.163	-4.5	137	0.00
14 S	2-Fluorobiphenyl	1.787	1.800	-0.7	129	0.00
15	Acenaphthylene	2.376	2.347	1.2	128	0.00
16	Acenaphthene	1.489	1.462	1.8	127	0.00
17	Fluorene	1.276	1.341	-5.1	120	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
19	4-Bromophenyl-phenylether	0.205	0.191	6.8	125	0.00
20	Hexachlorobenzene	0.141	0.132	6.4	122	0.00
21	Pentachlorophenol	0.063	0.060	4.8	125	0.00
22	Phenanthrene	1.713	1.494	12.8	112	0.00
23	Anthracene	0.724	0.709	2.1	119	0.00
24	Fluoranthene	1.080	0.936	13.3	110	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
26	Pyrene	1.741	1.706	2.0	110	0.00
27 S	Terphenyl-d14	0.737	0.751	-1.9	112	0.00
28	Benzo(a)anthracene	1.507	1.371	9.0	99	0.00
29	Chrysene	1.584	1.527	3.6	104	0.00
30	Bis(2-ethylhexyl)phthalate	0.869	0.755	13.1	100	0.00
31	Indeno(1,2,3-cd)pyrene	1.523	1.414	7.2	99	0.00
32 I	Perylene-d12	1.000	1.000	0.0	101	0.00
33	Benzo(b)fluoranthene	1.641	1.630	0.7	101	0.00
34	Benzo(k)fluoranthene	1.469	1.418	3.5	100	0.00
35 C	Benzo(a)pyrene	1.382	1.339	3.1	99	0.00
36	Dibenzo(a,h)anthracene	1.368	1.338	2.2	99	0.00
37	Benzo(g,h,i)perylene	1.439	1.402	2.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0