

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062022\
 Data File : BM035596.D
 Acq On : 22 Jun 2022 06:40
 Operator : CG/JU
 Sample : PB145517BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK517

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/22/2022
 Supervised By :Yogesh Patel 06/25/2022

Quant Time: Jun 22 07:18:35 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM061822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 20 01:56:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.820	152	5958	0.400	ng/u1	-0.04
4) Naphthalene-d8	10.605	136	19893	0.400	ng/u1	-0.04
9) Acenaphthene-d10	14.450	164	11368m	0.400	ng/u1	-0.02
13) Phenanthrene-d10	17.199	188	20856m	0.400	ng/u1	-0.03
17) Chrysene-d12	21.412	240	13834	0.400	ng/u1	0.00
23) Perylene-d12	23.733	264	16845	0.400	ng/u1	#-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	42977	5.127	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.244	152	8620	0.292	ng/u1	0.00
18) Fluoranthene-d10	19.233	212	14596	0.314	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062022\
Data File : BM035596.D
Acq On : 22 Jun 2022 06:40
Operator : CG/JU
Sample : PB145517BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK517

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 06/22/2022
Supervised By :Yogesh Patel 06/25/2022

Quant Time: Jun 22 07:18:35 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM061822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 20 01:56:13 2022
Response via : Initial Calibration

