

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062117\
 Data File : BM010617.D
 Acq On : 21 Jun 2017 11:14
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02040

Quant Time: Jun 22 00:41:38 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 21 01:33:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	62	0.00
2	1,4-Dioxane	0.390	0.352	9.7	51	0.00
3 S	1,4-Dioxane-d8	0.349	0.368	-5.4	61	0.00
4	Benzaldehyde	1.121	1.184	-5.6	54	0.00
5 S	Phenol-d5	1.905	1.603	15.9	52	0.00
6	Phenol	1.962	1.656	15.6	53	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.088	0.899	17.4	49	0.00
8	Bis(2-Chloroethyl)ether	1.429	1.223	14.4	52	0.00
9 S	2-Chlorophenol-d4	1.337	1.226	8.3	56	0.00
10	2-Chlorophenol	1.336	1.251	6.4	56	0.00
11	2-Methylphenol	1.496	1.320	11.8	54	0.00
12	2,2'-oxybis(1-Chloropropane	0.997	0.853	14.4	51	0.00
13 S	4-Methylphenol-d8	1.580	1.351	14.5	52	0.00
14	Acetophenone	2.579	2.275	11.8	54	0.00
15 P	N-Nitroso-di-n-propylamine	1.390	1.166	16.1	50	0.00
16	4-Methylphenol	1.646	1.426	13.4	53	0.00
17	Hexachloroethane	0.594	0.571	3.9	58	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	59	0.00
19 S	Nitrobenzene-d5	0.143	0.144	-0.7	59	0.00
20	Nitrobenzene	0.416	0.406	2.4	55	0.00
21	Isophorone	0.828	0.741	10.5	50	0.00
22 S	2-Nitrophenol-d4	0.164	0.170	-3.7	60	0.00
23 C	2-Nitrophenol	0.174	0.180	-3.4	58	0.00
24	2,4-Dimethylphenol	0.443	0.429	3.2	56	0.00
25	Bis(2-Chloroethoxy)methane	0.459	0.419	8.7	51	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.344	0.6	57	0.00
27 C	2,4-Dichlorophenol	0.337	0.352	-4.5	60	0.00
28	Naphthalene	0.995	0.986	0.9	57	0.00
29 S	4-Chloroaniline-d4	0.365	0.413	-13.2	56	0.00
30	4-Chloroaniline	0.367	0.396	-7.9	55	0.00
31 C	Hexachlorobutadiene	0.254	0.278	-9.4	62	0.00
32	Caprolactam	0.118	0.109	7.6	50	0.00
33 C	4-Chloro-3-methylphenol	0.430	0.425	1.2	55	0.00
34	2-Methylnaphthalene	0.800	0.813	-1.6	58	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	60	0.00
36	1,2,4,5-Tetrachlorobenzene	0.706	0.716	-1.4	60	0.00
37	Hexachlorocyclopentadiene	0.406	0.428	-5.4	67	0.00
38 C	2,4,6-Trichlorophenol	0.482	0.476	1.2	59	0.00
39	2,4,5-Trichlorophenol	0.498	0.493	1.0	59	0.00
40	1,1'-Biphenyl	1.564	1.528	2.3	59	0.00
41	2-Chloronaphthalene	1.230	1.231	-0.1	60	0.00
42	2-Nitroaniline	0.360	0.349	3.1	58	0.00
43 S	Dimethylphthalate-d6	1.756	1.777	-1.2	60	0.00
44	Dimethylphthalate	1.721	1.713	0.5	59	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062117\
 Data File : BM010617.D
 Acq On : 21 Jun 2017 11:14
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02040

Quant Time: Jun 22 00:41:38 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 21 01:33:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.298	0.318	-6.7	64	0.00
46 S	Acenaphthylene-d8	2.042	2.044	-0.1	60	0.00
47	Acenaphthylene	1.945	1.875	3.6	58	0.00
48	3-Nitroaniline	0.301	0.301	0.0	59	0.00
49 C	Acenaphthene	1.312	1.301	0.8	59	0.00
50	2,4-Dinitrophenol	0.216	0.223	-3.2	71	0.00
51 S	4-Nitrophenol-d4	0.309	0.295	4.5	59	0.00
52	4-Nitrophenol	0.392	0.458	-16.8	70	0.00
53	Dibenzofuran	1.987	1.998	-0.6	60	0.00
54	2,4-Dinitrotoluene	0.462	0.487	-5.4	63	0.00
55	2,3,4,6-Tetrachlorophenol	0.504	0.512	-1.6	60	0.00
56	Diethylphthalate	1.811	1.783	1.5	59	0.00
57 S	Fluorene-d10	1.518	1.557	-2.6	61	0.00
58	Fluorene	1.664	1.688	-1.4	61	0.00
59	4-Chlorophenyl-phenylether	0.904	0.926	-2.4	62	0.00
60	4-Nitroaniline	0.355	0.358	-0.8	63	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.123	0.120	2.4	67	0.00
63	4,6-Dinitro-2-methylphenol	0.131	0.129	1.5	68	0.00
64	N-Nitrosodiphenylamine	0.555	0.543	2.2	61	0.00
65	4-Bromophenyl-phenylether	0.229	0.223	2.6	62	0.00
66	Hexachlorobenzene	0.241	0.238	1.2	63	0.00
67	Atrazine	0.234	0.242	-3.4	64	0.00
68 C	Pentachlorophenol	0.170	0.157	7.6	59	0.00
69	Phenanthrene	1.063	1.052	1.0	62	0.00
70 S	Anthracene-d10	0.966	0.964	0.2	62	0.00
71	Anthracene	1.095	1.077	1.6	61	0.00
72	Carbazole	0.956	0.946	1.0	62	0.00
73	Di-n-butylphthalate	1.215	1.180	2.9	60	0.00
74 C	Fluoranthene	1.370	1.373	-0.2	63	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	68	0.00
76 S	Pyrene-d10	0.929	0.916	1.4	64	0.00
77	Pyrene	1.155	1.132	2.0	64	0.00
78	Butylbenzylphthalate	0.434	0.425	2.1	64	0.00
79	3,3'-Dichlorobenzidine	0.400	0.410	-2.5	66	0.00
80	Benzo(a)anthracene	1.165	1.178	-1.1	67	0.00
81	Bis(2-ethylhexyl)phthalate	0.663	0.665	-0.3	66	0.00
82	Chrysene	1.038	1.062	-2.3	67	0.00
83 I	Perylene-d12	1.000	1.000	0.0	71	0.00
84	Di-n-octyl phthalate	1.398	1.166	16.6	65	0.00
85	Benzo(b)fluoranthene	1.298	1.302	-0.3	72	0.00
86	Benzo(k)fluoranthene	1.231	1.177	4.4	68	0.00
87 S	Benzo(a)pyrene-d12	0.967	0.955	1.2	70	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062117\
Data File : BM010617.D
Acq On : 21 Jun 2017 11:14
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02040

Quant Time: Jun 22 00:41:38 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 21 01:33:50 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.195	1.166	2.4	69	0.00
89	Indeno(1,2,3-cd)pyrene	1.240	1.284	-3.5	72	0.00
90	Dibenzo(a,h)anthracene	1.033	1.061	-2.7	71	0.00
91	Benzo(g,h,i)perylene	0.987	1.038	-5.2	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0