

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062117\
 Data File : BM010632.D
 Acq On : 21 Jun 2017 20:40
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02041

Quant Time: Jun 22 01:36:51 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 00:43:46 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	62	0.00
2	1,4-Dioxane	0.390	0.392	-0.5	57	0.00
3 S	1,4-Dioxane-d8	0.349	0.378	-8.3	63	0.00
4	Benzaldehyde	1.121	1.249	-11.4	57	0.00
5 S	Phenol-d5	1.905	1.744	8.5	57	0.00
6	Phenol	1.962	1.763	10.1	57	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.088	0.934	14.2	51	0.00
8	Bis(2-Chloroethyl)ether	1.429	1.285	10.1	55	0.00
9 S	2-Chlorophenol-d4	1.337	1.320	1.3	61	0.00
10	2-Chlorophenol	1.336	1.312	1.8	59	0.00
11	2-Methylphenol	1.496	1.405	6.1	58	0.00
12	2,2'-oxybis(1-Chloropropane	0.997	0.852	14.5	51	0.00
13 S	4-Methylphenol-d8	1.580	1.501	5.0	59	0.00
14	Acetophenone	2.579	2.397	7.1	57	0.00
15 P	N-Nitroso-di-n-propylamine	1.390	1.258	9.5	54	0.00
16	4-Methylphenol	1.646	1.537	6.6	58	0.00
17	Hexachloroethane	0.594	0.596	-0.3	61	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
19 S	Nitrobenzene-d5	0.143	0.139	2.8	61	0.00
20	Nitrobenzene	0.416	0.409	1.7	59	0.00
21	Isophorone	0.828	0.757	8.6	54	0.00
22 S	2-Nitrophenol-d4	0.164	0.172	-4.9	64	0.00
23 C	2-Nitrophenol	0.174	0.182	-4.6	62	0.00
24	2,4-Dimethylphenol	0.443	0.435	1.8	60	0.00
25	Bis(2-Chloroethoxy)methane	0.459	0.416	9.4	54	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.348	-0.6	61	0.00
27 C	2,4-Dichlorophenol	0.337	0.337	0.0	61	0.00
28	Naphthalene	0.995	0.980	1.5	60	0.00
29 S	4-Chloroaniline-d4	0.365	0.423	-15.9	61	0.00
30	4-Chloroaniline	0.367	0.398	-8.4	58	0.00
31 C	Hexachlorobutadiene	0.254	0.278	-9.4	66	0.00
32	Caprolactam	0.118	0.113	4.2	55	0.00
33 C	4-Chloro-3-methylphenol	0.430	0.419	2.6	58	0.00
34	2-Methylnaphthalene	0.800	0.804	-0.5	61	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	64	0.00
36	1,2,4,5-Tetrachlorobenzene	0.706	0.690	2.3	62	0.00
37	Hexachlorocyclopentadiene	0.406	0.415	-2.2	70	0.00
38 C	2,4,6-Trichlorophenol	0.482	0.472	2.1	63	0.00
39	2,4,5-Trichlorophenol	0.498	0.497	0.2	64	0.00
40	1,1'-Biphenyl	1.564	1.519	2.9	63	0.00
41	2-Chloronaphthalene	1.230	1.176	4.4	62	0.00
42	2-Nitroaniline	0.360	0.365	-1.4	65	0.00
43 S	Dimethylphthalate-d6	1.756	1.751	0.3	63	0.00
44	Dimethylphthalate	1.721	1.685	2.1	62	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062117\
 Data File : BM010632.D
 Acq On : 21 Jun 2017 20:40
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02041

Quant Time: Jun 22 01:36:51 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 00:43:46 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.298	0.314	-5.4	68	0.00
46 S	Acenaphthylene-d8	2.042	1.942	4.9	61	0.00
47	Acenaphthylene	1.945	1.871	3.8	62	0.00
48	3-Nitroaniline	0.301	0.302	-0.3	64	0.00
49 C	Acenaphthene	1.312	1.251	4.6	61	0.00
50	2,4-Dinitrophenol	0.216	0.225	-4.2	77	0.00
51 S	4-Nitrophenol-d4	0.309	0.301	2.6	65	0.00
52	4-Nitrophenol	0.392	0.434	-10.7	71	0.00
53	Dibenzofuran	1.987	1.947	2.0	62	0.00
54	2,4-Dinitrotoluene	0.462	0.484	-4.8	67	0.00
55	2,3,4,6-Tetrachlorophenol	0.504	0.500	0.8	63	0.00
56	Diethylphthalate	1.811	1.768	2.4	62	0.00
57 S	Fluorene-d10	1.518	1.497	1.4	63	0.00
58	Fluorene	1.664	1.603	3.7	62	0.00
59	4-Chlorophenyl-phenylether	0.904	0.927	-2.5	66	0.00
60	4-Nitroaniline	0.355	0.334	5.9	63	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.123	0.125	-1.6	72	0.00
63	4,6-Dinitro-2-methylphenol	0.131	0.131	0.0	71	0.00
64	N-Nitrosodiphenylamine	0.555	0.541	2.5	63	0.00
65	4-Bromophenyl-phenylether	0.229	0.220	3.9	63	0.00
66	Hexachlorobenzene	0.241	0.239	0.8	65	0.00
67	Atrazine	0.234	0.247	-5.6	67	0.00
68 C	Pentachlorophenol	0.170	0.166	2.4	65	0.00
69	Phenanthrene	1.063	1.044	1.8	64	0.00
70 S	Anthracene-d10	0.966	0.958	0.8	64	0.00
71	Anthracene	1.095	1.075	1.8	63	0.00
72	Carbazole	0.956	0.925	3.2	63	0.00
73	Di-n-butylphthalate	1.215	1.159	4.6	61	0.00
74 C	Fluoranthene	1.370	1.367	0.2	65	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
76 S	Pyrene-d10	0.929	0.937	-0.9	65	0.00
77	Pyrene	1.155	1.148	0.6	64	0.00
78	Butylbenzylphthalate	0.434	0.448	-3.2	67	0.00
79	3,3'-Dichlorobenzidine	0.400	0.407	-1.7	65	0.00
80	Benzo(a)anthracene	1.165	1.194	-2.5	67	0.00
81	Bis(2-ethylhexyl)phthalate	0.663	0.647	2.4	63	0.00
82	Chrysene	1.038	1.057	-1.8	66	0.00
83 I	Perylene-d12	1.000	1.000	0.0	70	0.00
84	Di-n-octyl phthalate	1.398	1.195	14.5	66	0.00
85	Benzo(b)fluoranthene	1.298	1.254	3.4	68	0.00
86	Benzo(k)fluoranthene	1.231	1.229	0.2	70	0.00
87 S	Benzo(a)pyrene-d12	0.967	0.950	1.8	69	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062117\
Data File : BM010632.D
Acq On : 21 Jun 2017 20:40
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02041

Quant Time: Jun 22 01:36:51 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 22 00:43:46 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.195	1.184	0.9	70	0.00
89	Indeno(1,2,3-cd)pyrene	1.240	1.268	-2.3	71	0.00
90	Dibenzo(a,h)anthracene	1.033	1.040	-0.7	69	0.00
91	Benzo(g,h,i)perylene	0.987	1.017	-3.0	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0