

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062217\

Quantitation Report (Qedit)

Data File : BM010677.D

Acq On : 23 Jun 2017 01:32

Operator : SJ/JU

Sample : I3653-04 5X

Misc :

ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 05:10:15 2017

Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Jun 23 05:06:10 2017

Response via : Initial Calibration

Instrument :




BNA_M

ClientSampleId :

F5B21

Manual Integrations
APPROVED

mohammad
6/26/2017 8:26:25 AM



(29) 4-Chloroaniline-d4 (S)

10.363min (-0.006) 4.06ng/ul

response 27438

Ion	Exp%	Act%
131.00	100	100
133.00	31.40	18.16#
69.00	24.30	16.74#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062217\

Quantitation Report (Qedit)

Data File : BM010677.D

Acq On : 23 Jun 2017 01:32

Operator : SJ/JU

Sample : I3653-04 5X

Misc :

ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 05:10:15 2017

Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Jun 23 05:06:10 2017

Response via : Initial Calibration

Instrument :

BNA_M

ClientSampleId :


F5B21

Manual Integrations

APPROVED

mohammad

6/26/2017 8:26:25 AM



(29) 4-Chloroaniline-d4 (S)

10.416min (+0.047) 2.79ng/ul m

response 18824

Ion Exp% Act%

131.00 100 100

133.00 31.40 30.15

69.00 24.30 72.72#

0.00 0.00 0.00

S-J
06/30/2017

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062217\

Quantitation Report (Qedit)

Data File : BM010677.D

Acq On : 23 Jun 2017 01:32

Operator : SJ/JU

Sample : I3653-04 5X

Misc :

ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 05:10:15 2017

Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Jun 23 05:06:10 2017

Response via : Initial Calibration

Instrument :




BNA_M

ClientSampleId :

F5B21

Manual Integrations
APPROVED

mohammad
6/26/2017 8:26:25 AM



(86) Benzo(k)fluoranthene

22.592min (-0.041) 26.07ng/ul

response 885053

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	22.46
125.00	4.30	5.89#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062217\
Quantitation Report (Qedit)




Data File : BM010677.D
Acq On : 23 Jun 2017 01:32
Operator : SJ/JU
Sample : I3653-04 5X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 05:10:15 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 23 05:06:10 2017
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
F5B21

Manual Integrations
APPROVED

mohammad
6/26/2017 8:26:25 AM



(86) Benzo(k)fluoranthene

22.627min (-0.006) 10.89ng/ul m

response 369889

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	22.12
125.00	4.30	4.81
0.00	0.00	0.00

> S-J
06/30/2017

Quantitation Report (Qedit)


Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
 Data File : BM010677.D
 Acq On : 23 Jun 2017 01:32
 Operator : SJ/JU
 Sample : I3653-04 5X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 06:42:39 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 23 05:06:10 2017
 Response via : Initial Calibration

Instrument :
 BNA_M
 ClientSampleId :
 F5B21

Manual Integrations
 APPROVED

mohammad
 6/26/2017 8:26:25 AM



(24) 2,4-Dimethylphenol

9.422min (-0.000) 4.48ng/ul

response 36837

Ion	Exp%	Act%
107.00	100	100
121.00	39.90	48.01#
122.00	72.20	82.88
0.00	0.00	0.00

Quantitation Report (Qedit)


Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
 Data File : BM010677.D
 Acq On : 23 Jun 2017 01:32
 Operator : SJ/JU
 Sample : I3653-04 5X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 06:42:39 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 23 05:06:10 2017
 Response via : Initial Calibration

Instrument :
 BNA_M
 ClientSampleId :
 F5B21

Manual Integrations
 APPROVED

mohammad
 6/26/2017 8:26:25 AM



(24) 2,4-Dimethylphenol

9.422min (-0.000) 6.32ng/ul m

response 51898

Ion	Exp%	Act%
107.00	100	100
121.00	39.90	48.01#
122.00	72.20	82.88
0.00	0.00	0.00

S.J
 06/30/2017

Quantitation Report (Qedit)




Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
 Data File : BM010677.D
 Acq On : 23 Jun 2017 01:32
 Operator : SJ/JU
 Sample : I3653-04 5X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 06:42:39 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 23 05:06:10 2017
 Response via : Initial Calibration

Instrument :
 BNA_M
 ClientSampled :
 F5B21

Manual Integrations
 APPROVED

mohammad
 6/26/2017 8:26:25 AM



(29) 4-Chloroaniline-d4 (S)

10.416min (+0.047) 2.79ng/ul m

response 18824

Ion	Exp%	Act%
131.00	100	100
133.00	31.40	30.15
69.00	24.30	72.72#
0.00	0.00	0.00

> S.J
 06/30/2017

Quantitation Report (Qedit)


Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
 Data File : BM010677.D
 Acq On : 23 Jun 2017 01:32
 Operator : SJ/JU
 Sample : I3653-04 5X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F5B21

Quant Time: Jun 23 06:42:39 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 23 05:06:10 2017
 Response via : Initial Calibration

Manual Integrations
 APPROVED

mohammad
 6/26/2017 8:26:25 AM



(29) 4-Chloroaniline-d4 (S)

10.363min (-0.006) 7.16ng/ul m

response 48361

Ion Exp% Act%

131.00 100 100

133.00 31.40 18.16#

69.00 24.30 16.74#

0.00 0.00 0.00

SJ
 06/30/2017

Quantitation Report

(LSC Reviewed)

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
 Data File : BM010677.D
 Acq On : 23 Jun 2017 01:32
 Operator : SJ/JU
 Sample : I3653-04 5X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :

BNA_M

Client Sampled :

F5B21

Quant Time: Jun 23 13:22:35 2017

Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Jun 23 05:06:10 2017

Response via : Initial Calibration

Manual Integrations
APPROVEDmohammad
6/26/2017 8:26:25 AM

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Dichlorobenzene-d4	7.46	152	85704	20.00	ng/ul	0.00
18) Naphthalene-d8	10.23	136	370551	20.00	ng/ul	0.01
35) Acenaphthene-d10	14.12	164	254923	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.87	188	600751	20.00	ng/ul	0.00
75) Chrysene-d12	21.09	240	693880	20.00	ng/ul	0.00
83) Perylene-d12	23.23	264	551704	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.07	96	727	0.49	ng/uL	0.00
5) Phenol-d5	6.65	99	20538	2.52	ng/ul	0.00
7) Bis-(2-Chloroethyl) ether-d	6.81	67	11837	2.54	ng/ul	0.00
9) 2-Chlorophenol-d4	7.00	132	15653	2.73	ng/ul	0.00
13) 4-Methylphenol-d8	8.16	113	16954	2.50	ng/ul	0.00
19) Nitrobenzene-d5	8.61	128	9550	3.61	ng/ul	0.00
22) 2-Nitrophenol-d4	9.32	143	10604	3.48	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.85	165	18802	2.93	ng/ul	0.00
29) 4-Chloroaniline-d4	10.36	131	48361m	7.16	ng/ul	0.00
43) Dimethylphthalate-d6	13.53	166	73494	3.28	ng/ul	0.00
46) Acenaphthylene-d8	13.80	160	85394	3.28	ng/ul	0.00
51) 4-Nitrophenol-d4	14.32	143	10678	2.71	ng/ul	0.00
57) Fluorene-d10	15.12	176	68634	3.55	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.26	200	8328	2.26	ng/ul	0.02
70) Anthracene-d10	16.97	188	103337	3.56	ng/ul	0.00
76) Pyrene-d10	19.28	212	121676	3.77	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.09	264	92828	3.48	ng/ul	0.00

S.J
06/30/2017

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Phenol	6.67	94	12336	1.467	ng/ul#	85
11) 2-Methylphenol	7.90	108	11177	1.744	ng/ul	99
16) 4-Methylphenol	8.23	108	31019	4.399	ng/ul	97
24) 2,4-Dimethylphenol	9.42	107	51898m	6.316	ng/ul	97
28) Naphthalene	10.30	128	17192183	932.317	ng/ul#	93
34) 2-Methylnaphthalene	11.91	142	3469264	234.041	ng/ul	98
40) 1,1'-Biphenyl	12.94	154	765196	38.394	ng/ul	97
44) Dimethylphthalate	13.58	163	30702	1.400	ng/ul#	96
47) Acenaphthylene	13.83	152	112781	4.549	ng/ul#	93
49) Acenaphthene	14.19	153	2541536	151.968	ng/ul	98
53) Dibenzofuran	14.52	168	2796478	110.413	ng/ul	91
58) Fluorene	15.18	166	2810663	132.545	ng/ul	98
69) Phenanthrene	16.93	178	13067620	409.154	ng/ul	94
71) Anthracene	17.01	178	2088336	63.468	ng/ul	99
72) Carbazole	17.28	167	1022947	35.634	ng/ul	98
74) Fluoranthene	18.96	202	8128314	197.478	ng/ul#	94
77) Pyrene	19.32	202	5233393	130.578	ng/ul#	96
80) Benzo(a)anthracene	21.07	228	1770765	43.823	ng/ul	98
82) Chrysene	21.12	228	1269244	35.230	ng/ul	98
85) Benzo(b)fluoranthene	22.59	252	885053	24.713	ng/ul#	99
86) Benzo(k)fluoranthene	22.63	252	369889m	10.895	ng/ul	99
88) Benzo(a)pyrene	23.13	252	536392	16.269	ng/ul	99
89) Indeno(1,2,3-cd)pyrene	25.33	276	153803	4.498	ng/ul	95

S.J
06/30/2017

S.J
06/30/2017

OM-EPA-BM062017.M Fri Jun 23 13:28:48 2017

Quantitation Report

(LSC Reviewed)

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
Data File : BM010677.D
Acq On : 23 Jun 2017 01:32
Operator : SJ/JU
Sample : I3653-04 5X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Quant Time: Jun 23 13:22:35 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 23 05:06:10 2017
Response via : Initial Calibration

Instrument :

BNA_M

ClientSampleId :

F5B21

Manual Integrations

APPROVED

mohammad

6/26/2017 8:26:25 AM

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
90) Dibenzo(a,h)anthracene	25.33	278	47770	1.676	ng/ul#	95
91) Benzo(g,h,i)perylene	25.98	276	92135	3.385	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062217\
Data File : BM010677.D
Acq On : 23 Jun 2017 01:32
Operator : SJ/JU
Sample : I3653-04 5X
Misc :
ALS Vial : 28 Sample Multiplier: 1


Quant Time: Jun 23 13:22:35 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 23 05:06:10 2017
Response via : Initial Calibration

Instrument :

BNA_M

Client Sampled :

F5B21

Manual Integrations
APPROVEDmohammad
6/26/2017 8:26:25 AM

SOM-EPA-BM062017.M Fri Jun 23 13:28:49 2017