

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062221\
 Data File : BM030544.D
 Acq On : 23 Jun 2021 09:17
 Operator : CG/JU
 Sample : PB137214BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS214

Manual Integrations
 APPROVED

mohammad
 6/23/2021 4:49:05 PM

Quant Time: Jun 23 09:57:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM062221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 22 15:07:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	171866	20.000	ng/u1	0.00
20) Naphthalene-d8	10.593	136	737364	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.440	164	470417	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.175	188	879865	20.000	ng/u1	0.00
79) Chrysene-d12	21.345	240	669646	20.000	ng/u1	0.00
88) Perylene-d12	23.615	264	642446	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	25882	6.105	ng/uL	0.00
4) Pyridine-d5	3.693	84	311016	27.306	ng/u1	0.00
7) Phenol-d5	6.975	99	524697	35.399	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	325319	34.918	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	423222	36.991	ng/u1	0.00
15) 4-Methylphenol-d8	8.510	113	427097	37.026	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	207210	37.659	ng/u1	0.00
24) 2-Nitrophenol-d4	9.681	143	231415	37.845	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	455960	38.925	ng/u1	0.00
31) 4-Chloroaniline-d4	10.734	131	556477	34.066	ng/u1	0.00
46) Dimethylphthalate-d6	13.851	166	1298520	39.977	ng/u1	0.00
49) Acenaphthylene-d8	14.128	160	1650465	39.146	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	212535	34.539	ng/u1	0.00
60) Fluorene-d10	15.428	176	1127472	39.136	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.551	200	189067	32.844	ng/u1	0.00
73) Anthracene-d10	17.275	188	1535215	37.387	ng/u1	0.00
81) Pyrene-d10	19.563	212	1461405	41.649	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	1380513	41.331	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	59729	13.248	ng/uL	95
5) Pyridine	3.711	79	350846	28.615	ng/u1	98
6) Benzaldehyde	6.952	77	309348	42.923	ng/u1	96
8) Phenol	6.999	94	536287	35.782	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.234	93	412312	34.905	ng/u1	96
12) 2-Chlorophenol	7.369	128	422223	36.621	ng/u1	99
13) 2-Methylphenol	8.246	108	395732	36.437	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.328	45	660925	35.298	ng/u1	99
16) Acetophenone	8.628	105	672999	37.782	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.616	70	353404	40.110	ng/u1	99
18) 4-Methylphenol	8.575	108	441231	37.306	ng/u1	97
19) Hexachloroethane	8.881	117	168291	34.836	ng/u1	95
22) Nitrobenzene	9.005	77	504450	36.342	ng/u1	100
23) Isophorone	9.534	82	951469	39.584	ng/u1	98
25) 2-Nitrophenol	9.710	139	247605	38.136	ng/u1	99
26) 2,4-Dimethylphenol	9.775	107	366667	27.941	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.010	93	587908	37.594	ng/u1	100
29) 2,4-Dichlorophenol	10.246	162	436630	38.908	ng/u1	100
30) Naphthalene	10.646	128	1351194	36.193	ng/u1	100
32) 4-Chloroaniline	10.757	127	548033	33.794	ng/u1	99
33) Hexachlorobutadiene	10.928	225	272156	35.867	ng/u1	99
34) Caprolactam	11.540	113	138166m	42.419	ng/u1	
35) 4-Chloro-3-methylphenol	11.881	107	460864	41.825	ng/u1	98
36) 2-Methylnaphthalene	12.257	142	1007158	38.659	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062221\
 Data File : BM030544.D
 Acq On : 23 Jun 2021 09:17
 Operator : CG/JU
 Sample : PB137214BS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS214

Manual Integrations
 APPROVED


mohammad
 6/23/2021 4:49:05 PM

Quant Time: Jun 23 09:57:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM062221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 22 15:07:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.475	142	982245	37.248	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.628	216	554971	37.814	ng/ul	98
40) Hexachlorocyclopentadiene	12.604	237	183038	19.112	ng/ul	99
41) 2,4,6-Trichlorophenol	12.869	196	359228	38.865	ng/ul	98
42) 2,4,5-Trichlorophenol	12.940	196	390649	39.601	ng/ul	99
43) 1,1'-Biphenyl	13.269	154	1326500	38.207	ng/ul	100
44) 2-Chloronaphthalene	13.310	162	1034890	37.915	ng/ul	100
45) 2-Nitroaniline	13.516	65	314212	42.835	ng/ul	98
47) Dimethylphthalate	13.892	163	1313335	41.267	ng/ul	100
48) 2,6-Dinitrotoluene	14.016	165	275162	44.874	ng/ul	95
50) Acenaphthylene	14.157	152	1536369	39.284	ng/ul	100
51) 3-Nitroaniline	14.345	138	274549	42.031	ng/ul	96
52) Acenaphthene	14.498	153	1122715	39.557	ng/ul	99
53) 2,4-Dinitrophenol	14.551	184	126892	31.479	ng/ul	96
55) 4-Nitrophenol	14.651	109	178329	36.224	ng/ul	97
56) Dibenzofuran	14.834	168	1561554	39.465	ng/ul	100
57) 2,4-Dinitrotoluene	14.804	165	373596	43.514	ng/ul#	93
58) 2,3,4,6-Tetrachlorophenol	15.063	232	309781	38.793	ng/ul	100
59) Diethylphthalate	15.257	149	1305010	42.686	ng/ul	99
61) Fluorene	15.486	166	1335503	40.955	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.481	204	673861	41.494	ng/ul	97
63) 4-Nitroaniline	15.510	138	279661	44.959	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.569	198	187651	33.497	ng/ul#	95
67) N-Nitrosodiphenylamine	15.692	169	1107514	42.234	ng/ul	99
68) 4-Bromophenyl-phenylether	16.369	248	399068	44.450	ng/ul	98
69) Hexachlorobenzene	16.486	284	435873	41.973	ng/ul	99
70) Atrazine	16.639	200	216795	23.309	ng/ul	98
71) Pentachlorophenol	16.828	266	203376	31.253	ng/ul	98
72) Phenanthrene	17.216	178	1902474	40.264	ng/ul	100
74) Anthracene	17.310	178	1870696	38.722	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.234	216	557189	39.704	ng/ul	99
76) Pentachlorobenzene	14.757	250	526912	39.416	ng/ul	98
77) Carbazole	17.580	167	1593354	36.779	ng/ul	99
78) Di-n-butylphthalate	18.139	149	2019673	45.753	ng/ul	100
80) Fluoranthene	19.227	202	1888704	44.073	ng/ul	99
82) Pyrene	19.592	202	1918650	42.610	ng/ul	99
83) Butylbenzylphthalate	20.486	149	726362	49.005	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.263	252	504229	38.199	ng/ul	97
85) Benzo(a)anthracene	21.327	228	1660508	39.957	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.257	149	1032697	48.441	ng/ul	98
87) Chrysene	21.380	228	1582888	39.071	ng/ul	99
89) Di-n-octyl phthalate	22.139	149	1649831	44.642	ng/ul	100
90) Benzo(b)fluoranthene	22.933	252	1668172	40.899	ng/ul	99
91) Benzo(k)fluoranthene	22.974	252	1682486	42.922	ng/ul	99
93) Benzo(a)pyrene	23.515	252	1518365	41.400	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.921	276	2064076	44.705	ng/ul	98
95) Dibenzo(a,h)anthracene	25.927	278	1731627	45.242	ng/ul	99
96) Benzo(g,h,i)perylene	26.621	276	1702911	44.207	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Abundance Scan 39 (3.316 min): BM030542.D\data.ms (-33)



#2
1,4-Dioxane
Concen: 13.248 ng/uL
RT: 3.311 min Scan# 38
Delta R.T. -0.005 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion: 88 Resp: 59729
Ion Ratio Lower Upper
88 100
43 36.4 31.5 47.3
58 83.9 64.1 96.1

Instrument :
BNA_M
ClientSampleId :
SLCS214


Manual Integrations
APPROVED

mohammad
6/23/2021 4:49:05 PM


Abundance Scan 38 (3.311 min): BM030544.D\data.ms



Abundance Scan 38 (3.311 min): BM030544.D\data.ms (-5)





Abundance Scan 107 (3.716 min): BM030542.D\data.ms (-10)




#5
Pyridine
Concen: 28.615 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion: 79 Resp: 350846
Ion Ratio Lower Upper
79 100
52 86.6 69.9 104.9
51 40.3 33.8 50.6

Abundance Scan 106 (3.711 min): BM030544.D\data.ms




Abundance Scan 106 (3.711 min): BM030544.D\data.ms (-9)






#6
Benzaldehyde
Concen: 42.923 ng/ul
RT: 6.952 min Scan# 657
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion: 77 Resp: 309348
Ion Ratio Lower Upper
77 100
105 95.5 82.2 123.4
106 94.2 75.0 112.4




Instrument :
BNA_M
ClientSampled :
SLCS214

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#8
Phenol
Concen: 35.782 ng/ul
RT: 6.999 min Scan# 665
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion: 94 Resp: 536287
Ion Ratio Lower Upper
94 100
65 30.4 24.2 36.4
66 38.1 31.4 47.0









#10
Bis(2-Chloroethyl)ether
Concen: 34.905 ng/ul
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:	93	Resp:	412312
Ion	Ratio	Lower	Upper
93	100		
63	80.4	68.0	102.0
95	32.3	25.8	38.6

Instrument :
BNA_M
ClientSampleId :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#12
2-Chlorophenol
Concen: 36.621 ng/ul
RT: 7.369 min Scan# 728
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:	128	Resp:	422223
Ion	Ratio	Lower	Upper
128	100		
64	49.5	39.2	58.8
130	32.4	25.3	37.9









#13
2-Methylphenol
Concen: 36.437 ng/ul
RT: 8.246 min Scan# 877
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:108 Resp: 395732
Ion Ratio Lower Upper
108 100
107 91.0 70.6 105.8

Instrument :
BNA_M
ClientSampleId :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.298 ng/ul
RT: 8.328 min Scan# 891
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion: 45 Resp: 660925
Ion Ratio Lower Upper
45 100
77 12.4 10.2 15.4
79 9.5 7.8 11.8









#16
Acetophenone
Concen: 37.782 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:105 Resp: 672999
Ion Ratio Lower Upper
105 100
77 78.1 63.7 95.5
51 29.4 25.4 38.0

Instrument :
BNA_M
ClientSampleId :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#17
N-Nitroso-di-n-propylamine
Concen: 40.110 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion: 70 Resp: 353404
Ion Ratio Lower Upper
70 100
42 67.4 53.4 80.0
101 10.2 7.4 11.0
130 21.1 16.8 25.2





#18
4-Methylphenol
Concen: 37.306 ng/ul
RT: 8.575 min Scan# 933
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampleId :
SLCS214




Tgt Ion:108 Resp: 441231
Ion Ratio Lower Upper
108 100
107 112.3 87.0 130.6

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#19
Hexachloroethane
Concen: 34.836 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:117 Resp: 168291
Ion Ratio Lower Upper
117 100
201 106.5 80.1 120.1
199 66.1 50.8 76.2









#22
Nitrobenzene
Concen: 36.342 ng/ul
RT: 9.005 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion: 77 Resp: 504450
Ion Ratio Lower Upper
77 100
123 43.7 34.8 52.2
65 13.2 10.4 15.6



#23
Isophorone
Concen: 39.584 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion: 82 Resp: 951469
Ion Ratio Lower Upper
82 100
95 7.2 5.3 7.9
138 17.8 13.6 20.4



Instrument :
BNA_M
ClientSampled :
SLCS214

Manual Integrations
APPROVED

mohammad
6/23/2021 4:49:05 PM







#25
2-Nitrophenol
Concen: 38.136 ng/ul
RT: 9.710 min Scan# 1126
Delta R.T. -0.005 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:139 Resp: 247605
Ion Ratio Lower Upper
139 100
65 54.2 43.7 65.5
109 30.6 24.0 36.0

Instrument :
BNA_M
ClientSampled :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#26
2,4-Dimethylphenol
Concen: 27.941 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:107 Resp: 366667
Ion Ratio Lower Upper
107 100
121 50.6 41.0 61.6
122 85.1 68.2 102.2





#27
Bis(2-Chloroethoxy)methane
Concen: 37.594 ng/ul
RT: 10.010 min Scan# 1177
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion: 93 Resp: 587908

Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.9	38.9
123	11.2	8.5	12.7

Instrument :
BNA_M
ClientSampleId :
SLCS214

Manual Integrations
APPROVED




mohammad
6/23/2021 4:49:05 PM




#29
2,4-Dichlorophenol
Concen: 38.908 ng/ul
RT: 10.246 min Scan# 1217
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion: 162 Resp: 436630

Ion	Ratio	Lower	Upper
162	100		
164	64.8	51.6	77.4
98	35.4	28.2	42.4





#30
Naphthalene
Concen: 36.193 ng/u1
RT: 10.646 min Scan# 1285
Delta R.T. -0.005 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion:128 Resp: 1351194

Ion	Ratio	Lower	Upper
128	100		
129	10.9	9.0	13.4
127	13.2	10.6	15.8

Instrument :
BNA_M
ClientSampleId :
SLCS214

Manual Integrations
APPROVED




mohammad
6/23/2021 4:49:05 PM




#32
4-Chloroaniline
Concen: 33.794 ng/u1
RT: 10.757 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:127 Resp: 548033


Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.6	38.4





#33
Hexachlorobutadiene
Concen: 35.867 ng/ul
RT: 10.928 min Scan# 1333
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:225 Resp: 272156
Ion Ratio Lower Upper
225 100
223 64.2 50.7 76.1
227 62.8 50.7 76.1

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#34
Caprolactam
Concen: 42.419 ng/ul m
RT: 11.540 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion:113 Resp: 138166
Ion Ratio Lower Upper
113 100
55 187.4 155.2 232.8
56 137.9 101.9 152.9









#35
4-Chloro-3-methylphenol
Concen: 41.825 ng/ul
RT: 11.881 min Scan# 1495
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:107 Resp: 460864
Ion Ratio Lower Upper
107 100
144 27.5 22.3 33.5
142 84.8 65.6 98.4

Instrument :
BNA_M
ClientSampleId :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#36
2-Methylnaphthalene
Concen: 38.659 ng/ul
RT: 12.257 min Scan# 1559
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:142 Resp: 1007158
Ion Ratio Lower Upper
142 100
141 88.1 70.6 105.8









#37
1-Methylnaphthalene
Concen: 37.248 ng/ul
RT: 12.475 min Scan# 1596
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:142 Resp: 982245
Ion Ratio Lower Upper
142 100
141 92.9 72.8 109.2
116 3.8 3.1 4.7

Instrument :
BNA_M
ClientSampled :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.814 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion:216 Resp: 554971
Ion Ratio Lower Upper
216 100
214 79.6 62.2 93.4
179 20.5 16.6 24.8
108 16.6 13.5 20.3





#40
Hexachlorocyclopentadiene
Concen: 19.112 ng/ul
RT: 12.604 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:237 Resp: 183038
Ion Ratio Lower Upper
237 100
235 61.8 49.0 73.4
272 12.7 11.0 16.4

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#41
2,4,6-Trichlorophenol
Concen: 38.865 ng/ul
RT: 12.869 min Scan# 1663
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:196 Resp: 359228
Ion Ratio Lower Upper
196 100
198 96.8 78.8 118.2
200 31.0 26.0 39.0





#42
2,4,5-Trichlorophenol
Concen: 39.601 ng/ul
RT: 12.940 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:196 Resp: 390649
Ion Ratio Lower Upper
196 100
198 96.3 76.1 114.1
200 31.1 25.4 38.2

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#43
1,1'-Biphenyl
Concen: 38.207 ng/ul
RT: 13.269 min Scan# 1731
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:154 Resp: 1326500
Ion Ratio Lower Upper
154 100
153 40.4 32.3 48.5
76 12.9 10.6 15.8





#44
2-Chloronaphthalene
Concen: 37.915 ng/ul
RT: 13.310 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:162 Resp: 1034890
Ion Ratio Lower Upper
162 100
164 32.9 26.2 39.4
127 36.9 29.4 44.0

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#45
2-Nitroaniline
Concen: 42.835 ng/ul
RT: 13.516 min Scan# 1773
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion: 65 Resp: 314212
Ion Ratio Lower Upper
65 100
92 67.3 53.9 80.9
138 101.6 83.3 124.9









#47
Dimethylphthalate
Concen: 41.267 ng/ul
RT: 13.892 min Scan# 1837
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:163 Resp: 1313335
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.4 8.2 12.4

Instrument :
BNA_M
ClientSampleId :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#48
2,6-Dinitrotoluene
Concen: 44.874 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion:165 Resp: 275162
Ion Ratio Lower Upper
165 100
89 56.5 48.6 72.8
121 20.4 18.1 27.1





#50
Acenaphthylene
Concen: 39.284 ng/ul
RT: 14.157 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampleId :
SLCS214




Tgt Ion:152 Resp: 1536369
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.0 10.2 15.4

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#51
3-Nitroaniline
Concen: 42.031 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:138 Resp: 274549
Ion Ratio Lower Upper
138 100
108 10.3 7.4 11.2
92 115.6 96.5 144.7





#52
Acenaphthene
Concen: 39.557 ng/ul
RT: 14.498 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:153 Resp: 1122715
Ion Ratio Lower Upper
153 100
152 46.9 37.4 56.2
154 87.9 69.4 104.0

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#53
2,4-Dinitrophenol
Concen: 31.479 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:184 Resp: 126892
Ion Ratio Lower Upper
184 100
63 67.9 56.9 85.3
154 64.0 49.0 73.6








#55
4-Nitrophenol
Concen: 36.224 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Instrument :
BNA_M
ClientSampled :
SLCS214





Tgt Ion:109 Resp: 178329
Ion Ratio Lower Upper
109 100
139 128.1 100.3 150.5
65 135.5 104.6 156.8




#56
Dibenzofuran
Concen: 39.465 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion:168 Resp: 1561554
Ion Ratio Lower Upper
168 100
139 37.0 29.5 44.3









#57
2,4-Dinitrotoluene
Concen: 43.514 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:165 Resp: 373596
Ion Ratio Lower Upper
165 100
63 42.5 37.6 56.4
182 2.8 1.8 2.8#

Instrument :
BNA_M
ClientSampled :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.793 ng/ul
RT: 15.063 min Scan# 2036
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:232 Resp: 309781
Ion Ratio Lower Upper
232 100
131 43.4 35.0 52.6
130 2.3 1.7 2.5
166 30.4 24.4 36.6









#59
Diethylphthalate
Concen: 42.686 ng/ul
RT: 15.257 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:149 Resp: 1305010
Ion Ratio Lower Upper
149 100
177 22.0 17.0 25.4
150 12.4 9.9 14.9

Instrument :
BNA_M
ClientSampled :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#61
Fluorene
Concen: 40.955 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion:166 Resp: 1335503
Ion Ratio Lower Upper
166 100
165 98.3 78.4 117.6
167 13.0 10.6 15.8





#62
4-Chlorophenyl-phenylether
Concen: 41.494 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:204 Resp: 673861
Ion Ratio Lower Upper
204 100
206 33.5 26.0 39.0
141 54.8 45.6 68.4

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#63
4-Nitroaniline
Concen: 44.959 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:138 Resp: 279661
Ion Ratio Lower Upper
138 100
92 51.1 41.5 62.3
108 81.4 61.4 92.2





#66
4,6-Dinitro-2-methylphenol
Concen: 33.497 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:198 Resp: 187651
Ion Ratio Lower Upper
198 100
182 3.7 3.8 5.6#
77 22.4 20.1 30.1

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#67
N-Nitrosodiphenylamine
Concen: 42.234 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:169 Resp: 1107514
Ion Ratio Lower Upper
169 100
168 67.7 53.8 80.6
167 36.7 28.6 43.0





#68
4-Bromophenyl-phenylether
Concen: 44.450 ng/ul
RT: 16.369 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampleId :
SLCS214




Tgt Ion:248 Resp: 399068
Ion Ratio Lower Upper
248 100
250 97.9 78.6 118.0
141 67.7 56.2 84.2

Manual Integrations
APPROVED



mohammad
6/23/2021 4:49:05 PM




#69
Hexachlorobenzene
Concen: 41.973 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:284 Resp: 435873
Ion Ratio Lower Upper
284 100
142 35.2 29.0 43.4
249 31.9 25.0 37.4









#70
Atrazine
Concen: 23.309 ng/u1
RT: 16.639 min Scan# 2304
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion:200 Resp: 216795
Ion Ratio Lower Upper
200 100
173 24.9 18.7 28.1
215 47.3 38.9 58.3



#71
Pentachlorophenol
Concen: 31.253 ng/u1
RT: 16.828 min Scan# 2336
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Tgt Ion:266 Resp: 203376
Ion Ratio Lower Upper
266 100
264 63.2 48.2 72.4
268 65.2 51.6 77.4



Instrument :
BNA_M
ClientSampleId :
SLCS214

Manual Integrations
APPROVED

mohammad
6/23/2021 4:49:05 PM



#72
Phenanthrene
Concen: 40.264 ng/u1
RT: 17.216 min Scan# 2402
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion:178 Resp: 1902474

Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.1	18.1
176	19.3	15.2	22.8

Instrument :
BNA_M
ClientSampled :
SLCS214

Manual Integrations
APPROVED




mohammad
6/23/2021 4:49:05 PM




#74
Anthracene
Concen: 38.722 ng/u1
RT: 17.310 min Scan# 2418
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:178 Resp: 1870696

Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.1	18.1
176	18.7	15.2	22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 39.704 ng/uL
RT: 13.234 min Scan# 1725
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:216 Resp: 557189
Ion Ratio Lower Upper
216 100
214 78.3 61.9 92.9
218 47.9 39.5 59.3

Instrument :
BNA_M
ClientSampled :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM




#76
Pentachlorobenzene
Concen: 39.416 ng/uL
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:250 Resp: 526912
Ion Ratio Lower Upper
250 100
248 63.2 50.2 75.4
252 63.1 52.2 78.2






#77
Carbazole
Concen: 36.779 ng/ul
RT: 17.580 min Scan# 2464
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion:167 Resp: 1593354


Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.0	25.6
139	12.3	9.9	14.9




Instrument :
BNA_M
ClientSampleId :
SLCS214

Manual Integrations
APPROVED

mohammad
6/23/2021 4:49:05 PM






#78
Di-n-butylphthalate
Concen: 45.753 ng/ul
RT: 18.139 min Scan# 2559
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion:149 Resp: 2019673

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	5.0	4.1	6.1









#80
Fluoranthene
Concen: 44.073 ng/ul
RT: 19.227 min Scan# 2744
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:202 Resp: 1888704
Ion Ratio Lower Upper
202 100
101 10.5 8.8 13.2
100 8.1 6.8 10.2

Instrument :
BNA_M
ClientSampleId :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#82
Pyrene
Concen: 42.610 ng/ul
RT: 19.592 min Scan# 2806
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:202 Resp: 1918650
Ion Ratio Lower Upper
202 100
101 12.2 10.0 15.0
100 9.8 8.5 12.7









#83
Butylbenzylphthalate
Concen: 49.005 ng/ul
RT: 20.486 min Scan# 2958
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:149 Resp: 726362
Ion Ratio Lower Upper
149 100
91 68.5 57.4 86.2
206 21.1 15.9 23.9

Instrument :
BNA_M
ClientSampled :
SLCS214




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#84
3,3'-Dichlorobenzidine
Concen: 38.199 ng/ul
RT: 21.263 min Scan# 3090
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:252 Resp: 504229
Ion Ratio Lower Upper
252 100
254 65.1 49.9 74.9
126 10.3 8.9 13.3





#85
Benzo(a)anthracene
Concen: 39.957 ng/ul
RT: 21.327 min Scan# 3101
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17





Tgt Ion:228 Resp: 1660508

Ion	Ratio	Lower	Upper
228	100		
229	19.9	16.0	24.0
226	27.5	22.2	33.2

Instrument :
BNA_M
ClientSampled :
SLCS214

Manual Integrations
APPROVED




mohammad
6/23/2021 4:49:05 PM




#86
Bis(2-ethylhexyl)phthalate
Concen: 48.441 ng/ul
RT: 21.257 min Scan# 3089
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Tgt Ion:149 Resp: 1032697


Ion	Ratio	Lower	Upper
149	100		
167	26.9	20.7	31.1
279	4.2	3.0	4.4





#87
Chrysene
Concen: 39.071 ng/ul
RT: 21.380 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17

Instrument :
BNA_M
ClientSampleId :
SLCS214






Tgt Ion:228 Resp: 1582888

Ion	Ratio	Lower	Upper
228	100		
226	30.5	23.6	35.4
229	19.9	15.9	23.9




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#89
Di-n-octyl phthalate
Concen: 44.642 ng/ul
RT: 22.139 min Scan# 3239
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion:149 Resp: 1649831





#90
Benzo(b)fluoranthene
Concen: 40.899 ng/u1
RT: 22.933 min Scan# 3374
Delta R.T. -0.006 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214



Tgt Ion:252 Resp: 1668172
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.6
125 9.0 7.7 11.5




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#91
Benzo(k)fluoranthene
Concen: 42.922 ng/u1
RT: 22.974 min Scan# 3381
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion:252 Resp: 1682486
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 9.1 7.9 11.9





#93
Benzo(a)pyrene
Concen: 41.400 ng/ul
RT: 23.515 min Scan# 3473
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214



Tgt Ion:252 Resp: 1518365
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 10.4 8.6 13.0




Manual Integrations
APPROVED


mohammad
6/23/2021 4:49:05 PM



#94
Indeno(1,2,3-cd)pyrene
Concen: 44.705 ng/ul
RT: 25.921 min Scan# 3882
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17


Tgt Ion:276 Resp: 2064076
Ion Ratio Lower Upper
276 100
138 23.3 19.5 29.3
277 24.9 19.6 29.4





#95
Dibenzo(a,h)anthracene
Concen: 45.242 ng/ul
RT: 25.927 min Scan# 3883
Delta R.T. -0.011 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17




Instrument :
BNA_M
ClientSampled :
SLCS214




Tgt Ion:278 Resp: 1731627
Ion Ratio Lower Upper
278 100
139 14.7 12.8 19.2
279 23.8 19.1 28.7

Manual Integrations
APPROVED

mohammad
6/23/2021 4:49:05 PM



#96
Benzo(g,h,i)perylene
Concen: 44.207 ng/ul
RT: 26.621 min Scan# 4001
Delta R.T. -0.017 min
Lab File: BM030544.D
Acq: 23 Jun 2021 09:17



Tgt Ion:276 Resp: 1702911
Ion Ratio Lower Upper
276 100
138 19.4 16.5 24.7
277 24.3 18.5 27.7

