

Data Path : Z:\HPCHEM1\BNA M\DATA\BM062316\
 Data File : BM005949.D
 Acq On : 24 Jun 2016 02:59
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Jun 24 04:34:37 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM062216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 00:58:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.544	0.539	0.9	95	0.00
3 S	1,4-Dioxane-d8	0.502	0.514	-2.4	100	0.00
4	Benzaldehyde	0.332	0.401	-20.8#	96	0.00
5 S	Phenol-d5	1.722	1.730	-0.5	90	0.00
6	Phenol	1.874	1.905	-1.7	90	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.054	1.078	-2.3	90	0.00
8	Bis(2-Chloroethyl)ether	1.377	1.412	-2.5	90	0.00
9 S	2-Chlorophenol-d4	1.356	1.340	1.2	91	0.00
10	2-Chlorophenol	1.417	1.414	0.2	92	0.00
11	2-Methylphenol	1.360	1.355	0.4	87	0.00
12	2,2'-oxybis(1-Chloropropane	2.015	2.027	-0.6	87	0.00
13 S	4-Methylphenol-d8	1.349	1.320	2.1	84	0.00
14	Acetophenone	2.076	2.109	-1.6	84	0.00
15 P	N-Nitroso-di-n-propylamine	1.145	1.104	3.6	80	0.00
16	4-Methylphenol	1.472	1.453	1.3	84	0.00
17	Hexachloroethane	0.573	0.563	1.7	93	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
19 S	Nitrobenzene-d5	0.147	0.150	-2.0	86	0.00
20	Nitrobenzene	0.387	0.396	-2.3	86	0.00
21	Isophorone	0.710	0.694	2.3	77	0.00
22 S	2-Nitrophenol-d4	0.154	0.153	0.6	82	0.00
23 C	2-Nitrophenol	0.173	0.174	-0.6	83	0.00
24	2,4-Dimethylphenol	0.377	0.380	-0.8	82	0.00
25	Bis(2-Chloroethoxy)methane	0.442	0.440	0.5	81	0.00
26 S	2,4-Dichlorophenol-d3	0.295	0.304	-3.1	83	0.00
27 C	2,4-Dichlorophenol	0.299	0.308	-3.0	83	0.00
28	Naphthalene	0.998	1.008	-1.0	85	0.00
29 S	4-Chloroaniline-d4	0.323	0.373	-15.5	82	0.00
30	4-Chloroaniline	0.348	0.398	-14.4	81	0.00
31 C	Hexachlorobutadiene	0.184	0.195	-6.0	91	0.00
32	Caprolactam	0.110	0.101	8.2	71	0.00
33 C	4-Chloro-3-methylphenol	0.348	0.335	3.7	75	0.00
34	2-Methylnaphthalene	0.731	0.729	0.3	80	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
36	1,2,4,5-Tetrachlorobenzene	0.635	0.698	-9.9	84	0.00
37	Hexachlorocyclopentadiene	0.367	0.408	-11.2	83	0.00
38 C	2,4,6-Trichlorophenol	0.431	0.452	-4.9	76	0.00
39	2,4,5-Trichlorophenol	0.450	0.466	-3.6	75	0.00
40	1,1'-Biphenyl	1.643	1.728	-5.2	78	0.00
41	2-Chloronaphthalene	1.267	1.339	-5.7	79	0.00
42	2-Nitroaniline	0.425	0.427	-0.5	71	0.00
43 S	Dimethylphthalate-d6	1.589	1.562	1.7	69	0.00
44	Dimethylphthalate	1.604	1.573	1.9	69	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM062316\
 Data File : BM005949.D
 Acq On : 24 Jun 2016 02:59
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Jun 24 04:34:37 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM062216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 00:58:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.314	0.316	-0.6	69	0.00
46 S	Acenaphthylene-d8	1.985	2.065	-4.0	76	0.00
47	Acenaphthylene	2.108	2.184	-3.6	75	0.00
48	3-Nitroaniline	0.315	0.360	-14.3	73	0.00
49 C	Acenaphthene	1.334	1.369	-2.6	75	0.00
50	2,4-Dinitrophenol	0.172	0.156	9.3	78	0.00
51 S	4-Nitrophenol-d4	0.311	0.312	-0.3	73	0.00
52	4-Nitrophenol	0.297	0.297	0.0	74	0.00
53	Dibenzofuran	1.925	1.969	-2.3	74	0.00
54	2,4-Dinitrotoluene	0.462	0.465	-0.6	69	0.00
55	2,3,4,6-Tetrachlorophenol	0.400	0.410	-2.5	73	0.00
56	Diethylphthalate	1.658	1.592	4.0	67	0.00
57 S	Fluorene-d10	1.373	1.395	-1.6	73	0.00
58	Fluorene	1.500	1.546	-3.1	74	0.00
59	4-Chlorophenyl-phenylether	0.720	0.747	-3.8	75	0.00
60	4-Nitroaniline	0.390	0.400	-2.6	73	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.100	0.099	1.0	71	0.00
63	4,6-Dinitro-2-methylphenol	0.109	0.111	-1.8	72	0.00
64	N-Nitrosodiphenylamine	0.578	0.592	-2.4	71	0.00
65	4-Bromophenyl-phenylether	0.204	0.216	-5.9	72	0.00
66	Hexachlorobenzene	0.227	0.235	-3.5	72	0.00
67	Atrazine	0.204	0.211	-3.4	65	0.00
68 C	Pentachlorophenol	0.136	0.142	-4.4	72	0.00
69	Phenanthrene	1.086	1.107	-1.9	71	0.00
70 S	Anthracene-d10	0.911	0.925	-1.5	71	0.00
71	Anthracene	1.109	1.137	-2.5	71	0.00
72	Carbazole	0.954	1.037	-8.7	71	0.00
73	Di-n-butylphthalate	1.232	1.190	3.4	63	0.00
74 C	Fluoranthene	1.191	1.315	-10.4	71	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
76 S	Pyrene-d10	0.928	0.923	0.5	72	0.00
77	Pyrene	1.223	1.223	0.0	73	0.00
78	Butylbenzylphthalate	0.511	0.494	3.3	66	0.00
79	3,3'-Dichlorobenzidine	0.333	0.419	-25.8#	86	0.00
80	Benzo(a)anthracene	1.187	1.217	-2.5	79	0.00
81	Bis(2-ethylhexyl)phthalate	0.716	0.705	1.5	68	0.00
82	Chrysene	1.095	1.123	-2.6	79	0.00
83 I	Perylene-d12	1.000	1.000	0.0	97	-0.01
84	Di-n-octyl phthalate	1.195	1.197	-0.2	76	0.00
85	Benzo(b)fluoranthene	1.215	1.199	1.3	88	0.00
86	Benzo(k)fluoranthene	1.106	1.176	-6.3	94	0.00
87 S	Benzo(a)pyrene-d12	0.901	0.927	-2.9	95	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM062316\
Data File : BM005949.D
Acq On : 24 Jun 2016 02:59
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02053

Quant Time: Jun 24 04:34:37 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM062216.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 24 00:58:22 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.128	1.177	-4.3	96	0.00
89	Indeno(1,2,3-cd)pyrene	1.259	1.334	-6.0	108	-0.01
90	Dibenzo(a,h)anthracene	1.039	1.114	-7.2	109	0.00
91	Benzo(a,h,i)perylene	1.086	1.127	-3.8	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0