

Data File : BM010698.D
Acq On : 23 Jun 2017 20:02
Operator : SJ/JU
Sample : I3625-08
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
F5B82

Manual Integrations
APPROVED

mohammad
6/27/2017 5:03:00 PM

Quant Time: Jun 24 04:56:41 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jun 24 04:37:39 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.46	152	71150	20.00	ng/ul	0.00
18) Naphthalene-d8	10.22	136	327603	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.11	164	220250	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.87	188	507920	20.00	ng/ul	0.00
75) Chrysene-d12	21.09	240	562648	20.00	ng/ul	0.00
83) Perylene-d12	23.23	264	423712	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.08	96	6370	5.13	ng/uL	0.00
5) Phenol-d5	6.65	99	168275	24.83	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.81	67	100303	25.92	ng/ul	0.00
9) 2-Chlorophenol-d4	7.00	132	132551	27.88	ng/ul	0.00
13) 4-Methylphenol-d8	8.16	113	159556	28.39	ng/ul	0.00
19) Nitrobenzene-d5	8.60	128	76779	32.86	ng/ul	0.00
22) 2-Nitrophenol-d4	9.32	143	91764	34.08	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.85	165	178594	31.50	ng/ul	0.00
29) 4-Chloroaniline-d4	10.36	131	175099	29.31	ng/ul	0.00
43) Dimethylphthalate-d6	13.53	166	605436	31.31	ng/ul	0.00
46) Acenaphthylene-d8	13.80	160	719790	32.01	ng/ul	0.00
51) 4-Nitrophenol-d4	14.32	143	92909	27.31	ng/ul	0.00
57) Fluorene-d10	15.12	176	526686	31.51	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.24	200	97953	31.40	ng/ul	0.00
70) Anthracene-d10	16.97	188	851083m	34.70	ng/ul	0.00
76) Pyrene-d10	19.28	212	1138744	43.57	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.10	264	880855	43.01	ng/ul	0.01

Target Compounds				Qvalue		
6) Phenol	6.67	94	11330	1.62	ng/ul	90
44) Dimethylphthalate	13.58	163	83676	4.42	ng/ul#	96
47) Acenaphthylene	13.83	152	43581	2.03	ng/ul	98
69) Phenanthrene	16.91	178	42802m	1.59	ng/ul	
71) Anthracene	17.00	178	83336	3.00	ng/ul	98
72) Carbazole	17.27	167	34963	1.44	ng/ul#	95
73) Di-n-butylphthalate	17.86	149	74444	2.41	ng/ul	97
74) Fluoranthene	18.94	202	187583	5.39	ng/ul#	97
77) Pyrene	19.30	202	191935	5.91	ng/ul	98
80) Benzo(a)anthracene	21.07	228	161635m	4.93	ng/ul	
82) Chrysene	21.12	228	198483	6.79	ng/ul	96
85) Benzo(b)fluoranthene	22.59	252	403387	14.67	ng/ul#	97
86) Benzo(k)fluoranthene	22.63	252	103364m	3.96	ng/ul	
88) Benzo(a)pyrene	23.13	252	174586	6.89	ng/ul#	92
89) Indeno(1,2,3-cd)pyrene	25.34	276	182596	6.95	ng/ul	97
90) Dibenzo(a,h)anthracene	25.34	278	47377	2.16	ng/ul#	81
91) Benzo(g,h,i)perylene	25.98	276	176350	8.44	ng/ul#	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM010698.D

Acq On : 23 Jun 2017 20:02

Operator : SJ/JU

Sample : I3625-08

Misc :

ALS Vial : 15 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

F5B82

Manual Integrations
APPROVEDmohammad
6/27/2017 5:03:00 PM

Quant Time: Jun 24 04:56:41 2017

Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M

Quant Title : SVOA CALIBRATION

QLast Update : Sat Jun 24 04:37:39 2017

Response via : Initial Calibration

