

Data Path : Z:\HPCHEM1\BNA M\DATA\BM062317\
 Data File : BM010745.D
 Acq On : 25 Jun 2017 00:15
 Operator : SJ/JU
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02050

Quant Time: Jun 26 08:46:06 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 04:54:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.390	0.419	-7.4	67	0.00
3 S	1,4-Dioxane-d8	0.349	0.386	-10.6	70	0.00
4	Benzaldehyde	1.121	1.247	-11.2	62	0.00
5 S	Phenol-d5	1.905	1.673	12.2	60	0.00
6	Phenol	1.962	1.749	10.9	61	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.088	0.940	13.6	56	0.00
8	Bis(2-Chloroethyl)ether	1.429	1.262	11.7	59	0.00
9 S	2-Chlorophenol-d4	1.337	1.294	3.2	65	0.00
10	2-Chlorophenol	1.336	1.269	5.0	63	0.00
11	2-Methylphenol	1.496	1.352	9.6	61	0.00
12	2,2'-oxybis(1-Chloropropane	0.997	0.787	21.1#	52	0.00
13 S	4-Methylphenol-d8	1.580	1.442	8.7	61	0.00
14	Acetophenone	2.579	2.334	9.5	61	0.00
15 P	N-Nitroso-di-n-propylamine	1.390	1.280	7.9	60	0.00
16	4-Methylphenol	1.646	1.522	7.5	63	0.00
17	Hexachloroethane	0.594	0.617	-3.9	69	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
19 S	Nitrobenzene-d5	0.143	0.141	1.4	67	0.00
20	Nitrobenzene	0.416	0.433	-4.1	67	0.00
21	Isophorone	0.828	0.730	11.8	57	0.00
22 S	2-Nitrophenol-d4	0.164	0.167	-1.8	67	0.00
23 C	2-Nitrophenol	0.174	0.176	-1.1	64	0.00
24	2,4-Dimethylphenol	0.443	0.442	0.2	66	0.00
25	Bis(2-Chloroethoxy)methane	0.459	0.416	9.4	58	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.349	-0.9	66	0.00
27 C	2,4-Dichlorophenol	0.337	0.345	-2.4	67	0.00
28	Naphthalene	0.995	0.977	1.8	64	0.00
29 S	4-Chloroaniline-d4	0.365	0.414	-13.4	64	0.00
30	4-Chloroaniline	0.367	0.403	-9.8	63	0.00
31 C	Hexachlorobutadiene	0.254	0.276	-8.7	70	0.00
32	Caprolactam	0.118	0.107	9.3	57	0.00
33 C	4-Chloro-3-methylphenol	0.430	0.423	1.6	63	0.00
34	2-Methylnaphthalene	0.800	0.805	-0.6	66	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.00
36	1,2,4,5-Tetrachlorobenzene	0.706	0.731	-3.5	70	0.00
37	Hexachlorocyclopentadiene	0.406	0.326	19.7	59	0.00
38 C	2,4,6-Trichlorophenol	0.482	0.473	1.9	67	0.00
39	2,4,5-Trichlorophenol	0.498	0.508	-2.0	69	0.00
40	1,1'-Biphenyl	1.564	1.548	1.0	68	0.00
41	2-Chloronaphthalene	1.230	1.230	0.0	69	0.00
42	2-Nitroaniline	0.360	0.401	-11.4	76	0.00
43 S	Dimethylphthalate-d6	1.756	1.742	0.8	67	0.00
44	Dimethylphthalate	1.721	1.688	1.9	67	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM062317\
 Data File : BM010745.D
 Acq On : 25 Jun 2017 00:15
 Operator : SJ/JU
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02050

Quant Time: Jun 26 08:46:06 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 04:54:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.298	0.327	-9.7	76	0.00
46 S	Acenaphthylene-d8	2.042	2.018	1.2	68	0.00
47	Acenaphthylene	1.945	1.890	2.8	67	0.00
48	3-Nitroaniline	0.301	0.315	-4.7	71	0.00
49 C	Acenaphthene	1.312	1.263	3.7	66	0.00
50	2,4-Dinitrophenol	0.216	0.218	-0.9	79	0.00
51 S	4-Nitrophenol-d4	0.309	0.300	2.9	69	0.00
52	4-Nitrophenol	0.392	0.499	-27.3#	87	0.00
53	Dibenzofuran	1.987	2.025	-1.9	69	0.00
54	2,4-Dinitrotoluene	0.462	0.493	-6.7	73	0.00
55	2,3,4,6-Tetrachlorophenol	0.504	0.515	-2.2	69	0.00
56	Diethylphthalate	1.811	1.785	1.4	67	0.00
57 S	Fluorene-d10	1.518	1.587	-4.5	72	0.00
58	Fluorene	1.664	1.675	-0.7	69	0.00
59	4-Chlorophenyl-phenylether	0.904	0.941	-4.1	72	0.00
60	4-Nitroaniline	0.355	0.359	-1.1	72	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	73	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.123	0.121	1.6	78	0.00
63	4,6-Dinitro-2-methylphenol	0.131	0.122	6.9	74	0.00
64	N-Nitrosodiphenylamine	0.555	0.530	4.5	69	0.00
65	4-Bromophenyl-phenylether	0.229	0.224	2.2	72	0.00
66	Hexachlorobenzene	0.241	0.239	0.8	73	0.00
67	Atrazine	0.234	0.234	0.0	71	0.00
68 C	Pentachlorophenol	0.170	0.158	7.1	69	0.00
69	Phenanthrene	1.063	1.060	0.3	73	0.00
70 S	Anthracene-d10	0.966	0.969	-0.3	72	0.00
71	Anthracene	1.095	1.103	-0.7	72	0.00
72	Carbazole	0.956	0.932	2.5	71	0.00
73	Di-n-butylphthalate	1.215	1.149	5.4	68	0.00
74 C	Fluoranthene	1.370	1.394	-1.8	74	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
76 S	Pyrene-d10	0.929	0.949	-2.2	73	0.00
77	Pyrene	1.155	1.167	-1.0	73	0.00
78	Butylbenzylphthalate	0.434	0.446	-2.8	74	0.00
79	3,3'-Dichlorobenzidine	0.400	0.384	4.0	68	0.00
80	Benzo(a)anthracene	1.165	1.183	-1.5	73	0.00
81	Bis(2-ethylhexyl)phthalate	0.663	0.657	0.9	71	0.00
82	Chrysene	1.038	1.044	-0.6	72	0.00
83 I	Perylene-d12	1.000	1.000	0.0	63	0.00
84	Di-n-octyl phthalate	1.398	1.475	-5.5	73	0.00
85	Benzo(b)fluoranthene	1.298	1.364	-5.1	67	0.00
86	Benzo(k)fluoranthene	1.231	1.255	-1.9	64	0.00
87 S	Benzo(a)pyrene-d12	0.967	0.977	-1.0	64	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM062317\
Data File : BM010745.D
Acq On : 25 Jun 2017 00:15
Operator : SJ/JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02050

Quant Time: Jun 26 08:46:06 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 26 04:54:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.195	1.205	-0.8	64	0.00
89	Indeno(1,2,3-cd)pyrene	1.240	1.230	0.8	62	0.00
90	Dibenzo(a,h)anthracene	1.033	1.028	0.5	62	0.00
91	Benzo(a,h,i)perylene	0.987	1.000	-1.3	63	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0