


Data Path : \\terastorage\SV0ASRV\HPCHEM1\BNA_M\Data\BM062317\
Data File : BM010683.D
Acq On : 23 Jun 2017 09:52
Operator : SJ/JU
Sample : DFTPP18
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP18

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Title : SVOA CALIBRATION
Last Update : Tue Jun 27 02:29:28 2017



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.6	63768	PASS
68	69	0.00	2	1.8	1476	PASS
69	198	0.00	100	37.7	83832	PASS
70	69	0.00	2	0.4	321	PASS
127	198	10	80	44.2	98498	PASS
197	198	0.00	2	0.8	1857	PASS
198	198	100	100	100.0	222656	PASS
199	198	5	9	6.5	14395	PASS
275	198	10	60	26.5	58973	PASS
365	198	1	100	4.4	9893	PASS
441	443	0.01	100	73.0	23133	PASS
442	198	50	100	71.1	158272	PASS
443	442	15	24	20.0	31698	PASS

m/z	Abundance
38.10	973.0
39.10	4362.0
49.00	576.0
50.10	17688.0
51.10	60744.0
52.00	2371.0
54.90	352.0
56.00	1218.0
56.90	3540.0
61.00	770.0
62.00	1181.0
63.00	2970.0
64.00	443.0
65.00	1529.0
68.00	1798.0
69.00	73184.0
70.00	360.0
73.00	956.0
74.00	10195.0
75.00	14204.0
76.10	4255.0
77.00	87264.0
78.00	6282.0
79.00	6936.0
80.00	5055.0
81.00	6795.0
81.90	2073.0
82.90	1046.0
85.00	996.0
86.00	2435.0
87.00	762.0
88.00	323.0
91.00	1662.0
92.10	2097.0
93.00	14212.0
93.90	832.0
96.00	797.0
98.00	12444.0
99.00	7241.0
100.00	627.0
101.00	3603.0
103.10	1204.0
104.00	2937.0
105.00	3009.0
106.00	781.0
107.00	32392.0
108.00	4715.0
108.90	690.0
110.00	44200.0
111.00	7548.0
111.90	1141.0
112.90	324.0
116.00	1535.0
117.00	34376.0
118.00	2174.0
119.90	311.0
122.00	2472.0
123.00	3320.0
123.90	1807.0
124.90	1044.0
126.20	622.0
127.00	81296.0
128.00	6730.0
129.00	39728.0
130.00	3296.0
131.00	825.0
132.80	355.0
134.00	1232.0
134.90	3030.0
135.80	1345.0
137.00	1798.0
140.00	547.0
140.90	4746.0
141.90	2197.0
143.00	1325.0
144.00	449.0
144.90	303.0
146.10	948.0
147.00	3505.0
148.00	7963.0
148.90	1638.0
150.80	705.0
151.20	590.0
153.00	1648.0
153.90	1653.0
155.00	2912.0
156.00	5421.0
156.80	779.0
157.80	649.0
159.10	508.0
159.90	1951.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

161.00	2846.0
161.90	666.0
164.90	2025.0
166.00	1238.0
167.00	12386.0
168.00	6591.0
169.00	661.0
170.00	320.0
170.70	431.0
171.20	419.0
171.80	1286.0
173.10	989.0
173.90	2357.0
175.00	4535.0
176.00	1383.0
176.90	2124.0
178.10	811.0
179.00	8874.0
180.00	7311.0
181.00	2210.0
182.20	533.0
183.30	306.0
184.00	696.0
185.00	2941.0
186.00	29192.0
187.00	8704.0
187.90	1338.0
189.10	2336.0
190.90	1064.0
191.90	3123.0
192.90	3026.0
194.10	669.0
196.00	4494.0
197.00	1632.0
197.90	173440.0
198.90	12659.0
200.00	938.0
201.60	516.0
202.20	437.0
203.00	1692.0
204.00	7530.0
205.00	13487.0
206.00	51096.0
207.00	7641.0
207.90	2505.0
209.00	904.0
209.90	1012.0
210.90	1910.0
211.80	420.0
215.00	910.0
216.00	1711.0
217.00	15673.0
217.90	1609.0
218.90	340.0
221.00	4552.0
222.00	1723.0
223.00	2778.0
224.00	32040.0
225.00	7613.0
226.00	815.0
227.00	15902.0
227.90	2347.0
229.00	3466.0
230.00	559.0
231.10	1304.0
231.90	388.0
233.00	340.0
233.80	645.0
235.10	733.0
236.00	555.0
237.10	1316.0
238.80	439.0
239.80	518.0
241.00	655.0
241.90	1393.0
243.00	2045.0
244.00	21248.0
245.10	2832.0
246.00	6077.0
247.00	983.0
247.90	400.0
248.80	438.0
250.70	342.0
252.10	367.0
253.10	538.0
254.00	1033.0
255.00	112232.0
256.00	17888.0
257.00	1358.0
258.00	8592.0
259.00	1416.0
265.00	3882.0
266.10	690.0
270.90	335.0
271.80	369.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

273.10	3661.0
274.00	8714.0
275.00	41152.0
276.00	6123.0
277.00	5341.0
278.00	1007.0
278.80	323.0
283.00	815.0
284.00	417.0
285.00	661.0
290.10	390.0
291.90	362.0
293.00	1382.0
293.90	333.0
296.00	18376.0
297.00	2356.0
298.10	357.0
303.00	1950.0
303.90	483.0
314.00	634.0
315.00	1804.0
316.00	953.0
320.90	751.0
322.00	439.0
323.10	4781.0
324.00	596.0
327.00	721.0
328.10	650.0
331.90	413.0
333.20	470.0
334.00	2348.0
335.10	487.0
340.90	649.0
346.00	712.0
352.10	1453.0
353.00	1026.0
353.90	1501.0
365.00	6017.0
366.00	1016.0
372.20	1732.0
373.00	457.0
383.00	628.0
401.80	486.0
403.10	640.0
403.90	343.0
421.00	760.0
422.00	889.0
423.10	5453.0
424.00	1404.0
441.00	13976.0
442.10	94656.0
443.10	20504.0
444.10	1240.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

m/z	Abundance
37.00	503.0
38.10	1577.0
39.10	4693.0
48.90	587.0
50.10	21440.0
51.00	74000.0
52.00	3711.0
55.10	319.0
56.00	2325.0
57.00	5079.0
58.10	366.0
61.00	1061.0
62.00	1842.0
63.00	3241.0
63.90	644.0
65.10	1834.0
68.10	1597.0
69.00	97704.0
70.20	516.0
73.00	1241.0
74.10	11354.0
75.00	17136.0
76.10	5454.0
77.00	107544.0
78.10	7858.0
79.00	8818.0
80.00	5980.0
81.00	10030.0
81.80	1970.0
83.00	1708.0
85.00	1380.0
85.90	2967.0
87.00	1441.0
87.80	374.0
91.00	2680.0
92.10	2432.0
93.00	16680.0
94.00	1821.0
94.90	337.0
96.00	427.0
97.00	383.0
98.00	15886.0
99.00	9279.0
100.00	592.0
101.00	4195.0
101.90	452.0
103.10	1264.0
104.00	3608.0
105.00	3594.0
105.90	922.0
107.00	44456.0
108.00	6751.0
109.00	1102.0
110.00	60480.0
111.00	11111.0
111.90	1255.0
113.00	481.0
116.00	2162.0
117.00	47928.0
118.00	3492.0
119.00	355.0
120.00	335.0
122.10	2518.0
123.00	4593.0
124.00	2158.0
124.90	1567.0
126.20	893.0
127.00	120000.0
128.00	9285.0
129.00	52792.0
130.00	5044.0
131.00	894.0
131.80	404.0
134.00	1609.0
135.00	4807.0
136.00	2111.0
137.00	2698.0
138.20	389.0
139.00	365.0
139.90	857.0
141.00	8482.0
141.90	2337.0
143.00	1608.0
144.00	632.0
144.90	355.0
145.90	1329.0
146.90	3943.0
148.00	10386.0
149.00	1634.0
149.90	482.0
151.00	1050.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

151.70	341.0
153.00	2247.0
154.00	1612.0
155.00	4087.0
156.00	6746.0
157.10	797.0
157.90	1453.0
158.90	1402.0
159.90	3144.0
161.00	3500.0
162.00	832.0
164.10	534.0
165.00	2959.0
165.90	1998.0
167.00	16736.0
168.00	8144.0
169.00	1471.0
169.80	549.0
170.20	594.0
170.90	863.0
172.00	1582.0
172.90	1632.0
174.00	3608.0
175.00	6671.0
176.10	1359.0
177.00	3007.0
178.00	769.0
178.90	10534.0
180.00	7254.0
181.00	3562.0
181.80	519.0
182.20	508.0
184.00	1633.0
185.00	5769.0
186.00	42568.0
187.00	13980.0
188.00	1720.0
188.90	3900.0
190.00	483.0
191.00	1160.0
191.90	3486.0
193.00	3816.0
194.00	1409.0
194.80	480.0
195.90	6494.0
197.00	2038.0
197.90	261184.0
199.00	16968.0
199.90	1283.0
201.40	1198.0
202.90	2576.0
204.00	11048.0
205.00	19976.0
206.00	81616.0
207.00	12279.0
208.00	3646.0
209.10	1071.0
210.10	1215.0
211.00	3433.0
211.90	512.0
214.90	1285.0
216.00	1588.0
217.00	23904.0
218.00	3075.0
219.00	338.0
220.00	542.0
220.90	7084.0
221.90	2503.0
223.00	5573.0
224.00	45376.0
225.10	10999.0
225.90	1356.0
227.00	24048.0
227.90	3464.0
229.00	4584.0
230.10	707.0
231.00	1281.0
231.90	320.0
232.80	425.0
233.90	2468.0
235.10	1901.0
236.00	1191.0
237.10	1847.0
238.00	379.0
238.90	726.0
240.00	735.0
240.90	1255.0
242.10	2203.0
242.90	2240.0
243.10	2215.0
244.00	31184.0
245.00	4119.0
246.00	9765.0
247.10	1291.0
248.00	506.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

249.00	1327.0
251.10	463.0
252.00	378.0
253.00	1368.0
254.20	1871.0
255.00	179520.0
256.00	25848.0
257.00	2068.0
258.00	12939.0
258.90	2411.0
263.80	407.0
265.00	5999.0
266.00	502.0
266.90	342.0
270.90	688.0
271.90	588.0
273.00	6042.0
274.00	13470.0
275.00	68136.0
276.00	10656.0
277.00	8764.0
278.00	1410.0
283.10	758.0
284.00	670.0
285.00	1776.0
290.00	497.0
291.90	513.0
293.00	1454.0
293.90	514.0
295.10	560.0
296.00	27400.0
297.00	4759.0
302.10	655.0
303.10	2499.0
304.00	915.0
308.90	322.0
310.10	402.0
313.90	1504.0
315.00	2967.0
316.20	1289.0
317.10	432.0
320.90	896.0
322.00	384.0
323.00	7723.0
324.00	2228.0
327.00	1056.0
327.90	466.0
333.00	827.0
333.90	4565.0
335.00	1487.0
339.00	348.0
341.00	522.0
346.00	1378.0
347.00	440.0
352.00	2450.0
353.00	1032.0
354.00	1825.0
355.30	566.0
365.00	11623.0
365.90	1533.0
370.00	309.0
371.00	483.0
372.10	2486.0
373.10	1160.0
382.90	1107.0
384.00	435.0
389.90	653.0
402.10	1404.0
403.00	1579.0
404.10	755.0
421.00	1517.0
422.10	1402.0
423.00	9167.0
424.00	2362.0
425.00	843.0
439.30	358.0
440.10	323.0
441.10	27176.0
442.10	179008.0
443.10	34656.0
444.00	2896.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

m/z	Abundance
38.00	731.0
39.10	4544.0
40.00	321.0
49.20	663.0
50.00	14347.0
51.10	56560.0
52.10	2930.0
56.00	1705.0
57.00	4215.0
61.00	890.0
62.00	959.0
63.00	3386.0
64.00	468.0
65.10	1322.0
67.90	1035.0
69.00	80608.0
70.00	603.0
73.00	1290.0
74.10	9090.0
75.00	14721.0
76.10	5414.0
77.00	86384.0
78.10	5863.0
79.00	6350.0
80.00	4738.0
81.00	7144.0
81.90	1497.0
83.00	1639.0
83.90	362.0
84.90	1692.0
86.00	2702.0
87.00	1074.0
87.90	368.0
91.00	1885.0
92.00	2092.0
93.00	14001.0
93.90	673.0
96.10	461.0
98.00	11805.0
99.00	7755.0
100.00	594.0
101.00	3843.0
103.00	1566.0
104.00	2700.0
105.00	2787.0
105.80	617.0
107.00	35960.0
108.00	4877.0
109.10	1310.0
110.00	48592.0
111.00	9077.0
111.90	1013.0
113.00	392.0
115.90	2079.0
117.00	40000.0
118.00	2746.0
118.90	366.0
120.00	435.0
121.90	2635.0
123.00	4205.0
124.00	1590.0
125.00	1432.0
126.20	894.0
127.00	94200.0
128.00	7421.0
129.00	42384.0
130.00	3114.0
131.00	439.0
132.00	303.0
134.00	1409.0
135.00	3876.0
136.00	1832.0
137.00	1793.0
138.00	373.0
139.00	359.0
140.10	916.0
141.00	7006.0
142.00	2078.0
143.00	1123.0
144.10	769.0
146.00	820.0
147.00	2797.0
148.00	9930.0
149.00	1862.0
150.90	589.0
151.80	414.0
152.90	2071.0
154.00	1792.0
154.90	3692.0
156.10	5591.0
157.00	616.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

158.10	824.0
159.10	1056.0
160.00	2233.0
161.00	3033.0
162.00	1263.0
165.00	2382.0
165.90	2142.0
167.00	14604.0
168.00	7893.0
169.10	1145.0
170.10	553.0
170.80	454.0
171.90	1133.0
172.90	1942.0
174.00	3412.0
175.10	5631.0
175.90	547.0
176.80	1559.0
177.90	1084.0
178.90	10787.0
180.00	6244.0
180.90	3130.0
182.00	811.0
182.60	327.0
184.10	840.0
185.00	4407.0
186.00	39280.0
187.00	11307.0
188.00	1208.0
188.90	2687.0
190.00	371.0
191.10	892.0
192.00	4587.0
193.00	4300.0
194.00	800.0
195.00	323.0
195.90	5303.0
197.10	1901.0
197.90	233344.0
198.90	13559.0
199.90	1123.0
201.30	1189.0
202.90	2303.0
204.00	11780.0
205.00	15776.0
206.00	71648.0
207.00	9883.0
208.00	2884.0
209.00	1495.0
210.00	1902.0
211.00	3128.0
212.90	351.0
215.00	941.0
216.00	1691.0
217.00	20208.0
217.90	2491.0
218.90	365.0
221.10	7380.0
221.70	2531.0
223.00	3859.0
224.00	43240.0
225.00	10649.0
226.00	1338.0
227.00	23304.0
228.00	3144.0
229.00	4641.0
229.80	972.0
231.00	992.0
232.90	508.0
234.00	1056.0
234.80	1454.0
236.10	1189.0
237.10	1600.0
239.00	718.0
239.80	392.0
240.20	411.0
241.10	1197.0
242.00	2412.0
242.90	2147.0
244.00	30504.0
245.00	3477.0
246.00	8750.0
247.00	1695.0
248.90	1490.0
250.80	397.0
251.80	342.0
252.10	331.0
252.90	534.0
254.00	1536.0
255.00	162176.0
256.00	24488.0
256.90	1707.0
258.00	12188.0
258.90	2171.0
260.00	404.0

Instrument :

BNA_M

ClientSampleId :

DFTPP18

264.00	439.0
265.00	4504.0
266.10	649.0
269.90	321.0
270.90	737.0
272.20	509.0
273.00	4058.0
274.00	13542.0
275.00	67632.0
276.00	8555.0
277.00	6675.0
277.90	893.0
280.90	447.0
282.00	443.0
282.90	682.0
284.00	513.0
285.00	1566.0
289.00	412.0
289.90	327.0
293.00	1547.0
295.10	795.0
296.00	26336.0
297.00	4292.0
300.90	493.0
301.80	312.0
302.90	2755.0
304.00	692.0
313.90	737.0
315.10	2521.0
316.00	1441.0
317.10	391.0
321.00	747.0
322.00	496.0
323.00	7362.0
324.00	1536.0
327.10	1099.0
328.10	727.0
332.10	572.0
333.10	814.0
334.00	4419.0
335.10	1123.0
341.10	1002.0
346.00	2478.0
346.80	301.0
351.90	2225.0
353.00	1871.0
354.00	2391.0
355.10	419.0
365.00	12041.0
366.00	1431.0
371.20	330.0
372.10	3518.0
373.10	874.0
383.10	997.0
383.90	355.0
401.90	1050.0
403.10	1753.0
404.10	610.0
421.00	1222.0
422.10	1363.0
423.00	10877.0
424.00	2924.0
425.10	358.0
439.00	653.0
439.90	480.0
441.10	28248.0
442.00	201152.0
443.10	39936.0
444.00	5058.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18

Scan 2451 (17.303 min): BM010683.D
DFTPP18

m/z	Abundance
73.10	365.0
76.80	302.0
193.10	316.0
207.00	814.0
281.10	323.0

Instrument :
BNA_M
ClientSampleId :
DFTPP18