

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM062319\
 Data File : BM021137.D
 Acq On : 23 Jun 2019 18:06
 Operator : HP/JU
 Sample : AR1248 500PPM
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 AR1248 500PPM

Quant Time: Jun 24 00:56:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 20 03:45:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	273427	20.00	ng/ul	-0.02
18) Naphthalene-d8	10.56	136	1046508	20.00	ng/ul	-0.02
35) Acenaphthene-d10	14.41	164	646330	20.00	ng/ul	-0.02
61) Phenanthrene-d10	17.16	188	1471783	20.00	ng/ul	-0.02
77) Chrysene-d12	21.34	240	1050518	20.00	ng/ul	-0.01
85) Perylene-d12	23.61	264	1101941	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.11	96	239	0.04	ng/uL	-0.18
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul	
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul	
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul	
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	10.56	131	1097	0.05	ng/ul	-0.16
43) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul	
46) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul	
51) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
57) Fluorene-d10	15.42	176	2492	0.05	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
70) Anthracene-d10	17.16	188	1471783	19.24	ng/ul	-0.12
78) Pyrene-d10	19.56	212	1513	0.02	ng/ul	-0.01
89) Benzo(a)pyrene-d12	23.61	264	1074532	16.53	ng/ul	0.13

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM062319\
Data File : BM021137.D
Acq On : 23 Jun 2019 18:06
Operator : HP/JU
Sample : AR1248 500PPM
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
AR1248 500PPM

Quant Time: Jun 24 00:56:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 20 03:45:02 2019
Response via : Initial Calibration

