


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062320\
 Data File : BM026494.D
 Acq On : 23 Jun 2020 07:59
 Operator : JU/CG
 Sample : DFTPP03
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP03

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061220MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jun 23 10:25:31 2020



AutoFind: Scans 2446, 2447, 2448; Background Corrected with Scan 2440

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.1	90170	PASS
68	69	0.00	2	1.6	1398	PASS
69	198	0.00	100	44.8	85685	PASS
70	69	0.00	2	0.6	490	PASS
127	198	10	80	46.1	88237	PASS
197	198	0.00	2	0.5	969	PASS
198	198	100	100	100.0	191349	PASS
199	198	5	9	6.8	12919	PASS
275	198	10	60	29.1	55629	PASS
365	198	1	100	4.2	8086	PASS
441	198	0.01	100	12.8	24426	PASS
442	198	50	100	86.7	165909	PASS
443	442	15	24	19.2	31805	PASS

Scan 2440 (17.239 min): BM026494.D
DFTPP03

m/z	Abundance
36.00	170.0
40.00	685.0
44.00	279.0
73.20	458.0
147.00	187.0
207.10	834.0
208.00	150.0
281.10	150.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

m/z	Abundance
37.10	563.0
38.00	1612.0
39.10	8708.0
40.00	1121.0
41.20	331.0
43.90	400.0
45.00	225.0
49.00	820.0
50.10	33320.0
51.10	119240.0
52.10	6346.0
53.00	370.0
55.10	631.0
56.00	3482.0
57.00	6950.0
58.00	308.0
58.90	151.0
61.10	1117.0
62.10	1796.0
63.00	5175.0
64.10	549.0
65.10	2214.0
65.90	169.0
67.10	156.0
68.00	1913.0
69.00	113192.0
70.10	731.0
73.10	1381.0
74.10	12039.0
75.00	17184.0
76.10	6034.0
77.10	122600.0
78.10	7680.0
79.00	8550.0
80.00	6267.0
81.10	8381.0
82.00	2302.0
83.00	1551.0
83.90	355.0
85.00	1382.0
86.00	2664.0
87.00	973.0
88.00	419.0
89.00	202.0
91.00	1927.0
92.00	2084.0
93.00	14765.0
94.00	1063.0
95.10	260.0
96.00	679.0
97.10	430.0
98.00	12548.0
99.00	8865.0
100.00	1029.0
101.00	4649.0
103.00	1637.0
104.00	2889.0
105.00	2961.0
106.10	984.0
107.00	34160.0
108.00	4948.0
109.00	1089.0
110.00	59176.0
111.00	9832.0
111.90	1281.0
113.00	511.0
115.00	161.0
116.00	1784.0
117.00	35088.0
118.00	2395.0
118.90	230.0
120.00	314.0
120.90	279.0
122.00	2463.0
123.00	3387.0
124.00	1853.0
125.10	1442.0
126.20	584.0
127.00	107928.0
128.00	7864.0
129.00	44912.0
130.00	3935.0
131.00	806.0
132.00	549.0
132.90	297.0
134.00	1168.0
135.00	3525.0
136.00	1417.0
137.10	2510.0
138.00	527.0
139.00	211.0
139.90	523.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

141.00	6292.0
142.00	2112.0
143.10	1389.0
144.10	223.0
145.10	458.0
146.10	1026.0
147.00	3764.0
148.00	8179.0
149.10	1604.0
150.00	452.0
151.00	724.0
151.60	360.0
152.20	274.0
153.00	1691.0
154.00	1535.0
155.00	3220.0
156.00	4805.0
157.10	885.0
158.10	1252.0
159.00	807.0
160.10	2075.0
160.90	2681.0
162.00	838.0
163.00	299.0
164.10	299.0
165.00	2582.0
166.00	1791.0
167.10	11982.0
168.00	8491.0
169.00	1185.0
170.10	356.0
170.90	586.0
171.90	1127.0
173.00	1446.0
174.00	2441.0
175.00	4929.0
176.10	1564.0
177.00	2489.0
178.00	804.0
179.00	9556.0
180.00	7192.0
181.00	2650.0
182.10	328.0
183.10	232.0
184.00	886.0
185.00	3741.0
186.00	33112.0
187.00	9537.0
188.00	1171.0
189.00	2391.0
190.00	257.0
191.10	1007.0
192.00	2690.0
193.00	3653.0
194.10	568.0
195.00	367.0
196.00	6506.0
196.90	1149.0
198.00	222976.0
199.00	15393.0
200.00	1339.0
201.50	762.0
203.00	1909.0
204.00	8988.0
205.10	14206.0
206.10	61688.0
207.10	9049.0
208.00	2348.0
209.00	876.0
210.00	915.0
211.00	2732.0
211.80	515.0
212.90	227.0
214.90	778.0
216.00	1955.0
217.00	17800.0
218.00	2116.0
219.10	152.0
221.00	8262.0
221.80	2676.0
223.00	3850.0
224.00	33520.0
225.00	9420.0
226.10	1444.0
227.00	15694.0
228.00	2149.0
229.00	2870.0
230.00	384.0
231.00	1285.0
231.80	183.0
232.90	390.0
233.90	1089.0
234.90	1219.0
235.90	887.0
237.00	2012.0
238.00	312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

239.10	632.0
239.90	672.0
240.90	912.0
242.00	1810.0
243.10	1863.0
244.10	26520.0
245.00	3560.0
246.00	5511.0
247.00	1052.0
248.00	225.0
249.00	994.0
250.90	280.0
252.10	201.0
252.90	655.0
254.10	976.0
255.00	137536.0
256.00	20928.0
257.10	1776.0
258.00	8546.0
259.00	1281.0
260.00	226.0
261.00	215.0
264.10	231.0
265.00	3423.0
266.00	481.0
269.90	185.0
270.90	297.0
272.00	567.0
273.00	4213.0
274.00	9894.0
275.00	59968.0
276.00	7881.0
277.00	5675.0
278.00	983.0
279.00	276.0
280.90	377.0
281.90	261.0
282.90	680.0
284.00	563.0
285.00	922.0
286.00	227.0
288.90	164.0
290.90	165.0
291.90	307.0
293.00	1188.0
294.10	267.0
295.10	293.0
296.00	19096.0
297.00	2514.0
298.00	195.0
300.90	269.0
302.00	415.0
302.90	2020.0
304.10	617.0
308.00	325.0
308.90	169.0
310.20	198.0
314.00	712.0
314.90	1849.0
316.00	1165.0
316.90	191.0
321.00	501.0
322.00	355.0
323.00	5672.0
323.90	1101.0
325.00	166.0
327.00	778.0
327.90	504.0
332.10	332.0
333.10	532.0
334.00	3877.0
335.10	1012.0
341.00	698.0
346.00	978.0
346.90	253.0
352.00	1536.0
353.00	1064.0
354.00	1859.0
355.10	546.0
365.00	7920.0
365.90	1089.0
370.00	213.0
371.10	510.0
372.10	2717.0
373.00	582.0
383.00	828.0
384.00	313.0
389.90	234.0
391.10	346.0
391.90	195.0
401.90	930.0
403.00	1738.0
404.00	759.0
405.10	209.0
421.00	1285.0
422.00	1261.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

423.00	8500.0
424.00	3033.0
424.90	388.0
441.00	22288.0
442.00	147456.0
443.00	29280.0
444.00	2624.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

m/z	Abundance
36.00	153.0
37.00	447.0
38.10	1129.0
39.10	7805.0
40.00	875.0
41.00	270.0
44.10	306.0
45.00	279.0
49.00	864.0
50.10	27400.0
51.10	103280.0
52.10	5613.0
53.00	268.0
55.00	589.0
56.00	2834.0
57.00	6504.0
57.90	289.0
61.00	1464.0
62.00	1581.0
63.00	3960.0
64.10	485.0
65.10	1981.0
65.90	207.0
66.90	179.0
68.10	1630.0
69.00	96936.0
70.00	433.0
73.10	1449.0
74.10	11089.0
75.00	14826.0
76.10	5690.0
77.10	109184.0
78.10	6648.0
79.10	6917.0
80.00	6211.0
81.00	7676.0
82.00	1919.0
83.10	1575.0
83.90	351.0
85.00	1296.0
86.00	2341.0
87.10	1222.0
87.90	246.0
91.10	2092.0
92.00	2023.0
93.00	12756.0
94.00	1121.0
95.00	257.0
96.10	633.0
97.20	174.0
98.00	10151.0
99.00	7776.0
100.00	680.0
101.00	4206.0
102.10	296.0
103.00	1425.0
104.00	2609.0
105.00	2607.0
106.10	829.0
107.00	33112.0
108.00	4590.0
109.00	808.0
110.00	56128.0
111.00	8336.0
112.00	1119.0
113.00	354.0
115.20	236.0
116.00	1704.0
117.00	33432.0
118.00	2452.0
119.10	374.0
120.00	578.0
121.00	168.0
122.00	1776.0
123.00	3507.0
124.00	1660.0
125.10	1101.0
126.20	489.0
127.00	104784.0
128.00	7674.0
129.00	42072.0
130.00	3708.0
131.10	869.0
132.00	387.0
133.00	273.0
134.00	1357.0
135.00	3577.0
136.00	2028.0
137.10	2160.0
138.00	493.0
139.00	338.0
140.00	440.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

141.00	6490.0
142.00	1858.0
142.90	1531.0
144.00	496.0
144.80	367.0
146.00	1233.0
147.00	3345.0
148.00	7943.0
149.00	1598.0
149.90	491.0
151.20	750.0
152.10	487.0
153.00	1989.0
154.00	1283.0
155.00	3506.0
156.00	4339.0
157.00	816.0
157.90	1113.0
158.90	719.0
160.00	1835.0
161.00	2587.0
162.00	617.0
163.00	214.0
164.00	306.0
165.00	2391.0
166.00	1733.0
167.00	13421.0
168.00	7641.0
169.00	947.0
170.00	542.0
171.00	537.0
171.90	1005.0
173.00	1643.0
174.00	2761.0
175.10	4760.0
176.00	1241.0
177.00	2771.0
177.90	560.0
179.00	8569.0
180.00	6669.0
180.90	2705.0
182.10	624.0
183.00	212.0
184.00	908.0
185.00	4296.0
186.00	32936.0
187.10	9628.0
188.00	1142.0
188.90	2161.0
189.90	336.0
191.00	908.0
192.00	3177.0
193.00	3404.0
193.90	597.0
195.00	427.0
196.00	6748.0
196.90	1119.0
198.00	228352.0
199.00	14929.0
200.00	1531.0
201.40	1218.0
203.10	1899.0
204.10	9283.0
205.10	15510.0
206.00	63824.0
207.10	8663.0
208.10	2434.0
209.00	851.0
210.00	958.0
211.00	2312.0
211.80	451.0
212.80	185.0
215.00	718.0
216.00	1286.0
217.00	18344.0
218.00	2600.0
219.00	186.0
220.20	394.0
221.00	9738.0
221.90	2883.0
223.00	4206.0
224.00	35960.0
225.00	9747.0
226.00	1144.0
227.00	17072.0
228.00	2460.0
229.00	2921.0
230.00	411.0
230.90	1427.0
232.10	204.0
232.80	243.0
234.00	1309.0
235.00	1304.0
236.00	916.0
237.00	1653.0
238.10	334.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

239.00	690.0
240.00	557.0
240.90	1041.0
242.00	2411.0
243.00	1854.0
244.10	27552.0
245.00	3094.0
246.00	6021.0
246.90	1243.0
248.00	284.0
248.80	1042.0
250.10	289.0
251.00	234.0
251.90	203.0
253.00	637.0
254.00	1125.0
255.00	152512.0
256.00	22048.0
257.00	1919.0
258.00	8962.0
259.00	1745.0
260.00	255.0
261.00	287.0
263.90	172.0
265.00	3754.0
266.00	657.0
267.00	273.0
271.00	463.0
272.00	397.0
273.00	4593.0
274.00	13012.0
275.00	65832.0
276.00	8648.0
277.00	6685.0
278.00	1220.0
279.00	302.0
281.10	363.0
282.10	210.0
283.00	784.0
283.90	700.0
285.10	1118.0
286.00	318.0
291.10	184.0
292.10	219.0
292.90	1609.0
293.90	553.0
294.90	319.0
296.00	21408.0
297.00	2565.0
298.10	218.0
301.00	236.0
301.90	236.0
303.00	2512.0
304.00	621.0
308.00	679.0
309.00	221.0
310.00	272.0
313.00	201.0
314.00	1138.0
315.00	2320.0
316.00	1454.0
317.00	284.0
321.00	573.0
322.00	264.0
323.10	6235.0
324.00	1011.0
324.90	217.0
326.00	253.0
327.00	1591.0
327.90	629.0
331.90	492.0
333.00	491.0
334.00	4045.0
334.90	1116.0
339.90	191.0
341.10	758.0
341.90	181.0
346.00	1433.0
347.00	319.0
352.10	1881.0
353.00	1180.0
354.10	2388.0
354.90	306.0
365.00	9904.0
366.00	1385.0
369.90	255.0
371.00	528.0
372.00	3472.0
373.00	791.0
382.90	1055.0
383.90	356.0
390.00	452.0
390.90	293.0
391.90	228.0
400.90	163.0
402.00	1789.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

403.00	1859.0
404.10	662.0
405.10	168.0
421.00	1650.0
422.00	1744.0
423.00	11478.0
424.00	3226.0
425.00	526.0
440.20	214.0
441.00	30384.0
442.00	204224.0
443.10	38432.0
444.00	3403.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

m/z	Abundance
36.90	211.0
38.10	515.0
39.10	3872.0
40.00	886.0
41.10	180.0
44.00	234.0
45.00	244.0
49.00	426.0
50.10	12883.0
51.10	47992.0
52.10	2268.0
55.00	341.0
56.00	1608.0
57.00	3107.0
57.90	230.0
61.00	722.0
62.00	575.0
63.10	1824.0
64.00	324.0
65.00	817.0
68.10	652.0
69.00	46928.0
70.00	308.0
73.00	989.0
74.10	5256.0
75.00	7238.0
76.00	2707.0
77.10	50864.0
78.10	3551.0
79.00	3729.0
80.00	2655.0
81.00	4029.0
82.00	902.0
83.00	781.0
84.90	898.0
86.00	1177.0
87.00	311.0
88.10	222.0
91.00	760.0
92.00	999.0
93.00	6590.0
93.90	388.0
95.10	166.0
96.00	411.0
98.00	5238.0
99.00	4059.0
100.00	368.0
101.00	2337.0
102.90	563.0
104.00	1735.0
105.00	1501.0
106.10	392.0
107.00	15876.0
108.00	2353.0
109.00	536.0
110.00	27296.0
111.00	4286.0
111.80	332.0
113.10	202.0
116.10	850.0
117.00	17208.0
118.00	1288.0
118.90	204.0
120.00	197.0
122.00	958.0
123.00	1649.0
123.90	891.0
125.00	708.0
126.20	411.0
127.00	52000.0
128.00	3768.0
129.00	22456.0
130.00	2076.0
131.00	348.0
131.80	240.0
133.90	670.0
135.00	1782.0
136.00	744.0
137.00	880.0
137.70	243.0
139.90	304.0
141.10	2860.0
142.00	987.0
143.00	692.0
143.90	191.0
144.90	187.0
146.00	591.0
147.00	1697.0
148.00	4521.0
149.00	1006.0
150.20	201.0
151.10	505.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

151.80	323.0
153.00	993.0
154.00	654.0
155.10	1847.0
156.10	2477.0
157.10	463.0
157.90	789.0
159.10	469.0
159.90	968.0
161.00	1212.0
162.10	386.0
162.80	233.0
164.10	204.0
165.00	1254.0
166.00	927.0
167.00	7013.0
168.00	3792.0
169.00	701.0
170.10	224.0
171.00	262.0
171.90	475.0
173.00	897.0
174.00	1465.0
175.00	2399.0
176.00	794.0
177.00	1395.0
178.10	322.0
179.00	5024.0
180.00	3320.0
181.00	1690.0
182.00	388.0
184.00	414.0
185.00	2395.0
186.00	17320.0
187.00	5349.0
188.00	711.0
189.00	1273.0
190.10	151.0
190.90	657.0
192.00	1755.0
193.00	1562.0
194.00	366.0
194.90	279.0
196.00	3775.0
197.00	639.0
198.00	122720.0
199.00	8436.0
199.90	520.0
201.60	427.0
202.10	165.0
202.90	1014.0
204.00	5240.0
205.00	8707.0
206.10	35656.0
207.10	5026.0
208.00	1778.0
209.00	449.0
210.10	587.0
211.00	1575.0
212.00	289.0
214.90	348.0
216.00	971.0
217.00	10216.0
217.90	1293.0
218.90	159.0
221.00	4870.0
221.70	1546.0
223.00	2155.0
224.00	20360.0
225.10	4914.0
226.00	628.0
227.00	9103.0
227.90	1094.0
229.00	2031.0
230.00	286.0
230.90	595.0
232.90	192.0
233.90	522.0
235.10	683.0
235.90	520.0
237.00	1031.0
239.00	430.0
240.10	306.0
241.00	524.0
242.00	1009.0
243.00	1267.0
244.00	16311.0
245.00	1837.0
246.00	3631.0
247.00	817.0
248.90	626.0
251.00	218.0
252.00	188.0
253.00	370.0
254.10	576.0
255.00	91792.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03

256.00	13161.0
257.00	889.0
258.00	5992.0
259.00	776.0
265.10	2357.0
265.80	341.0
271.00	196.0
272.00	248.0
273.00	2755.0
274.00	7692.0
275.00	41088.0
276.00	5389.0
277.00	3991.0
278.00	641.0
281.00	379.0
281.90	282.0
283.00	462.0
284.00	284.0
285.00	577.0
292.00	241.0
292.90	811.0
293.90	160.0
295.00	397.0
296.00	12873.0
297.00	1936.0
301.00	181.0
302.00	216.0
303.00	1900.0
304.00	450.0
310.00	345.0
314.00	676.0
315.00	1567.0
316.00	1048.0
321.00	406.0
323.00	4098.0
324.10	544.0
327.00	1030.0
327.90	420.0
332.10	281.0
333.00	505.0
334.10	3047.0
335.00	555.0
341.00	643.0
342.10	228.0
346.00	1022.0
346.90	334.0
352.00	1267.0
353.00	1020.0
354.00	1265.0
354.90	234.0
365.00	6434.0
366.00	889.0
369.90	211.0
371.00	480.0
372.10	2344.0
373.00	702.0
383.00	600.0
383.90	322.0
390.10	363.0
390.90	199.0
401.90	1284.0
403.00	1632.0
404.00	396.0
421.00	1340.0
422.10	1231.0
423.00	8147.0
423.90	2265.0
425.00	159.0
440.00	160.0
441.00	20608.0
442.00	146048.0
443.00	27704.0
444.00	2783.0
444.90	171.0

Instrument :
BNA_M
ClientSampleId :
DFTPP03