


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062416\
 Data File : BM005962.D
 Acq On : 24 Jun 2016 11:26
 Operator : UM/SJ
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
 Title : SVOA CALIBRATION
 Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2536, 2537, 2538; Background Corrected with Scan 2530

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	97178	PASS
68	69	0.00	2	1.8	1593	PASS
69	198	0.00	100	40.7	90346	PASS
70	69	0.00	2	0.6	533	PASS
127	198	10	80	46.8	103808	PASS
197	198	0.00	2	0.3	718	PASS
198	198	100	100	100.0	221824	PASS
199	198	5	9	6.7	14934	PASS
275	198	10	60	25.0	55466	PASS
365	198	1	100	3.0	6708	PASS
441	443	0.01	100	79.3	26450	PASS
442	198	50	100	78.9	174970	PASS
443	442	15	24	19.1	33349	PASS

m/z	Abundance
36.90	443.0
37.90	1320.0
39.10	7391.0
40.00	1464.0
41.00	312.0
42.90	328.0
44.00	1724.0
49.00	785.0
50.10	24464.0
51.00	98576.0
52.00	5527.0
55.00	477.0
56.00	2712.0
57.00	6707.0
57.90	404.0
61.00	1418.0
62.00	1395.0
63.00	3742.0
64.10	614.0
65.00	1681.0
68.00	1440.0
69.00	87752.0
70.00	412.0
73.00	1181.0
74.10	9399.0
75.00	15768.0
76.00	4926.0
77.00	101016.0
78.00	7375.0
79.00	6465.0
80.00	5303.0
81.00	7560.0
82.00	2208.0
83.00	1497.0
85.00	1268.0
86.00	1965.0
87.00	1075.0
87.90	544.0
91.00	1854.0
92.00	1786.0
93.00	12270.0
94.00	850.0
96.00	741.0
98.00	9847.0
99.00	8191.0
100.00	659.0
101.00	3777.0
102.90	1395.0
103.90	2024.0
105.00	2011.0
106.00	786.0
107.00	31896.0
108.00	4291.0
109.00	775.0
110.00	53400.0
111.00	9878.0
112.00	756.0
112.90	388.0
116.00	1500.0
117.00	25680.0
118.00	2191.0
118.90	331.0
119.90	357.0
122.00	2082.0
123.00	3008.0
124.00	1282.0
124.90	1040.0
127.00	99824.0
128.00	7583.0
129.00	37072.0
130.00	3262.0
131.00	585.0
131.90	363.0
133.00	423.0
133.90	876.0
135.00	3260.0
136.00	1435.0
137.00	1577.0
137.90	369.0
139.00	332.0
140.10	463.0
141.00	4742.0
141.90	1288.0
142.90	1010.0
145.00	332.0
145.90	692.0
147.00	2486.0
148.00	5769.0
149.00	1704.0
150.00	395.0
151.00	729.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

152.90	1522.0
153.90	1006.0
155.00	2809.0
156.00	4020.0
157.10	325.0
158.00	813.0
159.00	681.0
159.90	1686.0
161.00	2436.0
161.90	720.0
162.90	310.0
163.80	319.0
165.00	1627.0
166.00	1377.0
167.00	9409.0
168.00	7458.0
169.00	1058.0
170.80	506.0
172.10	1173.0
173.00	1161.0
174.00	2595.0
175.00	3966.0
176.00	961.0
177.00	2048.0
177.90	584.0
178.90	7682.0
180.00	5615.0
181.00	2524.0
182.00	515.0
184.10	529.0
185.00	3575.0
186.00	25664.0
187.00	8597.0
188.00	969.0
189.00	1574.0
191.00	765.0
192.00	2459.0
193.00	2973.0
194.00	672.0
195.20	397.0
196.00	5241.0
196.90	1172.0
198.00	192448.0
199.00	13371.0
199.90	1080.0
201.40	797.0
203.10	1368.0
204.00	7673.0
205.00	12592.0
206.00	49920.0
207.00	8221.0
208.00	2136.0
208.90	957.0
210.00	780.0
211.00	2170.0
212.10	460.0
215.10	453.0
216.00	919.0
217.00	13760.0
218.00	1681.0
221.00	6873.0
223.00	3067.0
224.00	27320.0
225.00	7121.0
226.10	1039.0
227.00	12052.0
228.00	2361.0
229.00	2151.0
231.00	1242.0
234.00	724.0
235.00	975.0
236.00	683.0
237.00	1499.0
239.00	388.0
241.00	662.0
242.00	1667.0
243.00	1275.0
244.00	21016.0
245.10	2406.0
246.00	4361.0
246.90	1016.0
249.00	1134.0
253.00	500.0
254.10	518.0
255.00	110960.0
256.00	15224.0
257.00	1619.0
258.00	6037.0
259.00	1056.0
265.00	2555.0
266.00	335.0
271.00	354.0
273.00	2784.0
274.00	8583.0
275.00	45632.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

276.00	6937.0
277.00	3285.0
278.00	817.0
283.00	573.0
285.10	536.0
292.90	822.0
293.90	372.0
296.00	12968.0
297.00	1694.0
302.00	622.0
303.00	1460.0
304.00	403.0
313.90	757.0
315.00	1158.0
316.10	1039.0
321.00	357.0
323.00	4272.0
324.10	504.0
326.90	699.0
328.10	496.0
332.10	371.0
332.90	365.0
334.00	2801.0
335.00	812.0
341.00	749.0
346.00	794.0
347.00	347.0
352.00	1141.0
353.00	805.0
354.00	1141.0
365.00	5033.0
366.00	729.0
370.90	471.0
372.00	2173.0
373.00	444.0
383.00	505.0
390.00	303.0
402.00	942.0
403.00	1266.0
404.10	569.0
421.00	872.0
422.00	1195.0
423.00	6352.0
424.00	2059.0
425.10	481.0
441.00	18176.0
442.10	126832.0
443.10	24072.0
444.10	2390.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	500.0
38.00	1386.0
39.00	7882.0
40.00	1440.0
41.10	489.0
44.00	1381.0
49.00	911.0
50.00	28976.0
51.00	115112.0
52.10	5940.0
55.00	421.0
56.00	3380.0
57.00	7486.0
61.00	1292.0
62.00	1564.0
63.00	4613.0
63.90	873.0
65.00	2579.0
68.00	1908.0
69.00	108456.0
70.10	636.0
73.10	1254.0
74.00	10219.0
75.00	17888.0
76.00	5928.0
77.10	124144.0
78.00	8368.0
79.00	7486.0
80.00	6347.0
81.00	9118.0
82.00	2317.0
83.10	2117.0
84.00	387.0
84.90	1726.0
86.00	2345.0
87.00	1295.0
87.90	578.0
91.00	1767.0
92.00	2308.0
93.00	14625.0
94.00	1221.0
96.00	731.0
98.00	10661.0
99.00	10394.0
100.00	720.0
101.00	4312.0
103.00	1182.0
104.00	3135.0
105.00	3132.0
106.10	1185.0
107.00	39192.0
108.00	5601.0
109.00	902.0
110.00	67128.0
111.00	10686.0
112.00	1159.0
113.00	569.0
116.00	1843.0
117.00	32344.0
118.00	2320.0
120.00	410.0
122.00	2798.0
123.00	3576.0
124.00	1757.0
125.00	1910.0
126.10	386.0
127.00	121440.0
128.00	9512.0
129.00	46152.0
130.00	4189.0
131.00	888.0
132.00	443.0
133.10	467.0
133.90	1243.0
135.00	3697.0
135.90	1665.0
136.90	2327.0
138.10	481.0
140.00	558.0
141.00	5270.0
142.00	1925.0
143.00	1648.0
143.90	459.0
146.00	809.0
147.00	3318.0
148.00	7726.0
149.00	1734.0
149.90	550.0
151.00	864.0
153.00	2159.0
154.00	1466.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

155.00	3398.0
156.00	5538.0
156.90	1024.0
158.00	1097.0
158.90	786.0
160.00	1814.0
161.00	2851.0
162.00	668.0
163.00	494.0
165.00	2413.0
166.00	2201.0
167.00	13504.0
168.00	7710.0
169.00	1619.0
171.00	691.0
172.00	1105.0
173.00	1602.0
173.90	2902.0
175.00	4560.0
175.90	1606.0
177.00	2618.0
178.00	594.0
179.00	9502.0
180.00	7417.0
181.00	3038.0
182.00	452.0
183.90	697.0
185.00	4573.0
186.00	36752.0
187.00	10477.0
188.00	1125.0
189.00	2598.0
190.90	1054.0
192.00	2851.0
193.00	4109.0
194.00	669.0
194.90	303.0
196.00	7064.0
197.10	1107.0
198.00	262656.0
199.00	17784.0
200.00	1445.0
200.80	369.0
201.40	814.0
203.00	2055.0
204.00	9644.0
205.00	15158.0
206.10	67664.0
207.10	9562.0
208.00	2255.0
209.00	824.0
209.90	945.0
211.10	2918.0
215.00	1058.0
216.00	1835.0
217.00	18224.0
218.00	2435.0
220.30	359.0
221.00	9192.0
222.00	2252.0
223.00	4436.0
224.00	37600.0
225.00	9589.0
226.00	1544.0
227.00	16800.0
228.00	2142.0
229.00	3716.0
229.90	356.0
231.00	1231.0
234.00	969.0
235.00	1399.0
236.00	904.0
237.00	1816.0
237.90	300.0
239.00	472.0
240.00	402.0
240.90	923.0
242.00	2169.0
243.00	2379.0
244.10	29424.0
245.10	3739.0
246.00	5620.0
247.00	1345.0
249.00	941.0
252.90	632.0
254.00	996.0
255.00	150656.0
256.00	21544.0
257.00	1625.0
258.10	9202.0
259.00	1396.0
265.00	3946.0
266.00	534.0
272.00	459.0
273.00	4817.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

274.00	11133.0
275.00	63632.0
276.00	8405.0
277.00	5447.0
277.90	1080.0
279.10	306.0
283.10	793.0
284.10	345.0
285.00	652.0
288.90	366.0
292.90	1572.0
295.10	418.0
296.00	18912.0
297.00	2486.0
303.00	2704.0
304.00	954.0
314.00	940.0
314.90	2206.0
316.00	1217.0
321.10	641.0
323.10	5903.0
324.00	1015.0
326.90	1073.0
328.00	496.0
332.00	559.0
334.00	3839.0
335.00	1050.0
341.00	663.0
346.00	1474.0
347.00	324.0
352.00	1891.0
353.00	1384.0
354.00	2102.0
355.00	541.0
365.00	8160.0
366.00	1300.0
371.00	568.0
372.10	2800.0
373.00	563.0
383.10	745.0
383.90	495.0
390.00	344.0
391.00	446.0
402.00	1426.0
403.00	1753.0
404.00	447.0
421.00	1551.0
422.00	1467.0
423.10	10224.0
424.10	3349.0
425.10	850.0
441.00	30112.0
442.10	198592.0
443.10	38320.0
444.00	3692.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
38.00	1227.0
39.00	5779.0
40.00	1055.0
41.00	373.0
44.00	1221.0
49.00	738.0
50.00	20216.0
51.00	77848.0
52.00	4097.0
55.10	578.0
56.00	2402.0
57.00	5671.0
58.00	469.0
61.00	846.0
62.00	1094.0
63.00	3037.0
64.00	583.0
65.00	1227.0
67.90	1433.0
69.00	74832.0
70.00	551.0
73.00	820.0
74.00	7736.0
75.00	12753.0
76.00	4218.0
77.00	87040.0
78.10	6165.0
79.00	5315.0
80.00	4405.0
81.00	6308.0
82.00	1749.0
83.00	1441.0
84.90	1447.0
86.00	1781.0
87.00	912.0
88.10	377.0
91.00	1375.0
92.00	1449.0
93.00	10328.0
93.90	500.0
96.00	439.0
98.00	8306.0
99.00	6577.0
100.00	600.0
101.00	3415.0
101.90	311.0
103.00	1074.0
103.90	1887.0
105.00	2378.0
105.90	712.0
107.00	27432.0
108.00	3509.0
108.90	663.0
110.00	46696.0
111.00	7145.0
112.00	1022.0
116.00	1374.0
117.00	23312.0
118.00	1794.0
119.00	391.0
119.90	332.0
122.00	1345.0
123.00	2463.0
123.90	1141.0
125.00	1416.0
126.10	341.0
127.00	90160.0
128.00	6597.0
129.00	36848.0
129.90	2902.0
131.10	633.0
132.10	332.0
134.00	1264.0
135.00	2958.0
136.00	1344.0
137.00	1764.0
137.90	353.0
140.10	483.0
141.00	4729.0
142.00	1375.0
143.00	1204.0
144.00	347.0
146.10	870.0
147.00	2830.0
148.00	5603.0
149.00	1718.0
151.00	478.0
152.00	491.0
153.00	1698.0
154.00	1016.0
155.00	2998.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

156.00	3569.0
157.00	832.0
158.00	987.0
159.00	739.0
160.00	1711.0
161.00	2603.0
161.90	488.0
163.00	320.0
165.00	1611.0
166.00	1399.0
167.00	10534.0
168.00	6477.0
169.00	1038.0
171.00	461.0
171.90	942.0
173.00	887.0
174.00	1859.0
175.10	3950.0
176.00	1100.0
176.90	1473.0
178.00	685.0
179.00	6911.0
180.00	5823.0
181.00	2417.0
182.10	465.0
184.00	696.0
185.00	4014.0
186.00	28160.0
187.00	7608.0
188.10	974.0
189.00	1816.0
191.00	1240.0
192.00	1917.0
193.00	3177.0
194.00	344.0
196.00	5518.0
196.90	982.0
198.00	210368.0
199.00	13649.0
200.00	929.0
201.50	642.0
202.90	1476.0
204.00	7024.0
205.00	14217.0
206.00	53320.0
207.00	7384.0
208.00	1796.0
209.00	1114.0
210.00	917.0
211.10	1848.0
211.70	547.0
214.90	513.0
216.00	1467.0
217.00	13495.0
217.90	2097.0
220.20	425.0
221.00	5860.0
221.90	2155.0
223.00	3795.0
224.00	32432.0
225.00	7402.0
226.00	858.0
227.00	12312.0
228.00	2243.0
229.00	2443.0
229.90	463.0
231.00	1015.0
231.90	324.0
233.90	607.0
234.90	840.0
236.00	830.0
237.00	1378.0
238.90	362.0
240.10	416.0
241.00	819.0
242.00	1731.0
243.00	1705.0
244.10	25312.0
245.00	3552.0
246.00	4806.0
247.00	1046.0
249.00	872.0
252.90	332.0
254.10	712.0
255.00	128688.0
256.00	17856.0
257.00	1507.0
258.00	7687.0
259.00	1126.0
265.00	3347.0
265.90	458.0
272.10	418.0
273.00	4630.0
274.00	10744.0
275.00	57136.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

276.00	6855.0
277.00	4115.0
277.90	615.0
281.00	448.0
283.00	584.0
284.00	328.0
285.10	651.0
289.00	316.0
293.00	1122.0
293.80	373.0
296.00	15962.0
297.10	1936.0
301.00	341.0
302.00	351.0
303.00	1977.0
304.00	566.0
308.10	384.0
314.00	747.0
315.00	1600.0
316.00	1063.0
320.90	401.0
321.90	373.0
323.10	4888.0
324.10	660.0
326.90	1171.0
327.90	708.0
331.90	400.0
333.10	657.0
334.00	3239.0
335.00	869.0
340.90	768.0
346.10	787.0
352.00	1713.0
353.10	867.0
354.00	1789.0
355.00	359.0
365.00	6932.0
366.00	1187.0
371.10	358.0
372.00	2829.0
373.10	787.0
383.00	835.0
390.10	316.0
402.00	1146.0
403.00	2143.0
404.00	795.0
421.00	1733.0
422.00	1694.0
423.10	10211.0
424.00	3504.0
424.90	545.0
441.10	31064.0
442.10	199488.0
443.10	37656.0
444.10	3755.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
39.90	1392.0
41.10	334.0
44.00	1196.0
72.90	569.0
81.10	410.0
83.80	355.0
96.00	333.0
207.10	1643.0

Instrument :

BNA_M

ClientSampleId :

DFTPP66