

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062424\
 Data File : BM046331.D
 Acq On : 26 Jun 2024 15:53
 Operator : MA/JU
 Sample : SP6537
 Misc : SFAM SURROGATES
 ALS Vial : 92 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP6537

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/27/2024
 Supervised By :mohammad ahmed 06/27/2024

Quant Time: Jun 26 17:16:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 26 03:24:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.388	152	1982	0.400	ng/ul	-0.06
4) Naphthalene-d8	10.149	136	6473	0.400	ng/ul	-0.07
9) Acenaphthene-d10	14.038	164	3651	0.400	ng/ul	#-0.03
13) Phenanthrene-d10	16.785	188	8143m	0.400	ng/ul	-0.05
17) Chrysene-d12	21.023	240	5916	0.400	ng/ul	0.00
23) Perylene-d12	23.086	264	5136m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.012	96	31234	11.486	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.793	152	6697	0.771	ng/ul	-0.05
18) Fluoranthene-d10	18.829	212	13717	0.774	ng/ul	-0.03

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062424\
Data File : BM046331.D
Acq On : 26 Jun 2024 15:53
Operator : MA/JU
Sample : SP6537
Misc : SFAM SURROGATES
ALS Vial : 92 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SP6537

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/27/2024
Supervised By :mohammad ahmed 06/27/2024

Quant Time: Jun 26 17:16:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM061824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 26 03:24:28 2024
Response via : Initial Calibration

