

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062515\
 Data File : BM001794.D
 Acq On : 25 Jun 2015 12:23
 Operator : TP/IZ
 Sample : G2762-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C01L0

Quant Time: Jun 26 13:22:59 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 26 04:32:51 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	41529	20.00	ng/ul	0.00
18) Naphthalene-d8	10.45	136	167614	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	103843	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	249528	20.00	ng/ul	0.00
77) Chrysene-d12	21.25	240	303456	20.00	ng/ul	0.00
85) Perylene-d12	23.47	264	304966	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.14	96	4492	5.16	ng/uL	0.00
5) Phenol-d5	6.83	99	91386	28.65	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.00	67	49247	27.56	ng/ul	0.00
9) 2-Chlorophenol-d4	7.19	132	75679	30.00	ng/ul	0.00
13) 4-Methylphenol-d8	8.36	113	76956	30.06	ng/ul	0.00
19) Nitrobenzene-d5	8.82	128	34157	29.44	ng/ul	0.00
22) 2-Nitrophenol-d4	9.54	143	40019	31.54	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.07	165	79888	29.42	ng/ul	0.00
29) 4-Chloroaniline-d4	10.59	131	100135	36.24	ng/ul	0.00
43) Dimethylphthalate-d6	13.73	166	224127	29.11	ng/ul	0.00
46) Acenaphthylene-d8	14.01	160	289216	28.65	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	38707	27.25	ng/ul	0.00
57) Fluorene-d10	15.31	176	208269	29.15	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	32088	21.05	ng/ul	0.00
70) Anthracene-d10	17.16	188	335703	28.88	ng/ul	0.00
78) Pyrene-d10	19.46	212	385501	29.99	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.33	264	396567	29.50	ng/ul	0.00

Target Compounds						Qvalue
6) Phenol	6.86	94	3680	1.12	ng/ul	99
16) 4-Methylphenol	8.43	108	5301	2.00	ng/ul	80
44) Dimethylphthalate	13.77	163	93495	11.11	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062515\
Data File : BM001794.D
Acq On : 25 Jun 2015 12:23
Operator : TP/IZ
Sample : G2762-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C01L0

Quant Time: Jun 26 13:22:59 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 26 04:32:51 2015
Response via : Initial Calibration

