

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062515\
 Data File : BM001802.D
 Acq On : 25 Jun 2015 17:12
 Operator : TP/IZ
 Sample : G2760-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0BK6

Quant Time: Jun 26 05:51:11 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 26 04:32:51 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	40957	20.00	ng/ul	0.00
18) Naphthalene-d8	10.45	136	156186	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	100123	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	238568	20.00	ng/ul	0.00
77) Chrysene-d12	21.25	240	295143	20.00	ng/ul	0.00
85) Perylene-d12	23.47	264	293791	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.14	96	1026	1.19	ng/uL	0.00
5) Phenol-d5	6.83	99	21629	6.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.00	67	50651	28.74	ng/ul	0.00
9) 2-Chlorophenol-d4	7.19	132	63373	25.48	ng/ul	0.00
13) 4-Methylphenol-d8	8.36	113	36453	14.44	ng/ul	0.00
19) Nitrobenzene-d5	8.82	128	36076	33.37	ng/ul	0.00
22) 2-Nitrophenol-d4	9.54	143	42953	36.33	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.07	165	79160	31.28	ng/ul	0.00
29) 4-Chloroaniline-d4	10.59	131	3069	1.19	ng/ul	0.00
43) Dimethylphthalate-d6	13.73	166	241365	32.51	ng/ul	0.00
46) Acenaphthylene-d8	14.01	160	324200	33.31	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	8926	6.52	ng/ul	0.00
57) Fluorene-d10	15.31	176	237385	34.46	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	41955	28.79	ng/ul	0.00
70) Anthracene-d10	17.16	188	360191	32.41	ng/ul	0.00
78) Pyrene-d10	19.46	212	431663	34.53	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.33	264	385301	29.75	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062515\
Data File : BM001802.D
Acq On : 25 Jun 2015 17:12
Operator : TP/IZ
Sample : G2760-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0BK6

Quant Time: Jun 26 05:51:11 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061515.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 26 04:32:51 2015
Response via : Initial Calibration

