


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062520\
Data File : BM026567.D
Acq On : 25 Jun 2020 20:51
Operator : JU/CG
Sample : L3062-22
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ET153

Manual Integrations
APPROVED

mohammad
6/26/2020 12:20:14 PM

Quant Time: Jun 26 00:48:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 26 00:37:06 2020
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062520\
Data File : BM026567.D
Acq On : 25 Jun 2020 20:51
Operator : JU/CG
Sample : L3062-22
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ET153

Manual Integrations
APPROVED

mohammad
6/26/2020 12:20:14 PM

Quant Time: Jun 26 00:45:52 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 26 00:37:06 2020
Response via : Initial Calibration



TIC: BM026567.D

(52) 4-Nitrophenol-d4 (S)

14.310min (+0.035) 1.62ng/ul




response 7157

Ion	Exp%	Act%
143.00	100	100
113.00	81.70	90.04
41.00	49.50	62.90#
42.00	34.20	43.13#

```
Data Path   : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062520\  
Data File  : BM026567.D  
Acq On     : 25 Jun 2020   20:51  
Operator   : JU/CG  
Sample     : L3062-22  
Misc       :  
ALS Vial   : 18      Sample Multiplier: 1
```

mohammad
6/26/2020 12:20:14 PM

Quant Time: Jun 26 00:45:52 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 26 00:37:06 2020
Response via : Initial Calibration



TIC: BM026567.D

(52) 4-Nitrophenol-d4 (S)

14.310min (+0.035) 3.33ng/ul m 06/29/20 SD

response 14726

Ion	Exp%	Act%
143.00	100	100
113.00	81.70	90.04
41.00	49.50	62.90#
42.00	34.20	43.13#

Quantitation Report (QT/LSC Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062520\
 Data File : BM026567.D
 Acq On : 25 Jun 2020 20:51
 Operator : JU/CG
 Sample : L3062-22
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ET153

Manual Integrations
 APPROVED

mohammad
 6/26/2020 12:20:14 PM

Quant Time: Jun 26 00:48:07 2020

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM061220MA.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Jun 26 00:37:06 2020

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	97147	20.00	ng/ul	0.00
18) Naphthalene-d8	10.13	136	460439	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.03	164	316698	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.79	188	724321	20.00	ng/ul	0.00
78) Chrysene-d12	21.00	240	787403	20.00	ng/ul	0.00
86) Perylene-d12	23.09	264	800598	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.96	96	9421	3.40	ng/uL	0.00
5) Phenol-d5	6.57	99	144064	16.10	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.73	67	114472	20.22	ng/ul	0.00
9) 2-Chlorophenol-d4	6.91	132	119928	16.44	ng/ul	0.00
13) 4-Methylphenol-d8	8.09	113	126295	17.93	ng/ul	0.00
19) Nitrobenzene-d5	8.52	128	59754	14.61	ng/ul	0.00
22) 2-Nitrophenol-d4	9.23	143	56383	12.53	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.77	165	120836	14.58	ng/ul	0.00
29) 4-Chloroaniline-d4	10.28	131	164217	19.00	ng/ul	0.00
44) Dimethylphthalate-d6	13.46	166	457061	17.16	ng/ul	0.00
47) Acenaphthylene-d8	13.71	160	451561	14.14	ng/ul	0.00
52) 4-Nitrophenol-d4	14.31	143	14726m	3.33	ng/ul	0.04
58) Fluorene-d10	15.03	176	315162	13.79	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.19	200	12542	2.65	ng/ul	0.00
71) Anthracene-d10	16.88	188	487921	13.23	ng/ul	0.00
79) Pyrene-d10	19.19	212	524664	12.15	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.95	264	443913	10.03	ng/ul	0.00

06/24/20 JU

Target Compounds					Ovalue
45) Dimethylphthalate	13.50	163	161814	5.852	ng/ul 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed