


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062522\
 Data File : BM035680.D
 Acq On : 25 Jun 2022 09:58
 Operator : CG/JU
 Sample : DFTPP061
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP061

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM062522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 29 07:47:25 2022



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	146808	PASS
68	69	0.00	2	1.9	2783	PASS
69	198	0.00	100	36.4	146317	PASS
70	69	0.00	2	0.6	871	PASS
127	198	10	80	48.2	194027	PASS
197	198	0.00	2	0.5	2054	PASS
198	198	100	100	100.0	402475	PASS
199	198	5	9	6.8	27360	PASS
275	198	10	60	27.5	110829	PASS
365	198	1	100	3.3	13424	PASS
441	198	0.01	100	13.7	55019	PASS
442	198	50	100	89.9	361813	PASS
443	442	15	24	19.6	71029	PASS

m/z	Abundance
38.90	384.0
39.90	1469.0
41.00	955.0
42.00	207.0
43.00	683.0
44.00	1765.0
45.00	295.0
46.90	165.0
48.90	725.0
49.90	159.0
51.10	267.0
55.00	526.0
56.00	189.0
57.00	202.0
59.00	258.0
64.90	174.0
67.00	225.0
69.00	280.0
70.90	188.0
73.00	1527.0
73.90	196.0
74.90	313.0
77.00	546.0
80.90	334.0
82.00	184.0
83.10	397.0
83.90	513.0
85.80	344.0
88.90	151.0
91.00	521.0
94.90	277.0
96.10	291.0
96.90	263.0
103.80	175.0
104.90	230.0
107.00	203.0
109.00	184.0
110.90	181.0
115.00	298.0
119.00	329.0
120.90	191.0
127.90	191.0
128.80	184.0
131.00	157.0
133.00	519.0
134.90	395.0
135.90	151.0
140.90	161.0
147.10	587.0
148.00	253.0
164.90	267.0
178.90	154.0
190.90	520.0
191.90	246.0
193.00	476.0
195.00	174.0
197.10	203.0
207.10	3236.0
208.00	540.0
209.00	514.0
221.10	295.0
265.00	195.0
266.90	158.0
281.00	889.0
281.80	196.0
340.90	194.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

m/z	Abundance
37.00	797.0
38.00	1818.0
39.00	11730.0
39.90	1738.0
41.00	956.0
43.10	586.0
44.00	1743.0
45.00	458.0
47.00	271.0
48.10	153.0
49.00	1529.0
50.00	39928.0
51.00	161536.0
52.00	8438.0
53.00	434.0
54.90	1240.0
56.00	5348.0
57.00	11570.0
58.00	473.0
59.00	230.0
59.90	204.0
61.00	2364.0
62.00	2326.0
63.00	5886.0
64.00	1003.0
65.00	3045.0
66.10	267.0
67.00	414.0
68.00	3254.0
69.00	158208.0
70.00	868.0
71.10	229.0
72.00	184.0
73.00	2864.0
74.00	16246.0
75.00	25704.0
76.00	8669.0
77.00	180480.0
78.00	12795.0
79.00	12093.0
80.00	8536.0
81.00	13560.0
82.00	3263.0
83.00	3383.0
84.00	709.0
85.00	2246.0
85.90	3731.0
87.00	1549.0
87.90	659.0
89.00	267.0
91.00	3109.0
92.00	3358.0
93.00	19416.0
94.00	1790.0
95.00	574.0
96.00	1201.0
96.90	468.0
98.00	14746.0
99.00	13708.0
100.00	1114.0
101.00	7391.0
101.90	379.0
103.00	2732.0
104.00	4796.0
105.00	4588.0
106.00	1481.0
107.00	52200.0
108.00	8870.0
109.00	1494.0
110.00	101888.0
111.00	16055.0
111.90	1828.0
113.00	724.0
115.00	417.0
116.00	3074.0
117.00	41016.0
117.90	3745.0
119.00	589.0
119.90	639.0
120.90	414.0
121.90	3438.0
123.00	5404.0
124.00	2679.0
125.00	2316.0
126.10	947.0
127.00	200256.0
128.00	15362.0
129.00	76384.0
130.00	6407.0
130.90	1443.0
132.10	866.0
132.90	794.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

134.00	2224.0
135.00	5808.0
136.00	2274.0
137.00	3680.0
138.90	463.0
139.90	1146.0
141.00	10305.0
142.00	3432.0
143.00	2427.0
144.00	534.0
144.90	671.0
146.00	2198.0
147.00	5791.0
148.00	11055.0
149.00	2542.0
149.90	920.0
151.10	1311.0
151.90	858.0
153.00	3088.0
154.00	2050.0
155.00	5017.0
156.00	8086.0
157.00	1941.0
158.00	2016.0
159.00	1583.0
160.00	2755.0
161.00	4347.0
162.00	1235.0
163.00	585.0
163.90	605.0
165.00	3738.0
166.00	2724.0
167.00	17448.0
168.00	11272.0
169.00	1871.0
170.00	901.0
170.90	1184.0
172.00	1800.0
173.00	2388.0
174.00	3986.0
175.00	6944.0
175.90	2242.0
177.00	3986.0
177.90	1493.0
178.90	13515.0
180.00	10427.0
181.00	4389.0
182.00	825.0
182.90	453.0
184.00	1139.0
185.00	6726.0
186.00	54296.0
187.00	15282.0
188.00	1879.0
189.00	3700.0
189.90	657.0
191.00	1862.0
192.00	4648.0
193.00	5333.0
194.00	1162.0
195.00	638.0
196.00	11528.0
197.00	2324.0
198.00	393408.0
199.00	26984.0
199.90	2120.0
201.50	2269.0
202.20	572.0
203.00	2475.0
204.00	13775.0
205.00	23232.0
206.00	97984.0
207.00	16226.0
208.00	3610.0
209.00	1341.0
210.20	1668.0
211.00	3779.0
211.70	955.0
213.10	280.0
215.00	1142.0
216.00	2279.0
217.00	24448.0
218.00	3684.0
219.00	222.0
221.00	18200.0
221.80	3605.0
223.00	5578.0
224.00	53352.0
225.00	13905.0
226.00	1771.0
227.00	20744.0
228.00	2913.0
229.00	4499.0
229.90	670.0
231.00	1470.0
232.00	353.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

232.90	302.0
234.00	1233.0
235.00	1675.0
236.00	1301.0
237.00	2290.0
237.90	227.0
239.00	1052.0
240.00	771.0
240.90	1204.0
242.00	2879.0
243.00	3227.0
244.00	40936.0
245.00	5616.0
246.00	8671.0
247.00	1416.0
248.00	384.0
249.00	1634.0
249.90	373.0
250.90	361.0
251.90	373.0
253.00	927.0
254.10	1727.0
255.00	217472.0
256.00	32864.0
257.00	3367.0
258.00	11691.0
259.00	1978.0
259.80	231.0
261.00	194.0
264.00	261.0
265.00	4686.0
266.00	908.0
267.00	285.0
268.00	197.0
270.00	295.0
270.80	348.0
272.00	525.0
273.00	6300.0
274.00	16792.0
275.00	101872.0
276.00	12997.0
277.00	7589.0
278.00	1676.0
278.90	241.0
281.00	1119.0
282.00	420.0
283.10	874.0
283.90	614.0
285.00	1180.0
285.90	331.0
288.00	257.0
289.00	391.0
290.00	343.0
291.10	256.0
292.00	337.0
292.90	1605.0
294.00	609.0
295.00	807.0
296.00	24248.0
297.00	3181.0
297.80	276.0
300.90	360.0
302.00	463.0
303.00	2519.0
304.00	826.0
304.90	207.0
307.80	377.0
308.80	272.0
309.90	394.0
313.00	306.0
314.00	1229.0
315.00	2775.0
316.00	1550.0
316.90	273.0
321.00	725.0
322.00	592.0
323.00	7895.0
324.00	1433.0
325.90	221.0
327.00	1600.0
327.90	1031.0
332.00	617.0
332.90	802.0
334.00	5012.0
335.00	1297.0
336.00	203.0
339.00	167.0
341.10	1158.0
342.00	256.0
346.00	1654.0
346.90	345.0
352.10	2392.0
353.00	1646.0
354.10	2766.0
355.00	696.0
359.00	215.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

365.00	10658.0
365.90	1737.0
370.10	164.0
371.00	752.0
372.00	4325.0
373.10	1122.0
374.00	157.0
382.90	1187.0
383.90	327.0
390.00	699.0
390.90	217.0
392.00	225.0
401.00	344.0
402.00	1715.0
403.00	2624.0
404.00	1044.0
404.90	182.0
421.00	2047.0
422.00	2188.0
423.10	15638.0
424.00	5028.0
425.00	635.0
440.00	204.0
441.10	41792.0
442.10	270464.0
443.10	55000.0
444.00	5337.0
445.00	410.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

m/z	Abundance
37.00	757.0
38.00	2376.0
39.00	13096.0
40.00	1738.0
41.00	923.0
42.00	162.0
43.10	520.0
43.90	1828.0
44.90	487.0
47.00	153.0
49.00	1963.0
50.00	42792.0
51.00	171456.0
52.00	8847.0
52.90	588.0
55.00	1260.0
56.00	5262.0
57.00	11868.0
58.00	708.0
59.00	296.0
60.00	237.0
61.00	2314.0
62.00	2519.0
63.00	6421.0
64.00	1191.0
65.00	3632.0
66.00	306.0
67.00	426.0
68.00	3122.0
69.00	174080.0
70.00	1011.0
71.00	299.0
72.00	177.0
73.00	3002.0
74.00	17856.0
75.00	27640.0
76.00	8836.0
77.00	196992.0
78.00	13493.0
79.00	12615.0
80.00	10557.0
81.00	13700.0
82.00	3731.0
83.00	3336.0
83.90	974.0
85.00	2442.0
86.00	4157.0
87.00	2037.0
88.00	790.0
88.90	322.0
91.00	3840.0
92.00	3729.0
93.00	22312.0
94.00	1657.0
94.90	519.0
96.00	1281.0
97.00	588.0
98.00	16744.0
99.00	13827.0
99.90	1250.0
101.00	7809.0
102.20	465.0
103.00	2927.0
104.00	5021.0
105.00	5255.0
106.00	1738.0
107.00	59960.0
108.00	8720.0
108.90	1874.0
110.00	115432.0
111.00	17112.0
112.00	2273.0
112.90	566.0
113.80	173.0
114.90	598.0
116.00	3764.0
117.00	48736.0
118.00	3545.0
119.00	561.0
119.90	990.0
121.10	353.0
122.00	3710.0
123.00	6454.0
124.00	2708.0
125.00	2629.0
126.10	1170.0
127.00	231360.0
128.00	16050.0
129.00	86752.0
130.00	7644.0
131.00	1649.0
132.00	843.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

133.00	709.0
133.90	2698.0
135.00	7571.0
136.00	2960.0
137.00	4155.0
138.00	661.0
138.80	529.0
140.00	1136.0
141.00	10989.0
142.00	3902.0
142.90	2430.0
144.00	870.0
145.00	913.0
146.00	1880.0
147.00	5978.0
148.00	12716.0
149.00	2892.0
150.00	966.0
151.30	1344.0
152.00	1028.0
153.00	3540.0
154.00	2545.0
155.00	6192.0
156.00	9557.0
157.00	1866.0
157.90	2126.0
159.00	1383.0
160.00	3456.0
161.00	5319.0
162.00	1763.0
163.10	446.0
164.00	869.0
165.00	4200.0
166.00	3370.0
167.00	21912.0
168.00	12160.0
169.00	2513.0
170.00	792.0
171.00	858.0
171.90	2045.0
173.00	2885.0
174.00	4907.0
175.00	8430.0
176.00	2949.0
177.00	4699.0
178.00	1832.0
179.00	17064.0
180.00	12271.0
181.00	5890.0
181.90	641.0
182.90	600.0
184.00	1315.0
185.00	8107.0
186.00	64080.0
187.00	18496.0
188.00	2180.0
188.90	3897.0
190.00	464.0
191.00	2161.0
192.00	5205.0
193.00	6560.0
194.00	1366.0
195.00	1111.0
196.00	15690.0
197.00	2559.0
198.00	484032.0
199.00	32456.0
200.00	2766.0
201.60	2273.0
203.00	3332.0
204.00	15220.0
205.00	27584.0
206.00	120704.0
207.00	17936.0
208.00	4462.0
209.00	1593.0
210.00	1872.0
211.00	4574.0
211.80	1146.0
213.00	296.0
214.00	161.0
215.00	1095.0
216.00	2391.0
217.00	30352.0
218.00	4290.0
219.10	489.0
221.00	23376.0
221.70	4893.0
223.00	7309.0
224.00	67744.0
225.00	17304.0
226.00	1716.0
227.00	26360.0
228.00	3778.0
229.00	6165.0
230.10	838.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

231.00	2707.0
232.00	409.0
232.90	472.0
234.00	1685.0
234.90	2153.0
236.00	1387.0
237.00	2686.0
238.10	260.0
239.00	1016.0
239.90	889.0
241.00	1672.0
242.00	3452.0
243.10	3571.0
244.00	54976.0
245.10	6728.0
246.00	10243.0
247.00	1915.0
247.90	484.0
249.00	2349.0
250.00	371.0
250.80	396.0
252.10	575.0
253.10	1148.0
254.10	1985.0
255.00	277760.0
256.00	40440.0
257.00	3689.0
258.00	15499.0
259.00	2678.0
260.00	260.0
261.00	534.0
263.10	182.0
264.10	469.0
265.00	5882.0
266.00	1088.0
267.00	441.0
270.00	403.0
271.00	680.0
271.90	899.0
273.00	7900.0
274.00	21616.0
275.00	133120.0
276.00	16680.0
277.00	9885.0
278.00	2184.0
279.00	342.0
281.00	943.0
282.10	454.0
283.00	1337.0
284.00	794.0
285.10	1987.0
286.00	244.0
289.00	387.0
290.00	369.0
291.00	252.0
292.00	313.0
293.00	2154.0
294.00	673.0
295.00	774.0
296.00	33304.0
297.00	4512.0
297.90	334.0
300.90	372.0
302.00	555.0
303.00	3748.0
304.00	1243.0
308.00	369.0
308.90	434.0
310.00	566.0
310.90	175.0
312.00	159.0
312.90	407.0
314.10	1654.0
315.00	3731.0
316.00	2155.0
317.00	312.0
321.00	1029.0
322.00	693.0
323.10	11352.0
324.00	2223.0
324.90	220.0
326.00	328.0
327.00	2131.0
328.10	1144.0
328.90	301.0
332.00	797.0
333.00	1133.0
334.00	7331.0
335.00	1772.0
335.90	201.0
339.00	195.0
341.10	1292.0
342.00	351.0
343.00	196.0
346.00	2659.0
346.90	536.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

351.00	215.0
352.00	3407.0
353.00	2630.0
354.10	4300.0
355.10	927.0
358.00	198.0
359.00	267.0
365.00	16180.0
365.90	2390.0
370.00	398.0
370.90	937.0
372.00	6530.0
373.10	1305.0
373.90	150.0
377.10	164.0
383.00	1467.0
383.90	560.0
390.00	775.0
391.00	731.0
392.10	443.0
400.90	288.0
402.00	2826.0
403.10	3830.0
404.00	1707.0
405.10	282.0
415.00	199.0
421.00	3284.0
422.00	2960.0
423.00	23008.0
424.10	6525.0
425.10	765.0
440.10	272.0
441.10	62808.0
442.10	425408.0
443.10	83592.0
444.10	7286.0
445.10	554.0

Instrument :

BNA_M

ClientSampleId :

DFTPP061

m/z	Abundance
36.00	201.0
36.90	456.0
38.00	1389.0
39.00	7651.0
40.00	1853.0
41.10	789.0
42.00	278.0
43.00	785.0
44.00	1736.0
44.90	544.0
46.80	182.0
49.00	1258.0
50.00	27816.0
51.00	108232.0
52.00	4834.0
52.90	417.0
55.00	1125.0
56.00	3235.0
57.00	7196.0
57.90	364.0
58.90	270.0
60.10	170.0
61.00	1450.0
62.00	1881.0
63.00	4516.0
64.00	693.0
65.00	2411.0
66.00	260.0
66.90	377.0
68.00	1972.0
69.00	107504.0
70.00	734.0
70.80	236.0
73.00	2652.0
74.00	11005.0
75.00	17456.0
76.00	5424.0
77.00	120360.0
78.00	8173.0
79.00	8189.0
80.00	5681.0
81.00	8547.0
82.00	2116.0
83.00	1922.0
84.00	828.0
85.00	2024.0
86.00	2815.0
86.90	1167.0
87.90	561.0
88.90	309.0
91.00	2367.0
91.90	2087.0
93.00	14186.0
94.00	1131.0
95.10	487.0
96.00	1025.0
97.00	389.0
98.00	11181.0
99.00	8932.0
100.00	736.0
101.00	5112.0
102.00	313.0
103.10	1860.0
104.00	3483.0
105.00	3240.0
106.00	1032.0
107.00	37824.0
108.00	6264.0
109.00	1552.0
110.00	73952.0
111.00	11006.0
111.90	1439.0
113.00	525.0
114.10	244.0
114.90	361.0
116.00	2721.0
117.00	32448.0
118.00	2613.0
119.00	411.0
120.00	603.0
121.00	354.0
121.90	2679.0
123.00	4175.0
124.00	2124.0
125.00	1947.0
126.10	713.0
127.00	150464.0
128.00	11918.0
129.00	55392.0
130.00	4605.0
130.90	1147.0
131.90	484.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

133.00	610.0
134.00	1902.0
135.00	5324.0
136.00	1934.0
137.00	2643.0
137.80	641.0
138.90	374.0
139.90	738.0
141.00	6931.0
142.00	2205.0
143.00	1828.0
144.00	723.0
145.00	565.0
146.00	1623.0
147.00	4360.0
148.00	8789.0
149.00	2148.0
150.00	568.0
151.00	997.0
151.70	539.0
153.00	2576.0
154.00	2074.0
155.00	4531.0
156.00	6720.0
157.10	1453.0
157.90	1618.0
159.00	1279.0
160.00	2127.0
161.00	3884.0
161.90	1286.0
162.90	395.0
164.00	624.0
165.00	3171.0
166.00	2410.0
167.00	15537.0
168.00	7974.0
169.00	1271.0
170.10	528.0
171.00	691.0
172.00	1036.0
173.00	1633.0
174.00	3234.0
175.00	5423.0
176.00	1472.0
177.00	2811.0
178.00	1211.0
179.00	11735.0
180.00	7576.0
181.00	3742.0
182.00	798.0
182.90	408.0
184.00	878.0
185.00	5307.0
186.00	43048.0
187.00	11977.0
188.00	1413.0
189.00	2566.0
190.00	508.0
191.10	1600.0
192.00	3580.0
193.00	4608.0
194.00	1031.0
195.00	607.0
196.00	10841.0
196.90	1889.0
198.00	329984.0
199.00	22640.0
200.00	1822.0
201.30	1770.0
202.90	1959.0
204.00	12165.0
205.00	18424.0
206.00	87592.0
207.10	13937.0
208.00	3408.0
209.00	1260.0
210.00	1359.0
211.00	2950.0
211.80	865.0
213.00	251.0
214.90	1005.0
216.00	1753.0
217.00	21560.0
218.00	2646.0
219.00	339.0
220.20	799.0
221.10	15470.0
223.00	5551.0
224.00	46624.0
225.00	11946.0
226.10	1254.0
227.00	19352.0
228.00	2969.0
229.00	4091.0
229.90	451.0
231.00	1461.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

232.10	346.0
233.10	280.0
234.00	1280.0
235.00	1681.0
236.00	996.0
237.00	1768.0
237.90	297.0
239.00	738.0
239.90	603.0
241.00	1179.0
242.00	2612.0
243.00	2570.0
244.10	41472.0
245.00	4930.0
246.00	7010.0
247.00	1519.0
248.00	201.0
249.00	1284.0
250.00	310.0
251.00	589.0
251.90	416.0
253.00	705.0
254.10	1336.0
255.00	204160.0
256.00	28800.0
257.00	2351.0
258.00	10902.0
259.00	1767.0
259.90	266.0
260.90	328.0
264.00	242.0
265.00	4589.0
265.90	587.0
266.90	465.0
270.10	194.0
271.00	467.0
272.00	523.0
273.00	6248.0
274.00	16488.0
275.00	97496.0
276.00	13105.0
277.00	8544.0
278.00	1338.0
278.90	212.0
281.10	1090.0
281.90	513.0
283.00	928.0
283.90	761.0
285.00	1618.0
285.90	226.0
288.90	342.0
289.90	219.0
291.00	187.0
291.90	416.0
293.00	1779.0
293.90	441.0
295.10	492.0
296.00	26064.0
297.00	3335.0
298.10	276.0
300.90	450.0
301.80	400.0
303.00	3613.0
304.00	1072.0
305.00	218.0
307.90	306.0
308.90	352.0
310.00	420.0
313.00	254.0
314.00	1347.0
315.00	2888.0
316.00	1560.0
316.90	258.0
319.90	161.0
321.00	919.0
322.10	530.0
323.10	9023.0
324.10	1634.0
325.00	204.0
326.00	206.0
326.90	1678.0
327.90	865.0
331.90	701.0
332.90	821.0
334.00	5417.0
335.00	1636.0
335.90	263.0
341.00	1258.0
342.00	495.0
346.00	2217.0
346.90	416.0
351.10	192.0
352.00	2590.0
353.00	1795.0
354.10	3201.0
355.00	844.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061

358.90	329.0
365.00	13435.0
366.00	1804.0
370.00	481.0
371.00	651.0
372.10	4831.0
373.10	1347.0
383.00	1567.0
384.00	325.0
385.10	150.0
390.10	665.0
391.10	412.0
391.90	314.0
400.90	417.0
402.10	2361.0
403.10	3551.0
404.00	1166.0
404.80	157.0
409.90	160.0
414.90	208.0
421.00	3102.0
422.00	2889.0
423.00	21032.0
424.10	5297.0
425.00	749.0
439.20	208.0
440.00	238.0
441.10	60456.0
442.10	389568.0
443.10	74496.0
444.00	6828.0
445.00	305.0

Instrument :
BNA_M
ClientSampleId :
DFTPP061