

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062616\
 Data File : BM006063.D
 Acq On : 27 Jun 2016 05:22
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02037

Quant Time: Jun 27 07:05:18 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 27 05:15:31 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	215#	0.00
2 S	1,4-Dioxane-d8	0.448	0.421	6.0	197#	0.00
3 S	Phenol-d5	2.104	1.894	10.0	196#	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.170	1.101	5.9	198#	0.00
5 S	2-Chlorophenol-d4	1.465	1.418	3.2	207#	0.00
6 S	4-Methylphenol-d8	1.776	1.519	14.5	185#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	192#	0.00
8 S	Nitrobenzene-d5	0.152	0.155	-2.0	193#	0.00
9 S	2-Nitrophenol-d4	0.169	0.171	-1.2	189#	0.00
10 S	2,4-Dichlorophenol-d3	0.325	0.327	-0.6	190#	0.00
11	Naphthalene	0.999	1.015	-1.6	192#	0.00
12 S	4-Chloroaniline-d4	0.353	0.397	-12.5	172#	0.00
13	2-Methylnaphthalene	0.804	0.770	4.2	178#	0.00
14	1-Methylnaphthalene	0.770	0.728	5.5	176#	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	158#	0.00
16 S	Dimethylphthalate-d6	1.712	1.566	8.5	144	0.00
17 S	Acenaphthylene-d8	1.986	2.046	-3.0	159#	0.00
18	Acenaphthylene	2.094	2.156	-3.0	158#	0.00
19 C	Acenaphthene	1.330	1.362	-2.4	160#	0.00
20 S	4-Nitrophenol-d4	0.365	0.324	11.2	131	0.00
21 S	Fluorene-d10	1.462	1.410	3.6	150	0.00
22	Fluorene	1.602	1.527	4.7	146	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.111	0.071	36.0#	86	0.00
25	Phenanthrene	1.076	1.110	-3.2	141	0.00
26 S	Anthracene-d10	0.919	0.947	-3.0	139	0.00
27	Anthracene	1.111	1.143	-2.9	139	0.00
28 C	Fluoranthene	1.378	1.326	3.8	128	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
30 S	Pyrene-d10	0.876	1.138	-29.9#	126	0.00
31	Pyrene	1.145	1.479	-29.2#	124	0.00
32	Benzo(a)anthracene	1.188	1.243	-4.6	101	0.00
33	Chrysene	1.085	1.128	-4.0	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	76	0.00
35	Benzo(b)fluoranthene	1.162	1.249	-7.5	81	0.00
36	Benzo(k)fluoranthene	1.081	1.166	-7.9	79	0.00
37 S	Benzo(a)pyrene-d12	0.903	0.931	-3.1	77	0.00
38 C	Benzo(a)pyrene	1.114	1.151	-3.3	77	0.00
39	Indeno(1,2,3-cd)pyrene	1.328	1.283	3.4	71	-0.02
40	Dibenzo(a,h)anthracene	1.104	1.068	3.3	71	-0.02
41	Benzo(g,h,i)perylene	1.149	1.111	3.3	71	-0.01

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062616\
Data File : BM006063.D
Acq On : 27 Jun 2016 05:22
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02037

Quant Time: Jun 27 07:05:18 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062516.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 27 05:15:31 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0