

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_M\Data\BM062617\
 Data File : BM010747.D
 Acq On : 26 Jun 2017 15:31
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02001

Quant Time: Jun 27 01:26:23 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 16:50:55 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.390	0.444	-13.8	81	0.00
3 S	1,4-Dioxane-d8	0.349	0.393	-12.6	83	0.00
4	Benzaldehyde	1.121	1.261	-12.5	72	0.00
5 S	Phenol-d5	1.905	1.693	11.1	70	0.00
6	Phenol	1.962	1.751	10.8	71	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.088	0.992	8.8	68	0.00
8	Bis(2-Chloroethyl)ether	1.429	1.266	11.4	68	0.00
9 S	2-Chlorophenol-d4	1.337	1.302	2.6	76	0.00
10	2-Chlorophenol	1.336	1.272	4.8	72	0.00
11	2-Methylphenol	1.496	1.428	4.5	74	0.00
12	2,2'-oxybis(1-Chloropropane	0.997	0.811	18.7	61	0.00
13 S	4-Methylphenol-d8	1.580	1.439	8.9	70	0.00
14	Acetophenone	2.579	2.469	4.3	74	0.00
15 P	N-Nitroso-di-n-propylamine	1.390	1.294	6.9	70	0.00
16	4-Methylphenol	1.646	1.512	8.1	71	0.00
17	Hexachloroethane	0.594	0.602	-1.3	78	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
19 S	Nitrobenzene-d5	0.143	0.142	0.7	79	0.00
20	Nitrobenzene	0.416	0.418	-0.5	77	0.00
21	Isophorone	0.828	0.738	10.9	68	0.00
22 S	2-Nitrophenol-d4	0.164	0.171	-4.3	81	0.00
23 C	2-Nitrophenol	0.174	0.166	4.6	72	0.00
24	2,4-Dimethylphenol	0.443	0.439	0.9	78	0.00
25	Bis(2-Chloroethoxy)methane	0.459	0.410	10.7	68	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.350	-1.2	79	0.00
27 C	2,4-Dichlorophenol	0.337	0.346	-2.7	80	0.00
28	Naphthalene	0.995	0.981	1.4	76	0.00
29 S	4-Chloroaniline-d4	0.365	0.415	-13.7	77	0.00
30	4-Chloroaniline	0.367	0.404	-10.1	75	0.00
31 C	Hexachlorobutadiene	0.254	0.278	-9.4	84	0.00
32	Caprolactam	0.118	0.106	10.2	66	0.00
33 C	4-Chloro-3-methylphenol	0.430	0.421	2.1	74	0.00
34	2-Methylnaphthalene	0.800	0.802	-0.3	78	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
36	1,2,4,5-Tetrachlorobenzene	0.706	0.719	-1.8	83	0.00
37	Hexachlorocyclopentadiene	0.406	0.335	17.5	73	0.00
38 C	2,4,6-Trichlorophenol	0.482	0.476	1.2	81	0.00
39	2,4,5-Trichlorophenol	0.498	0.481	3.4	79	0.00
40	1,1'-Biphenyl	1.564	1.513	3.3	80	0.00
41	2-Chloronaphthalene	1.230	1.201	2.4	81	0.00
42	2-Nitroaniline	0.360	0.373	-3.6	86	0.00
43 S	Dimethylphthalate-d6	1.756	1.667	5.1	77	0.00
44	Dimethylphthalate	1.721	1.633	5.1	78	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_M\Data\BM062617\
 Data File : BM010747.D
 Acq On : 26 Jun 2017 15:31
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02001

Quant Time: Jun 27 01:26:23 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 16:50:55 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.298	0.307	-3.0	86	0.00
46 S	Acenaphthylene-d8	2.042	1.971	3.5	80	0.00
47	Acenaphthylene	1.945	1.848	5.0	79	0.00
48	3-Nitroaniline	0.301	0.316	-5.0	86	0.00
49 C	Acenaphthene	1.312	1.260	4.0	80	0.00
50	2,4-Dinitrophenol	0.216	0.220	-1.9	96	0.00
51 S	4-Nitrophenol-d4	0.309	0.290	6.1	80	0.00
52	4-Nitrophenol	0.392	0.461	-17.6	97	0.00
53	Dibenzofuran	1.987	1.956	1.6	81	0.00
54	2,4-Dinitrotoluene	0.462	0.488	-5.6	87	0.00
55	2,3,4,6-Tetrachlorophenol	0.504	0.509	-1.0	82	0.00
56	Diethylphthalate	1.811	1.736	4.1	79	0.00
57 S	Fluorene-d10	1.518	1.539	-1.4	84	0.00
58	Fluorene	1.664	1.639	1.5	82	0.00
59	4-Chlorophenyl-phenylether	0.904	0.917	-1.4	84	0.00
60	4-Nitroaniline	0.355	0.352	0.8	85	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.123	0.120	2.4	91	0.00
63	4,6-Dinitro-2-methylphenol	0.131	0.127	3.1	91	0.00
64	N-Nitrosodiphenylamine	0.555	0.529	4.7	81	0.00
65	4-Bromophenyl-phenylether	0.229	0.228	0.4	86	0.00
66	Hexachlorobenzene	0.241	0.237	1.7	85	0.00
67	Atrazine	0.234	0.235	-0.4	85	0.00
68 C	Pentachlorophenol	0.170	0.156	8.2	80	0.00
69	Phenanthrene	1.063	1.057	0.6	86	0.00
70 S	Anthracene-d10	0.966	0.961	0.5	84	0.00
71	Anthracene	1.095	1.083	1.1	84	0.00
72	Carbazole	0.956	0.936	2.1	84	0.00
73	Di-n-butylphthalate	1.215	1.128	7.2	78	0.00
74 C	Fluoranthene	1.370	1.376	-0.4	86	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
76 S	Pyrene-d10	0.929	0.961	-3.4	86	0.00
77	Pyrene	1.155	1.178	-2.0	85	0.00
78	Butylbenzylphthalate	0.434	0.443	-2.1	85	0.00
79	3,3'-Dichlorobenzidine	0.400	0.391	2.3	81	0.00
80	Benzo(a)anthracene	1.165	1.176	-0.9	85	0.00
81	Bis(2-ethylhexyl)phthalate	0.663	0.666	-0.5	84	0.00
82	Chrysene	1.038	1.062	-2.3	85	0.00
83 I	Perylene-d12	1.000	1.000	0.0	74	0.00
84	Di-n-octyl phthalate	1.398	1.458	-4.3	84	0.00
85	Benzo(b)fluoranthene	1.298	1.360	-4.8	78	0.00
86	Benzo(k)fluoranthene	1.231	1.256	-2.0	75	0.00
87 S	Benzo(a)pyrene-d12	0.967	0.965	0.2	73	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_M\Data\BM062617\
Data File : BM010747.D
Acq On : 26 Jun 2017 15:31
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02001

Quant Time: Jun 27 01:26:23 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM062017.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 26 16:50:55 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.195	1.201	-0.5	74	0.00
89	Indeno(1,2,3-cd)pyrene	1.240	1.205	2.8	71	0.00
90	Dibenzo(a,h)anthracene	1.033	1.005	2.7	70	0.00
91	Benzo(g,h,i)perylene	0.987	0.995	-0.8	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0