

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062716\
 Data File : BM006065.D
 Acq On : 27 Jun 2016 09:21
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02038

Quant Time: Jun 28 02:09:03 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 27 10:49:48 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	228#	0.00
2 S	1,4-Dioxane-d8	0.448	0.415	7.4	206#	0.00
3 S	Phenol-d5	2.104	1.867	11.3	205#	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.170	1.087	7.1	208#	0.00
5 S	2-Chlorophenol-d4	1.465	1.413	3.5	219#	0.00
6 S	4-Methylphenol-d8	1.776	1.497	15.7	194#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	201#	0.00
8 S	Nitrobenzene-d5	0.152	0.157	-3.3	205#	0.00
9 S	2-Nitrophenol-d4	0.169	0.170	-0.6	198#	0.00
10 S	2,4-Dichlorophenol-d3	0.325	0.325	0.0	198#	0.00
11	Naphthalene	0.999	1.016	-1.7	201#	0.00
12 S	4-Chloroaniline-d4	0.353	0.399	-13.0	182#	0.00
13	2-Methylnaphthalene	0.804	0.768	4.5	187#	0.00
14	1-Methylnaphthalene	0.770	0.724	6.0	184#	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	165#	0.00
16 S	Dimethylphthalate-d6	1.712	1.589	7.2	152#	0.00
17 S	Acenaphthylene-d8	1.986	2.081	-4.8	169#	0.00
18	Acenaphthylene	2.094	2.173	-3.8	167#	0.00
19 C	Acenaphthene	1.330	1.364	-2.6	168#	0.00
20 S	4-Nitrophenol-d4	0.365	0.315	13.7	133	0.00
21 S	Fluorene-d10	1.462	1.435	1.8	159#	0.00
22	Fluorene	1.602	1.552	3.1	155#	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	147	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.111	0.080	27.9#	102	0.00
25	Phenanthrene	1.076	1.110	-3.2	149	0.00
26 S	Anthracene-d10	0.919	0.948	-3.2	147	0.00
27	Anthracene	1.111	1.144	-3.0	147	0.00
28 C	Fluoranthene	1.378	1.340	2.8	137	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
30 S	Pyrene-d10	0.876	1.092	-24.7#	134	0.00
31	Pyrene	1.145	1.423	-24.3#	133	0.00
32	Benzo(a)anthracene	1.188	1.228	-3.4	111	0.00
33	Chrysene	1.085	1.126	-3.8	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	84	0.00
35	Benzo(b)fluoranthene	1.162	1.279	-10.1	90	0.00
36	Benzo(k)fluoranthene	1.081	1.154	-6.8	86	0.00
37 S	Benzo(a)pyrene-d12	0.903	0.928	-2.8	84	0.00
38 C	Benzo(a)pyrene	1.114	1.148	-3.1	84	0.00
39	Indeno(1,2,3-cd)pyrene	1.328	1.258	5.3	76	0.00
40	Dibenzo(a,h)anthracene	1.104	1.044	5.4	76	0.00
41	Benzo(g,h,i)perylene	1.149	1.082	5.8	76	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062716\
Data File : BM006065.D
Acq On : 27 Jun 2016 09:21
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02038

Quant Time: Jun 28 02:09:03 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062516.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 27 10:49:48 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0