

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062816\
 Data File : BM006107.D
 Acq On : 28 Jun 2016 22:02
 Operator : UM/SJ
 Sample : H3780-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9YF6

Quant Time: Jun 29 01:15:45 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 29 00:57:09 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	208499	20.00	ng/ul	0.00
7) Naphthalene-d8	10.43	136	1005571	20.00	ng/ul	0.00
15) Acenaphthene-d10	14.30	164	618541	20.00	ng/ul	0.00
23) Phenanthrene-d10	17.05	188	1425333	20.00	ng/ul	0.00
29) Chrysene-d12	21.24	240	1512196	20.00	ng/ul	0.00
34) Perylene-d12	23.46	264	1221077	20.00	ng/ul	0.00

System Monitoring Compounds

2) 1,4-Dioxane-d8	3.17	96	8140	1.81	ng/uL	0.00
3) Phenol-d5	6.82	99	406317	23.85	ng/ul	0.00
4) Bis-(2-Chloroethyl)ether-d	6.99	67	257474	25.64	ng/ul	0.00
5) 2-Chlorophenol-d4	7.18	132	305900	23.40	ng/ul	0.00
6) 4-Methylphenol-d8	8.36	113	356260	25.99	ng/ul	0.00
8) Nitrobenzene-d5	8.80	128	169245	24.99	ng/ul	0.00
9) 2-Nitrophenol-d4	9.52	143	190365	24.56	ng/ul	0.00
10) 2,4-Dichlorophenol-d3	10.06	165	335417	23.71	ng/ul	0.00
12) 4-Chloroaniline-d4	10.57	131	438389	39.21	ng/ul	0.00
16) Dimethylphthalate-d6	13.72	166	1209935	27.29	ng/ul	0.00
17) Acenaphthylene-d8	13.99	160	1435908	26.69	ng/ul	0.00
20) 4-Nitrophenol-d4	14.51	143	205096	24.38	ng/ul	0.00
21) Fluorene-d10	15.30	176	998795	26.40	ng/ul	0.00
24) 4,6-Dinitro-2-methylphenol	15.42	200	154456	30.60	ng/ul	0.00
26) Anthracene-d10	17.15	188	1539583	26.93	ng/ul	0.00
30) Pyrene-d10	19.45	212	1726813	28.36	ng/ul	0.00
37) Benzo(a)pyrene-d12	23.32	264	1326096	27.29	ng/ul	0.00

Target Compounds

						Qvalue
25) Phenanthrene	17.09	178	78718	1.17	ng/ul	99
28) Fluoranthene	19.12	202	224761	2.72	ng/ul	98
31) Pyrene	19.48	202	171359	2.15	ng/ul	97
32) Benzo(a)anthracene	21.23	228	104809	1.33	ng/ul	99
33) Chrysene	21.28	228	103627	1.43	ng/ul	97
35) Benzo(b)fluoranthene	22.80	252	115967	1.83	ng/ul#	95
38) Benzo(a)pyrene	23.36	252	71881	1.20	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062816\
Data File : BM006107.D
Acq On : 28 Jun 2016 22:02
Operator : UM/SJ
Sample : H3780-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9YF6

Quant Time: Jun 29 01:15:45 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062816.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 29 00:57:09 2016
Response via : Initial Calibration

