

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062821\
 Data File : BM030653.D
 Acq On : 28 Jun 2021 13:10
 Operator : CG/JU
 Sample : SSTD1.606
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.606

Quant Time: Jun 28 13:42:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM062821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 12:59:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	1784	0.40	ng/ul	0.00
4) Naphthalene-d8	10.58	136	6460	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.42	164	3473	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.16	188	6557	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	6407	0.40	ng/ul	0.00
23) Perylene-d12	23.58	264	6060	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.28	96	3157	1.65	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.16	152	16794	1.63	ng/ul	0.00
18) Fluoranthene-d10	19.18	212	30184	1.54	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.31	88	3287	1.652	ng/ul#	81
5) Naphthalene	10.63	128	30232	1.573	ng/ul	98
7) 2-Methylnaphthalene	12.24	142	20479	1.580	ng/ul	99
8) 1-Methylnaphthalene	12.46	142	19983	1.586	ng/ul	100
10) Acenaphthylene	14.14	152	25500	1.569	ng/ul	98
11) Acenaphthene	14.48	153	20537	1.551	ng/ul	98
12) Fluorene	15.47	166	22869	1.560	ng/ul	99
14) Pentachlorophenol	16.82	266	3414	1.787	ng/ul	96
15) Phenanthrene	17.20	178	35449	1.518	ng/ul	98
16) Anthracene	17.29	178	32048	1.567	ng/ul	97
19) Fluoranthene	19.21	202	41819	1.420	ng/ul	98
20) Pyrene	19.57	202	42426	1.406	ng/ul	97
21) Benzo(a)anthracene	21.31	228	38242	1.503	ng/ul	99
22) Chrysene	21.36	228	40855	1.446	ng/ul	98
24) Benzo(b)fluoranthene	22.90	252	43294	1.642	ng/ul	89
25) Benzo(k)fluoranthene	22.94	252	43127	1.560	ng/ul#	89
26) Benzo(a)pyrene	23.48	252	36898	1.577	ng/ul#	85
27) Indeno(1,2,3-cd)pyrene	25.86	276	50057	1.683	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.87	278	40457	1.705	ng/ul	91
29) Benzo(g,h,i)perylene	26.56	276	43241	1.683	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062821\
 Data File : BM030653.D
 Acq On : 28 Jun 2021 13:10
 Operator : CG/JU
 Sample : SSTD1.606
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 Client Sampled :
 SSTD1.6606

Quant Time: Jun 28 13:42:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM062821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 12:59:24 2021
 Response via : Initial Calibration

