

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062821\
 Data File : BM030654.D
 Acq On : 28 Jun 2021 13:47
 Operator : CG/JU
 Sample : SSTD3.201
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2601

Quant Time: Jun 28 14:30:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM062821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 12:59:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	1806	0.40	ng/ul	0.00
4) Naphthalene-d8	10.58	136	6526	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.42	164	3501	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.16	188	6626	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	6608	0.40	ng/ul	0.00
23) Perylene-d12	23.58	264	6150	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.28	96	6370	3.30	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.16	152	34628	3.32	ng/ul	0.00
18) Fluoranthene-d10	19.18	212	62621	3.09	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.31	88	6758	3.355	ng/ul#	77
5) Naphthalene	10.63	128	62400	3.214	ng/ul	97
7) 2-Methylnaphthalene	12.24	142	42577	3.251	ng/ul	99
8) 1-Methylnaphthalene	12.46	142	41103	3.229	ng/ul	99
10) Acenaphthylene	14.14	152	54601	3.333	ng/ul	98
11) Acenaphthene	14.48	153	42647	3.195	ng/ul	99
12) Fluorene	15.46	166	47807	3.234	ng/ul	99
14) Pentachlorophenol	16.81	266	7280	3.771	ng/ul	95
15) Phenanthrene	17.19	178	73423	3.112	ng/ul	98
16) Anthracene	17.29	178	68633	3.321	ng/ul	97
19) Fluoranthene	19.21	202	86720	2.855	ng/ul	98
20) Pyrene	19.57	202	88039	2.829	ng/ul	98
21) Benzo(a)anthracene	21.31	228	81218	3.094	ng/ul	99
22) Chrysene	21.36	228	84774	2.909	ng/ul	99
24) Benzo(b)fluoranthene	22.90	252	89167	3.333	ng/ul	87
25) Benzo(k)fluoranthene	22.94	252	90349	3.220	ng/ul#	87
26) Benzo(a)pyrene	23.48	252	77336	3.258	ng/ul#	83
27) Indeno(1,2,3-cd)pyrene	25.86	276	103668	3.434	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.87	278	84357	3.504	ng/ul#	90
29) Benzo(a,h,i)perylene	26.56	276	88557	3.395	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM062821\
Data File : BM030654.D
Acq On : 28 Jun 2021 13:47
Operator : CG/JU
Sample : SST3.201
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2601

Quant Time: Jun 28 14:30:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM062821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 28 12:59:24 2021
Response via : Initial Calibration

