

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM062821\
 Data File : BM030655.D
 Acq On : 28 Jun 2021 14:38
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV710

Quant Time: Jun 28 15:22:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM062821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 14:33:12 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	2064	0.40	ng/ul	0.00
4) Naphthalene-d8	10.58	136	7218	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.42	164	3798	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.16	188	6994	0.40	ng/ul	0.00
17) Chrysene-d12	21.33	240	6704	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	6285	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.28	96	855	0.38	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.17	152	4308	0.38	ng/ul	0.00
18) Fluoranthene-d10	19.18	212	7437	0.39	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.32	88	861	0.371	ng/ul	88
5) Naphthalene	10.63	128	7519	0.366	ng/ul	100
7) 2-Methylnaphthalene	12.24	142	5014	0.362	ng/ul	99
8) 1-Methylnaphthalene	12.46	142	4589	0.341	ng/ul	99
10) Acenaphthylene	14.14	152	6366	0.375	ng/ul	99
11) Acenaphthene	14.48	153	5010	0.365	ng/ul	99
12) Fluorene	15.47	166	5448	0.359	ng/ul	99
14) Pentachlorophenol	16.82	266	739	0.324	ng/ul	97
15) Phenanthrene	17.20	178	8574	0.365	ng/ul	99
16) Anthracene	17.29	178	7186	0.349	ng/ul	99
19) Fluoranthene	19.21	202	10525	0.356	ng/ul	99
20) Pyrene	19.57	202	10525	0.354	ng/ul	99
21) Benzo(a)anthracene	21.31	228	9046	0.338	ng/ul	99
22) Chrysene	21.36	228	10788	0.360	ng/ul	99
24) Benzo(b)fluoranthene	22.91	252	10688	0.384	ng/ul	99
25) Benzo(k)fluoranthene	22.95	252	10613	0.370	ng/ul	99
26) Benzo(a)pyrene	23.49	252	9496	0.401	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.87	276	11675	0.379	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.88	278	9424	0.381	ng/ul	100
29) Benzo(a,h,i)perylene	26.57	276	10461	0.390	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM062821\  
Data File : BM030655.D  
Acq On    : 28 Jun 2021  14:38  
Operator  : CG/JU  
Sample    : SSTDICV0.4  
Misc      :  
ALS Vial  : 9      Sample Multiplier: 1
```

Quant Time: Jun 28 15:22:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM062821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 28 14:33:12 2021
Response via : Initial Calibration