

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062822\
 Data File : BM035743.D
 Acq On : 29 Jun 2022 03:27
 Operator : CG/JU
 Sample : PB145703BL
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK703

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/29/2022
 Supervised By :mohammad ahmed 07/01/2022

Quant Time: Jun 29 05:19:49 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM062522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 05:17:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.795	152	4351m	0.400	ng/u1	-0.07
4) Naphthalene-d8	10.572	136	13208	0.400	ng/u1	-0.05
9) Acenaphthene-d10	14.419	164	6882m	0.400	ng/u1	-0.01
13) Phenanthrene-d10	17.166	188	12724m	0.400	ng/u1	-0.03
17) Chrysene-d12	21.391	240	10595	0.400	ng/u1	# 0.02
23) Perylene-d12	23.685	264	15580	0.400	ng/u1	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.209	96	36652	6.207	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.228	152	5429	0.262	ng/u1	0.00
18) Fluoranthene-d10	19.202	212	10531	0.295	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM062822\
Data File : BM035743.D
Acq On : 29 Jun 2022 03:27
Operator : CG/JU
Sample : PB145703BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK703

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 06/29/2022
Supervised By :mohammad ahmed 07/01/2022

Quant Time: Jun 29 05:19:49 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM062522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 05:17:22 2022
Response via : Initial Calibration

