

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062916\
 Data File : BM006119.D
 Acq On : 29 Jun 2016 10:45
 Operator : UM/SJ
 Sample : PB91640BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK40

Quant Time: Jun 29 13:49:21 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 29 00:57:09 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	192219	20.00	ng/ul	0.00
7) Naphthalene-d8	10.43	136	915893	20.00	ng/ul	0.00
15) Acenaphthene-d10	14.30	164	580354	20.00	ng/ul	0.00
23) Phenanthrene-d10	17.05	188	1393995	20.00	ng/ul	0.00
29) Chrysene-d12	21.25	240	1712496	20.00	ng/ul	0.00
34) Perylene-d12	23.47	264	1627640	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.17	96	11773	2.85	ng/uL	0.00
3) Phenol-d5	6.83	99	488274	31.09	ng/ul	0.00
4) Bis-(2-Chloroethyl)ether-d	6.99	67	314687	33.99	ng/ul	0.00
5) 2-Chlorophenol-d4	7.19	132	373682	31.00	ng/ul	0.00
6) 4-Methylphenol-d8	8.36	113	405423	32.08	ng/ul	0.00
8) Nitrobenzene-d5	8.80	128	200993	32.58	ng/ul	0.00
9) 2-Nitrophenol-d4	9.52	143	220706	31.26	ng/ul	0.00
10) 2,4-Dichlorophenol-d3	10.06	165	378350	29.37	ng/ul	0.00
12) 4-Chloroaniline-d4	10.57	131	578906	56.85	ng/ul	0.00
16) Dimethylphthalate-d6	13.72	166	1412736	33.96	ng/ul	0.00
17) Acenaphthylene-d8	13.99	160	1663788	32.96	ng/ul	0.00
20) 4-Nitrophenol-d4	14.51	143	255683	32.39	ng/ul	0.00
21) Fluorene-d10	15.30	176	1178344	33.20	ng/ul	0.00
24) 4,6-Dinitro-2-methylphenol	15.42	200	183300	37.12	ng/ul	0.00
26) Anthracene-d10	17.15	188	1919386	34.32	ng/ul	0.00
30) Pyrene-d10	19.45	212	2209015	32.03	ng/ul	0.00
37) Benzo(a)pyrene-d12	23.33	264	2060624	31.81	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM062916\
Data File : BM006119.D
Acq On : 29 Jun 2016 10:45
Operator : UM/SJ
Sample : PB91640BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK40

Quant Time: Jun 29 13:49:21 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062816.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 29 00:57:09 2016
Response via : Initial Calibration

