

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062916\
 Data File : BM006144.D
 Acq On : 30 Jun 2016 02:27
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02052

Quant Time: Jun 30 03:32:16 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 29 00:57:09 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	136	0.00
2 S	1,4-Dioxane-d8	0.431	0.322	25.3#	127	0.00
3 S	Phenol-d5	1.634	1.352	17.3	136	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.963	0.813	15.6	134	0.00
5 S	2-Chlorophenol-d4	1.254	1.013	19.2	135	0.00
6 S	4-Methylphenol-d8	1.315	1.099	16.4	136	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	135	0.00
8 S	Nitrobenzene-d5	0.135	0.111	17.8	139	0.00
9 S	2-Nitrophenol-d4	0.154	0.122	20.8#	133	0.00
10 S	2,4-Dichlorophenol-d3	0.281	0.228	18.9	136	0.00
11	Naphthalene	0.876	0.711	18.8	137	0.00
12 S	4-Chloroaniline-d4	0.222	0.180	18.9	120	0.00
13	2-Methylnaphthalene	0.665	0.544	18.2	136	0.00
14	1-Methylnaphthalene	0.632	0.512	19.0	136	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
16 S	Dimethylphthalate-d6	1.434	1.115	22.2#	129	0.00
17 S	Acenaphthylene-d8	1.740	1.412	18.9	134	0.00
18	Acenaphthylene	1.822	1.496	17.9	133	0.00
19 C	Acenaphthene	1.162	0.937	19.4	131	0.00
20 S	4-Nitrophenol-d4	0.272	0.215	21.0#	123	0.00
21 S	Fluorene-d10	1.223	0.989	19.1	133	0.00
22	Fluorene	1.323	1.079	18.4	131	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.071	0.079	-11.3	216#	0.00
25	Phenanthrene	0.946	0.771	18.5	129	0.00
26 S	Anthracene-d10	0.802	0.660	17.7	128	0.00
27	Anthracene	0.973	0.806	17.2	129	0.00
28 C	Fluoranthene	1.159	0.918	20.8	122	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
30 S	Pyrene-d10	0.805	0.766	4.8	120	0.00
31	Pyrene	1.055	0.999	5.3	119	0.00
32	Benzo(a)anthracene	1.040	0.848	18.5	100	0.00
33	Chrysene	0.957	0.781	18.4	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	73	0.00
35	Benzo(b)fluoranthene	1.038	0.846	18.5	75	0.00
36	Benzo(k)fluoranthene	0.974	0.829	14.9	76	0.00
37 S	Benzo(a)pyrene-d12	0.796	0.634	20.4#	72	0.00
38 C	Benzo(a)pyrene	0.981	0.787	19.8	72	0.00
39	Indeno(1,2,3-cd)pyrene	1.019	0.885	13.2	74	-0.01
40	Dibenzo(a,h)anthracene	0.843	0.738	12.5	74	0.00
41	Benzo(g,h,i)perylene	0.853	0.769	9.8	77	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM062916\
Data File : BM006144.D
Acq On : 30 Jun 2016 02:27
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02052

Quant Time: Jun 30 03:32:16 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM062816.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 29 00:57:09 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0